Research Article

Antibacterial Activity of As-Annealed TiO₂ Nanotubes Doped with Ag Nanoparticles against Periodontal Pathogens

Sinem Yeniyol,¹ Zhiming He,² Behiye Yüksel,³ Robert Joseph Boylan,² Mustafa Ürgen,⁴ Tayfun Özdemir,¹ and John Lawrence Ricci⁵

¹ Department of Oral Implantology, Faculty of Dentistry, Istanbul University, 34093 Istanbul, Turkey

² Department of Basic Science and Craniofacial Biology, New York University College of Dentistry, New York, NY 10010, USA

³ Department of Mechanical Engineering, Istanbul Aydın University, 34668 Istanbul, Turkey

⁴ Department of Metallurgical and Materials Engineering, Faculty of Chemical and Metallurgical Engineering, Istanbul Technical University, 34469 Istanbul, Turkey

⁵ Department of Biomaterials and Biomimetics, New York University College of Dentistry, New York, NY 10010, USA

Correspondence should be addressed to Sinem Yeniyol; yeniyols@istanbul.edu.tr

Received 29 May 2014; Revised 26 July 2014; Accepted 26 July 2014; Published 18 August 2014

Academic Editor: Ian S. Butler

Copyright © 2014 Sinem Yeniyol et al. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

It is important to develop functional transmucosal implant surfaces that reduce the number of initially adhering bacteria and they need to be modified to improve the anti-bacterial performance. Commercially pure Ti sheets were anodized in an electrolyte containing ethylene glycol, distilled water and ammonium fluoride at room temperature to produce TiO₂ nanotubes. These structures were then annealed at 450°C to transform them to anatase. As-annealed TiO₂ nanotubes were then treated in an electrolyte containing 80.7 g/L NiSO₄·7H₂O, 41 g/L MgSO₄·7H₂O, 45 g/L H₃BO₃, and 1.44 g/L Ag₂SO₄ at 20°C by the application of 9 V AC voltage for doping them with silver. As-annealed TiO₂ nanotubes and as-annealed Ag doped TiO₂ nanotubes were evaluated by SEM, FESEM, and XRD. Antibacterial activity was assessed by determining the adherence of *A. actinomycetemcomitans*, *T. forsythia*, and *C. rectus* to the surface of the nanotubes. Bacterial morphology was examined using an SEM. As-annealed Ag doped TiO₂ nanotubes revealed intense peak of Ag. Bacterial death against the as-annealed Ag doped TiO₂ nanotubes were detected against *A. actinomycetemcomitans*, *T. forsythia*, and *C. rectus* indicating antibacterial efficacy.

1. Introduction

Dental implants have become a widely applied treatment option in dentistry to replace missing teeth for function and esthetics. However, implant failure and peri-implantitis are problems to be solved to provide long-term stability of the dental implants which depend not only on the integration into the surrounding bone [1], but also on the presence of the protective soft tissue sealing around the implant [2].

The composition, configuration, and density of the proteins in the pellicle derived from the saliva and gingival crevicular fluid are largely dependent on the physical and chemical nature of the underlying surface and thus the properties of the surface influence bacterial adhesion through the pellicle [2]. The phrase “race for the surface” was coined

by Gristina in 1987 to describe the competition between bacterial adhesion and tissue integration [3]. If the native host bacteria win the race, tissue cells will not be able to displace these primary colonizers, and biofilm formation will occur developing into peri-implantitis [4, 5]. It is important to develop functional transmucosal implant surfaces that reduce the number of initially adhering bacteria. The first method is to inhibit the initial adhesion of oral bacteria. An ideal transmucosal implant surface exposed to the oral cavity is recommended to be highly polished to resist bacterial colonization and it is expected to allow the formation of an epithelial seal that prevents plaque accumulation leading to peri-implantitis [5, 6]. The second method is to inhibit the colonization of the oral bacteria, which involves surface antibacterial activity [7]. The surface needs modification to

optimize the antibacterial properties of the implant. The antibacterial characteristics of implants can be enhanced by mechanical, physical, chemical, and biochemical surface treatments. Electrochemical anodization has been receiving increasing attention as a chemical surface modification method for fabrication of highly ordered nanotubular titanium oxide (TiO_2) layers for the medical implants as a cost-effective, versatile, and simple technique [8–11]. Anodization leads to an oxidation of metal species that form a solid oxide on the metal surface. Depending on the anodization conditions (potential, nature of the electrolyte, concentration of the electrolyte, temperature, potential sweep rate, pH, and anodizing time) [12, 13], the solid oxide layer can be either compact, or nanotubular [14]. Ordered nanotubular structures of TiO_2 , with a controlled and uniform diameter, length, and wall thickness, can be formed if the dissolution is enhanced by fluoride containing electrolytes and suitable anodization conditions are used [14].

The nanotubular TiO_2 surface layers play an important role in the improvement of osseointegration through the enhancement of bone cell adhesion, differentiation, ALP activity, bone matrix deposition, apatite deposition rates [15], and hemocompatibility of Ti and Ti-based materials [16–18]. There are also studies validating TiO_2 nanotubes as promising bioactive coatings with predictable drug release characteristics for local drug delivery systems [19], killing of cancer cells [20], and bacteria cells [21] for anticancer and antibacterial treatments. With the increase of microorganisms resistant to multiple antibiotics, there is a need for alternative antibacterial agents [22]. Silver with its nontoxicity to human cells [23] is widely used as an antibacterial coating to avoid initial adhesion of bacteria onto the implant surface [23, 24]. It is difficult for bacteria to develop resistance against this element. It is effective at low concentrations and relatively large reservoir provided by the nanotubes can give rise to long-term antibacterial effects [25]. With this concept, antibacterial studies have focused on fabrication of TiO_2 nanotubes serving as carriers for Ag as an antibacterial agent. TiO_2 nanotubes were loaded with Ag by ultrasonication [26], soaking in AgNO_3 solutions [25], electrodeposition [27], and sputter deposition techniques [22] to generate surfaces showing adherent Ag nanoparticles uniformly distributed on the TiO_2 nanotube walls. In our study, electrochemical anodization technique, which is easy and cost effective, was used to fill in the TiO_2 nanotubes with Ag instead of generating distributed Ag particles on these nanotubes.

Periodontopathogen bacteria can attach intraoral components of implants that are exposed to saliva, plaque, and crevicular fluid and increase the risk for peri-implantitis infections. Therefore, the aim of this study was to indicate the possible clinical benefit of as-annealed Ag doped TiO_2 nanotubes in providing antimicrobial properties due to their Ag content against the adhesion of peri-implantitis-associated bacteria *Aggregatibacter actinomycetemcomitans*, *Tannerella forsythia*, and *Campylobacter rectus* for transmucosal components of dental implants.

2. Materials and Methods

2.1. Preparation of Samples. Commercially pure titanium (cpTi) sheets in squares ($10 \times 10 \times 1$ mm, 99.6% purity) were used as substrates for the experiments. These sheets were ultrasonically cleaned in acetone, distilled water, and methanol, respectively. The electrochemical anodization was employed to form a layer of TiO_2 nanotubes on the cpTi sheets. Anodization voltage was kept at 40 V with a DC power supply for 30 min at room temperature. Electrolyte was ethylene glycol with 0.5 wt% ammonium fluoride (NH_4F) and 3 vol% distilled water [28]. These sheets were then annealed at 450°C in air for 30 min to convert the amorphous TiO_2 nanotubes into the anatase phase [29]. These sheets containing as-annealed TiO_2 nanotubes were named as Group TiO_2 . Group TiO_2 sheets were cleaned with acetone and rinsed with distilled water after anodic oxidation, and they were immediately doped with Ag at 20°C with a constant voltage of 9 V with a DC power supply for 30 sec. Electrolyte contained 80.7 g/L $\text{NiSO}_4 \cdot 7\text{H}_2\text{O}$, 41 g/L $\text{MgSO}_4 \cdot 7\text{H}_2\text{O}$, 45 g/L H_3BO_3 , and 1.44 g/L Ag_2SO_4 [30, 31]. The sheets containing as-annealed TiO_2 nanotubes served as the cathode electrode and a platinum sheet as the counter electrode. These sheets containing as-annealed Ag doped TiO_2 nanotubes were named as Group Ag. Untreated cpTi sheets were named as Group Ti. 24-well cell culture plate bottoms were used as the control group named as Control Group at the antibacterial assay.

2.2. Characterization Methods. The surface morphologies of the sheets were observed using a scanning electron microscope (SEM) (JSM5410, JEOL, Tokyo, Japan) at a 25 kV acceleration voltage. Field emission scanning electron microscope (FESEM) (JSM-7000F, JEOL, Tokyo, Japan) was used to observe the microstructures of the thin films at 5 and 10 kV acceleration voltages and at various magnifications. The structure of the films and corresponding orientations of Ag- TiO_2 films were determined by utilizing an X-ray diffractometer (Philips PW 3710, Cu-K α radiation). A scan rate of $0.01^\circ/\text{sec}$ was used for cpTi surface and as-annealed TiO_2 films with a grazing incidence of 0.2° , but this grazing incidence was not sufficient to detect the Ag in the as-annealed TiO_2 nanotubes doped by Ag. For this reason, the grazing incidence used for the as-annealed TiO_2 nanotubes doped by Ag was as 0.5° .

2.3. Antibacterial Assay. The tests were performed using *A. actinomycetemcomitans* (ATCC 43718; ATCC, Rockville MD, USA), *T. forsythia* (ATCC43037A), and *C. rectus* (ATCC 33238). Bacteria cells were cultured in brain heart infusion (BHI) broth (Thermo Scientific Remel, Lenexa, KS, USA) overnight at 37°C . Based on our pilot studies, the bacteria were grown to mid-log phase and centrifuged and resuspended in trypticase soy broth to optical densities of approximately 0.40 for *A. actinomycetemcomitans*, 0.25 for *T. forsythia*, and 0.10 for *C. rectus* at the wavelength of 600 nm. Sheets of the Groups TiO_2 , Ag, Ti, and Control Group were placed into individual wells of the sterile 24-well culture plates with their modified surfaces placed facing

upward and bacteria cells were pipetted onto the samples for the three different bacteria experiments. The culture plates were covered by their lids and incubated at 37°C in an anaerobic environment (Modular Atmosphere Controlled System, DW Scientific, Shipley, Yorkshire, UK) for 18 h. The supernatant fluid from each well was appropriately diluted and plated on TSBN media (personal communication, S.S. Socransky, Forsyth Institute, Cambridge, MA, USA) for *A. actinomycetemcomitans*, *T. forsythia*, and *C. rectus* and incubated anaerobically for 4 days at 37°C and the number of colonies (colony-forming unit: CFU) was counted. TSBN was prepared as follows: solution A—26 g of brain heart infusion agar (Thermo Scientific Remel), 20 g of trypticase soy agar (Thermo Scientific Remel), 10 g of yeast extract (Difco Laboratories, Detroit, MI, USA), and 5 mg of hemin (Sigma-Aldrich, St. Louis, MO, USA) were added to 930 mL of distilled water. Solution A was autoclaved and then placed in a water bath. When a temperature of 52°C was reached, the following solutions were added aseptically: 10 mL of a menadione stock solution (5 mg/100 mL) (Sigma-Aldrich), 1 mL of an N-acetylmuramic acid stock solution (1 g/100 mL) (Sigma-Aldrich), and 50 mL of sheep blood (Thermo Scientific Remel).

The cell densities were chosen at different concentrations based on pilot adhesion experiments for being able to view different colonization concentrations of the selected bacteria cells on the material surfaces by SEM. Sheets at the Groups TiO₂, Ag, and Ti were subjected to fixation followed by SEM as described next.

2.4. Bacterial Morphology. Representative sheets colonized with the selected oral bacteria were prepared for SEM following standard procedures. Sheets were fixed in 2.5% glutaraldehyde for 1 h. After washing three times in the 0.1 M phosphate buffer, bacteria were postfixed with 1% OsO₄ for 1 h. After sheets were rinsed twice in the 0.1 M phosphate buffer, they were dehydrated through a graded alcohol series (25–100%). Hexamethyldisilazane was applied twice. Sheets were subsequently critical-point dried; sputter coated with gold, and examined using SEM (Philips XL 30, Eindhoven, The Netherlands) at 20 and 25 kV accelerating voltages.

2.5. Statistical Analysis. Statistical analysis was done online with VassarStats: Statistical Computation Web Site. Differences in the mean numbers of the microbes (CFUs) harvested from the experimental materials were tested with one-way analysis of variance (ANOVA) and post hoc analyses were performed using the Tukey's studentized range (HSD) test. All results were reported as mean ± standard deviation (SD). Threshold for significance was set for $P < 0.05$.

3. Results and Discussion

The goal of this study was to develop an antibacterial surface for the transmucosal components of dental implants less prone to periodontopathogen bacteria colonization. This objective was achieved via surface nanostructural modification by electrochemical anodization and annealing followed by Ag doping. This study showed that as-annealed Ag doped

TiO₂ nanotubes inhibited adhesion of *A. actinomycetemcomitans*, *T. forsythia*, or *C. rectus*.

It is generally accepted that the periodontopathogen bacteria play a crucial role in peri-implantitis through an assembly of putative virulent factors. *A. actinomycetemcomitans* is responsible for the induction of inflammation of the gingivae and destruction of the periodontal ligament and alveolar bone by modulating inflammation, inducing tissue destruction, and inhibiting tissue repair [32]. Epithelial cell invasion by *T. forsythia* is considered to be an important virulence mechanism and it has putative virulent factors such as trypsin-like protease, sialidase, hemagglutinin, components of the bacterial S-layer, and cell surface-associated and secreted protein (BspA) [33]. *C. rectus* is a bacterium reaching the deeper parts of the subgingival pockets using the motility of its flagellum that appear to be the major pathogenic factor [34]. To examine the antimicrobial properties of the surfaces, *A. actinomycetemcomitans*, *T. forsythia*, and *C. rectus* were chosen for this study as they are bacterial species associated with periodontal disease considering their individual putative virulent factors.

In our study, cpTi surfaces were observed to be smooth, with features of grooves, valleys, and peaks at the micron scale before electrochemical anodization (Figure 1(c)). These cpTi surfaces were electrochemically anodized with ethylene glycol, distilled water, and ammonium fluoride to produce highly ordered TiO₂ nanotube formation on the Ti substrates. Vertically orientated as-annealed TiO₂ nanotubes, with an inner diameter of 70–100 nm as grown on Ti substrates after electrochemical anodization, were formed in Group TiO₂ (Figure 1(a)). The side view image (inset of Figure 1(a)) indicated that the as-annealed TiO₂ nanotubes were straight with uniform pore walls opened at the top with a length of 4.5 μm. Photoresponse of these nanotubes is affected by tube geometry (length, diameter, and tube wall thickness) and structure (anatase, anatase/rutile). The longer the tube, the larger the surface area with higher total light absorption [14]. TiO₂ is usually used as a photocatalyst in two crystal structures: rutile and anatase. Anatase generally has much higher activity than rutile [35]. As-synthesized nanotubes are amorphous, and postannealing is required to crystallize them into anatase, rutile, or bookite structure [12]. They can be crystallized into anatase at temperatures higher than approximately 280°C in air or a mixture of anatase and rutile at temperatures higher than approximately 450°C [14, 36–39]. In this regard, we aimed to form crystallized TiO₂ nanotubes by annealing the sheets at 450°C after electrochemical anodization. This is confirmed by the results in the XRD pattern showing several dominant peaks of anatase phase after annealing process. They indicate diffraction peaks at $2\theta = 25.5^\circ$, 38.1° , and 48.3° that are identified to be (1 0 1), (0 0 4), and (2 0 0) crystal faces, respectively. All the anodized TiO₂ films after the annealing process contained anatase (JCPDS 21-1272) without any evidence for rutile structure (Figure 2). Anodization of cpTi surfaces (Group Ti) clearly diminished the adhesion of *C. rectus* on as-annealed TiO₂ nanotubes (Group TiO₂) ($P < 0.05$), whereas no significant differences were found in *A. actinomycetemcomitans* and *T. forsythia* adhesion ($P > 0.05$; all) (Figure 3). UV exposure for

FIGURE 1: (a) Top-view FESEM micrograph of the surface of Group TiO₂ displaying as-annealed TiO₂ nanotubes (bar = 100 nm). Inset: side view of the as-annealed TiO₂ nanotubes (bar = 1 μm). (b) Top-view FESEM micrograph of the surface of Group Ag displaying as-annealed Ag doped TiO₂ nanotubes (bar = 100 nm). Inset: side view of the as-annealed Ag doped TiO₂ nanotubes (bar = 1 μm). (c) Top-view SEM micrograph of the surface of Group Ti displaying cpTi sheet (bar = 10 μm).

FIGURE 2: X-ray diffraction patterns of TiO₂: as-annealed TiO₂ nanotubes at Group TiO₂; Ag: as-annealed Ag doped TiO₂ nanotubes at Group Ag (◆: Ti; ●: anatase; ■: Ag); Ti: cpTi surface at Group Ti.

photocatalytic activity was not applied in our experiments. It is widely known that TiO₂ photocatalysts have minimal antibacterial efficacy in visible light [40]. Antibacterial effect against selected periodontopathogen bacteria on the as-annealed TiO₂ nanotubes can be ascribed to either the visible light exposure or the bacterial sensitivity of *C. rectus* against the crystal structure of anatase.

The high efficacy of Ag at very low concentrations and the relatively large reservoir provided by the nanotubes offer long-term antibacterial effects. In our study, vertically aligned as-annealed TiO₂ nanotubes as grown on Ti-surfaces doped with Ag were formed in Group Ag (Figure 1(b)). The side view image (inset of Figure 1(b)) indicated that the as-annealed Ag doped TiO₂ nanotubes were grown in vertical direction in length up to 6 μm after doping. The mechanism of this reservoir was reported by Zhao et al. [25] as the release of the oxidized Ag⁺ by the slow infiltration of body fluids into the nanotubes leading to antibacterial

effect. They considered selecting TiO₂ nanotubes with a size smaller than 130 nm which should reduce water infiltration and accomplish controlled release of Ag⁺. In accordance with this finding, controlled release of Ag⁺ was expected for the nanotubes in our study with a diameter of 70–100 nm for the antibacterial effect. According to our data, The XRD pattern exhibited diffraction peaks in the pattern corresponding to anatase phase of TiO₂ and cpTi surface while the small peak at 2θ = 44.5 which was allocated to the diffraction of (2 0 0) plane of face centered cubic (FCC) silver marked with Ag (Figure 2) [41]. Anodized surfaces with as-annealed Ag doped TiO₂ nanotubes (Group Ag) did not enhance *A. actinomycetemcomitans*, *T. forsythia*, and *C. rectus* adhesion on titanium surface compared to the Control Group (*P* < 0.01; all) (Figure 3). This reduction in bacterial activity on the as-annealed Ag doped TiO₂ nanotubes showed that the use of as-annealed TiO₂ nanotubes containing Ag was effective in improving the antimicrobial properties of Ti-based materials. In accordance with our findings, on the surfaces of the Group TiO₂ and Group Ag, bacteria cells showed highly deteriorated morphologies (Figures 4(a) and 4(b)), while bacteria cells at Group Ti showed no morphological change indicating bacterial death (Figure 4(c)). As a consequence of bacterial deterioration, surfaces at Group TiO₂ and Group Ag were covered with dead biofilms composed of the bacteria remnants.

Antibacterial action of silver is not yet well understood, but it is suggested that Ag nanoparticles can cause bacterial penetration through interaction with sulfur-containing proteins at the bacterial membrane and the phosphorus containing compounds like DNA, finally leading to cell death by attacking the respiratory chain. Ag nanoparticles also release Ag⁺ ions enhancing their bactericidal activity by converting DNA from relaxed state into its condensed form and thereby preventing its replication which would lead to cell death [42]. It is reported that cell proliferation, adhesion, and spreading are improved by the TiO₂ layer. Therefore, it was suggested that the combination of antibacterial properties (from Ag) and biocompatibility (from TiO₂) of the TiO₂/Ag compound coating might be advantageous for medical use [43]. In this regard, titanium surface can be modified by TiO₂

FIGURE 3: Descriptive analysis of adhesion of (a) *A. actinomycetemcomitans*, (b) *T. forsythia*, and (c) *C. rectus* on all groups tested (Group TiO₂: as-annealed TiO₂ nanotubes; Group Ag: as-annealed Ag doped TiO₂ nanotubes; Group Ti: commercially pure Ti sheet; Control Group: 24-well cell culture plate bottoms). Data are presented as the mean $\pm$ SD (standard deviation). Results were analyzed using a one-way ANOVA and post hoc analyses were performed using Tukey's studentized range (HSD) test (* $P < 0.05$ and † $P < 0.01$).

nanotubes to enhance bone-cell materials interactions, and the nanoporous surface can then be silver coated to further improve antibacterial activity on the surface [44]. According to the investigators, eukaryotic cells show more structural and functional redundancy as bigger targets for attacking silver ions compared to prokaryotic cells acquiring higher silver ion concentrations to achieve comparable toxic effects, relative to the bacteria cells [45]. Though there are limited reports about the antimicrobial effect of the nanostructure of materials, the results of the research by Zheng et al. [46] suggest that not only was Ag-implanted titanium reported to promote osteogenesis with increased cell attachment, viability, and

osteogenic gene expression, but also it appeared to show a strong antimicrobial effect against oral microorganisms including *S. mutans*, *P. gingivalis*, and *C. albicans*. Das et al. [44] revealed that Ag-treated TiO₂ nanotube surface provided antibacterial properties against *P. aeruginosa* without interference to the attachment and proliferation of human osteoblasts. Ewald et al. [47] achieved the establishment of titanium/silver hard coatings via physical vapor deposition with significant antimicrobial potency and absence of cytotoxic effects on osteoblasts and epithelial cells. The antibacterial resistance demonstrated in our study is consistent with the previous studies demonstrating antimicrobial effects

FIGURE 4: SEM micrographs after adhesion of *A. actinomycetemcomitans*, *T. forsythia*, and *C. rectus* on the surface of (a) Group TiO₂: as-annealed TiO₂ nanotubes; (b) Group Ag: as-annealed Ag doped TiO₂ nanotubes; (c) Group Ti: commercially pure Ti sheet.

of silver ions which make them promising for combating postoperative infection for application in dental implant placement procedures.

4. Conclusions

It is of great importance to provide antibacterial activity for maintaining plaque-free surfaces on transmucosal parts exposed to the oral cavity as a future strategy. This study demonstrates that the use of electrochemical anodization and Ag doping provides the required antibacterial surface properties against the selected periodontal pathogens, *A. actinomycetemcomitans*, *T. forsythia*, and *C. rectus*, resulting in reproducible antibacterial coatings on transmucosal parts of dental implants. The findings, however, have to be verified in clinical settings.

Conflict of Interests

The authors declare that there is no conflict of interests regarding the publication of this paper.

Acknowledgment

This research was supported by Istanbul University.

References

- [1] J. Deng, T. Arimoto, Y. Shibata, S. Omori, T. Miyazaki, and T. Igarashi, "Role of chloride formed on anodized titanium surfaces against an oral microorganism," *Journal of Biomaterials Applications*, vol. 25, no. 2, pp. 179–189, 2010.
- [2] V. Fröjd, P. Linderbäck, A. Wennerberg, L. Chávez de Paz, G. Svensäter, and J. R. Davies, "Effect of nanoporous TiO₂ coating and anodized Ca²⁺ modification of titanium surfaces on early microbial biofilm formation," *BMC Oral Health*, vol. 11, no. 8, 2011.
- [3] A. G. Gristina, "Biomaterial-centered infection: microbial adhesion versus tissue integration," *Science*, vol. 237, no. 4822, pp. 1588–1595, 1987.
- [4] K. G. Neoh, X. Hu, D. Zheng, and E. T. Kang, "Balancing osteoblast functions and bacterial adhesion on functionalized titanium surfaces," *Biomaterials*, vol. 33, no. 10, pp. 2813–2822, 2012.

- [5] B. Grössner-Schreiber, M. Griepentrog, I. Haustein et al., "Plaque formation on surface modified dental implants—an in vitro study," *Clinical Oral Implants Research*, vol. 12, no. 6, pp. 543–551, 2001.
- [6] R. Bürgers, T. Gerlach, S. Hahnel, F. Schwarz, G. Handel, and M. Gosau, "In vivo and in vitro biofilm formation on two different titanium implant surfaces," *Clinical Oral Implants Research*, vol. 21, no. 2, pp. 156–164, 2010.
- [7] M. Yoshinari, Y. Oda, T. Kato, and K. Okuda, "Influence of surface modifications to titanium on antibacterial activity in vitro," *Biomaterials*, vol. 22, no. 14, pp. 2043–2048, 2001.
- [8] B. S. Smith, S. Yoriya, T. Johnson, and K. C. Papat, "Dermal fibroblast and epidermal keratinocyte functionality on titania nanotube arrays," *Acta Biomaterialia*, vol. 7, no. 6, pp. 2686–2696, 2011.
- [9] K. Gulati, M. S. Aw, and D. Losic, "Drug-eluting Ti wires with titania nanotube arrays for bone fixation and reduced bone infection," *Nanoscale Research Letters*, vol. 6, article 571, pp. 1–6, 2011.
- [10] S. D. Puckett, P. P. Lee, D. M. Ciombor, R. K. Aaron, and T. J. Webster, "Nanotextured titanium surfaces for enhancing skin growth on transcutaneous osseointegrated devices," *Acta Biomaterialia*, vol. 6, no. 6, pp. 2352–2362, 2010.
- [11] J. Karlsson, R. Jimbo, H. M. Fathali et al., "In vivo biomechanical stability of osseointegrating mesoporous TiO₂ implants," *Acta Biomaterialia*, vol. 8, no. 12, pp. 4438–4446, 2012.
- [12] A. W. Tan, B. Pingguan-Murphy, R. Ahmad, and S. A. Akbar, "Review of titania nanotubes: fabrication and cellular response," *Ceramics International*, vol. 38, no. 6, pp. 4421–4435, 2012.
- [13] V. S. Saji, H. C. Choe, and W. A. Brantley, "An electrochemical study on self-ordered nanoporous and nanotubular oxide on Ti-35Nb-5Ta-7Zr alloy for biomedical applications," *Acta Biomaterialia*, vol. 5, no. 6, pp. 2303–2310, 2009.
- [14] J. M. Macak, H. Tsuchiya, A. Ghicov et al., "TiO₂ nanotubes: Self-organized electrochemical formation, properties and applications," *Current Opinion in Solid State and Materials Science*, vol. 11, no. 1-2, pp. 3–18, 2007.
- [15] A. Kodama, S. Bauer, A. Komatsu, H. Asoh, S. Ono, and P. Schmuki, "Bioactivation of titanium surfaces using coatings of TiO₂ nanotubes rapidly pre-loaded with synthetic hydroxyapatite," *Acta Biomaterialia*, vol. 5, no. 6, pp. 2322–2330, 2009.
- [16] B. S. Smith, S. Yoriya, L. Grissom, C. A. Grimes, and K. C. Papat, "Hemocompatibility of titania nanotube arrays," *Journal of Biomedical Materials Research A*, vol. 95, no. 2, pp. 350–360, 2010.
- [17] K. C. Papat, L. Leoni, C. A. Grimes, and T. A. Desai, "Influence of engineered titania nanotubular surfaces on bone cells," *Biomaterials*, vol. 28, no. 21, pp. 3188–3197, 2007.
- [18] K. S. Brammer, S. Oh, C. J. Cobb, L. M. Bjursten, H. van der Heyde, and S. Jin, "Improved bone-forming functionality on diameter-controlled TiO₂ nanotube surface," *Acta Biomaterialia*, vol. 5, no. 8, pp. 3215–3223, 2009.
- [19] Y. Hu, K. Cai, Z. Luo et al., "TiO₂ nanotubes as drug nanoreservoirs for the regulation of mobility and differentiation of mesenchymal stem cells," *Acta Biomaterialia*, vol. 8, no. 1, pp. 439–448, 2012.
- [20] K. Gulati, M. S. Aw, and D. Losic, "Nanoengineered drug-releasing Ti wires as an alternative for local delivery of chemotherapeutics in the brain," *International Journal of Nanomedicine*, vol. 7, pp. 2069–2076, 2012.
- [21] X. Chen, K. Cai, J. Fang et al., "Fabrication of selenium-deposited and chitosan-coated titania nanotubes with anticancer and antibacterial properties," *Colloids and Surfaces B: Biointerfaces*, vol. 103, pp. 149–157, 2013.
- [22] A. Roguska, M. Pisarek, M. Andrzejczuk, M. Lewandowska, K. J. Kurzydłowski, and M. Janik-Czachor, "Surface characterization of Ca-P/Ag/TiO₂ nanotube composite layers on Ti intended for biomedical applications," *Journal of Biomedical Materials Research A*, vol. 100, no. 8, pp. 1954–1962, 2012.
- [23] T. J. Berger, J. A. Spadaro, S. E. Chapin, and R. O. Becker, "Electrically generated silver ions: quantitative effects on bacterial and mammalian cells," *Antimicrobial Agents and Chemotherapy*, vol. 9, no. 2, pp. 357–358, 1976.
- [24] J. Zhao, X. M. Cai, H. Q. Tang, T. Liu, H. Q. Gu, and R. Z. Cui, "Bactericidal and biocompatible properties of TiN/Ag multilayered films by ion beam assisted deposition," *Journal of Materials Science: Materials in Medicine*, vol. 20, no. 1, pp. S101–S105, 2009.
- [25] L. Zhao, H. Wang, K. Huo et al., "Antibacterial nano-structured titania coating incorporated with silver nanoparticles," *Biomaterials*, vol. 32, no. 24, pp. 5706–5716, 2011.
- [26] L. Sun, J. Li, C. Wang et al., "Ultrasound aided photochemical synthesis of Ag loaded TiO₂ nanotube arrays to enhance photocatalytic activity," *Journal of Hazardous Materials*, vol. 171, no. 1–3, pp. 1045–1050, 2009.
- [27] Q. Ma, S. Mei, K. Ji, Y. Zhang, and P. K. Chu, "Immobilization of Ag nanoparticles/FGF-2 on a modified titanium implant surface and improved human gingival fibroblasts behavior," *Journal of Biomedical Materials Research A*, vol. 98, no. 2, pp. 274–286, 2011.
- [28] G. Liu, N. Hoivik, X. Wang, S. Lu, K. Wang, and H. Jakobsen, "Photoconductive, free-standing crystallized TiO₂ nanotube membranes," *Electrochimica Acta*, vol. 93, pp. 80–86, 2013.
- [29] D. Fang, Z. Luo, K. Huang, and D. C. Lagoudas, "Effect of heat treatment on morphology, crystalline structure and photocatalysis properties of TiO₂ nanotubes on Ti substrate and freestanding membrane," *Applied Surface Science*, vol. 257, no. 15, pp. 6451–6461, 2011.
- [30] G. J. Chi, S. W. Yao, J. Fan, W. G. Zhang, and H. Z. Wang, "Antibacterial activity of anodized aluminum with deposited silver," *Surface and Coatings Technology*, vol. 157, no. 2–3, pp. 162–165, 2002.
- [31] A. Hakimzad, K. Raeissi, and F. Ashrafzadeh, "Characterization of aluminum anodized layers modified in sulfuric and phosphoric acid baths and their effect on conventional electrolytic coloring," *Surface and Coatings Technology*, vol. 206, no. 8–9, pp. 2438–2445, 2012.
- [32] B. Henderson, M. Wilson, L. Sharp, and J. M. Ward, "Actinobacillus actinomycetemcomitans," *Journal of Medical Microbiology*, vol. 51, no. 12, pp. 1013–1020, 2002.
- [33] N. Suzuki, M. Yoneda, and T. Hirofujii, "Mixed red-complex bacterial infection in periodontitis," *International Journal of Dentistry*, vol. 2013, Article ID 587279, 6 pages, 2013.
- [34] H. Ihara, T. Miura, T. Kato et al., "Detection of Campylobacter rectus in periodontitis sites by monoclonal antibodies," *Journal of Periodontal Research*, vol. 38, no. 1, pp. 64–72, 2003.
- [35] T. Kawahara, Y. Konishi, H. Tada, N. Tohge, J. Nishii, and S. Ito, "A patterned TiO₂ (anatase)/TiO₂ (rutile) bilayer-type photocatalyst: effect of the anatase/rutile junction on the photocatalytic activity," *Angewandte Chemie*, vol. 41, no. 15, pp. 2811–2813, 2002.

- [36] O. K. Varghese, D. Gong, M. Paulose, C. A. Grimes, and E. C. Dickey, "Crystallization and high-temperature structural stability of titanium oxide nanotube arrays," *Journal of Materials Research*, vol. 18, no. 1, pp. 156–165, 2003.
- [37] J. M. Macak, S. Aldabergerova, A. Ghicov, and P. Schmuki, "Smooth anodic TiO₂ nanotubes: annealing and structure," *Physica Status Solidi A: Applications and Materials Science*, vol. 203, no. 10, pp. R67–R69, 2006.
- [38] A. Ghicov, H. Tsuchiya, J. M. Macak, and P. Schmuki, "Annealing effects on the photoresponse of TiO₂ nanotubes," *Physica Status Solidi A*, vol. 203, no. 4, pp. R28–R30, 2006.
- [39] J. Zhao, X. Wang, T. Sun, and L. Li, "In situ templated synthesis of anatase single-crystal nanotube arrays," *Nanotechnology*, vol. 16, no. 10, pp. 2450–2454, 2005.
- [40] K. J. Shieh, M. Li, Y. H. Lee, S. D. Sheu, Y. T. Liu, and Y. C. Wang, "Antibacterial performance of photocatalyst thin film fabricated by defection effect in visible light," *Nanomedicine: Nanotechnology, Biology, and Medicine*, vol. 2, no. 2, pp. 121–126, 2006.
- [41] A. A. Hebeish, M. M. Abdelhady, and A. M. Youssef, "TiO₂ nanowire and TiO₂ nanowire doped Ag-PVP nanocomposite for antimicrobial and self-cleaning cotton textile," *Carbohydrate Polymers*, vol. 91, no. 2, pp. 549–559, 2013.
- [42] M. Rai, A. Yadav, and A. Gade, "Silver nanoparticles as a new generation of antimicrobials," *Biotechnology Advances*, vol. 27, no. 1, pp. 76–83, 2009.
- [43] Y. Y. Chang, C. H. Lai, J. T. Hsu, C. H. Tang, W. C. Liao, and H. L. Huang, "Antibacterial properties and human gingival fibroblast cell compatibility of TiO₂/Ag compound coatings and ZnO films on titanium-based material," *Clinical Oral Investigations*, vol. 16, no. 1, pp. 95–100, 2012.
- [44] K. Das, S. Bose, A. Bandyopadhyay, B. Karandikar, and B. L. Gibbins, "Surface coatings for improvement of bone cell materials and antimicrobial activities of Ti implants," *Journal of Biomedical Materials Research B: Applied Biomaterials*, vol. 87, no. 2, pp. 455–460, 2008.
- [45] J. Liao, M. Anchun, Z. Zhu, and Y. Quan, "Antibacterial titanium plate deposited by silver nanoparticles exhibits cell compatibility," *International Journal of Nanomedicine*, vol. 5, no. 1, pp. 337–342, 2010.
- [46] Y. Zheng, J. Li, X. Liu, and J. Sun, "Antimicrobial and osteogenic effect of Ag-implanted titanium with a nanostructured surface," *International Journal of Nanomedicine*, vol. 13, no. 5, pp. 337–342, 2010.
- [47] A. Ewald, S. K. Glückermann, R. Thull, and U. Gbureck, "Antimicrobial titanium/silver PVD coatings on titanium," *BioMedical Engineering Online*, vol. 5, article 22, 2006.