Letter to the Editor

Formetric 4D Rasterstereography

Johnny Padulo¹ and Luca Paolo Ardigo²

¹ *Tunisian Research Laboratory “Sports Performance Optimization”, National Center of Medicine and Science in Sport, 263 Tunis, Tunisia*

² *School of Exercise and Sport Science, Department of Neurological and Movement Sciences, University of Verona, Via Felice Casorati, No. 43, 37131 Verona, Italy*

Correspondence should be addressed to Luca Paolo Ardigo; luca.ardigo@univr.it

Received 27 August 2013; Accepted 20 March 2014; Published 31 March 2014

Academic Editor: Matteo Paci

Copyright © 2014 J. Padulo and L. P. Ardigo. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

This paper is a comment on “Intra- and interday reliability of spine rasterstereography” [1]. We think that supporting the feasibility of using the rasterstereographic tool Formetric 4D, both as a diagnostic tool and in scientific research, requires much more scientific exactness. The purpose of this paper is to underline what we consider to be substantial omission or error.

In research, validity is the extent to which a conclusion corresponds accurately to the reality. Differently, reliability is the overall consistency of a measure [2]. Validity means a high level of accuracy, while reliability means a high level of precision [3]. The conclusions of the paper are supported by the results in terms of reliability (“In conclusion, the present study reveals a good to great reliability of the DIERS Formetric 4D system depending on the typology of the measured parameter.”) but not in terms of validity [4–6] (“Therefore, this study validated aspects of the rasterstereographic measuring system that potentially could replace X-rays in follow-up of spinal deformities helping to reduce X-rays irradiation.”) [1]. A clarification should be done in this paper to better explain the state of the art about the topic described in this paper.

In conclusion, “Therefore, this study validated aspects of the rasterstereographic measuring system that potentially could replace X-rays in follow-up of spinal deformities helping to reduce X-rays irradiation” [1]. Discussion should provide only those conclusions that are supported by the study [7]. At present, Formetric 4D cannot be considered

validated but only reliable. Further evidence-based research is necessary to support Formetric 4D effectiveness.

We should verify the data from this study with caution before suggesting such a clear indication [8, 9]. Attention of clinical researchers should focus on patients’ needs. Patients deserve both examination validity and reliability, and reduced exposure to ionizing radiation. Firstly, a minimum acceptable examination validity and reliability level has to be established. Then, researchers should investigate alternative methodologies capable of achieving at least that specific level. Sound research on this topic is continuously needed.

Conflict of Interests

There is no conflict of interests in this paper.

References

- [1] L. Guidetti, V. Bonavolonta, A. Tito, V. M. Reis, M. C. Gallotta, and C. Baldari, “Intra- and interday reliability of spine rasterstereography,” *BioMed Research International*, vol. 2013, Article ID 745480, 5 pages, 2013.
- [2] W. G. Hopkins, “Measures of reliability in sports medicine and science,” *Sports Medicine*, vol. 30, no. 1, pp. 1–15, 2000.
- [3] N. R. Carlson, W. Buskist, C. D. Heth, and R. Schmaltz, *Psychology: The Science of Behaviour*, Pearson Education, Ontario, Canada, 2009.

- [4] M. Mangone, P. Raimondi, M. Paoloni et al., "Vertebral rotation in adolescent idiopathic scoliosis calculated by radiograph and back surface analysis-based methods: correlation between the Raimondi method and rasterstereography," *European Spine Journal*, vol. 22, no. 2, pp. 367–371, 2013.
- [5] J. Padulo and L. P. Ardigò, "Letter to the Editor concerning, "Vertebral rotation in adolescent idiopathic scoliosis calculated by radiograph and back surface analysis-based methods: correlation between the Raimondi method and rasterstereography," *European Spine Journal*, vol. 22, no. 10, pp. 2336–2337, 2013.
- [6] J. Padulo and L. P. Ardigò, "Formetric rasterstereography: a new perspective," *Proceedings of the National Academy of Sciences of the United States of America*, 2014.
- [7] J. Padulo, N. Maffulli, and L. P. Ardigò, "Signal or noise, a statistical perspective," *Proceedings of the National Academy of Sciences of the United States of America*, 2014.
- [8] R. M. Rosenfeld, R. N. Shiffman, and P. Robertson, "Clinical Practice Guideline Development Manual, Third Edition: a quality-driven approach for translating evidence into action," *Otolaryngology: Head and Neck Surgery*, vol. 148, pp. S1–S55, 2013.
- [9] R. N. Shiffman and A. Wright, "Evidence-based clinical decision support," *Yearbook of Medical Informatics*, vol. 8, pp. 120–127, 2013.