Corrigendum

Corrigendum to “Advances of Techniques in Deep Regional Blocks”

Jui-An Lin ^{1,2} **Rafael Blanco**,³ **Yasuyuki Shibata** ⁴,
Tatsuo Nakamoto ⁵ and **Ko-Huan Lin**⁶

¹Department of Anesthesiology, Wan Fang Hospital, Taipei Medical University, Taipei, Taiwan

²Department of Anesthesiology, School of Medicine, College of Medicine, Taipei Medical University, Taipei, Taiwan

³Department of Anesthesia, Corniche Hospital, Abu Dhabi, UAE

⁴Department of Anesthesiology, Nagoya University Hospital, Nagoya, Japan

⁵Department of Anesthesiology, Kansai Medical University, Hirakata, Japan

⁶Division of Psychiatry, Hualien Armed Forces General Hospital, Hualien, Taiwan

Correspondence should be addressed to Jui-An Lin; juian.lin@tmu.edu.tw

Received 3 March 2018; Accepted 6 March 2018; Published 5 July 2018

Copyright © 2018 Jui-An Lin et al. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

In the article titled “Advances of Techniques in Deep Regional Blocks” [1], Table 1 was missing. Accordingly, the table is shown below, and its in-text citation is added as follows:

Our preliminary qualitative analysis demonstrated that, by using the pressure management system of Injectomat Agilia® pump (Fresenius Vial, Brezins, France) as an in-line manometer between the needle (the tip inserted 3 cm into the pork model) and the low-dead space extension tube, pushing pressure generated by the act of half-the-air was below 15 psi during injection (experiment was run in triplicate, and occlusion alarm did not occur after the flow had commenced in response to half-the-air pressure exerted in the 20 mL local anesthetic syringe with the syringe pump set to an infusion rate of 0.1 ml/h and a pressure limit of 750 mmHg) (Table 1).

TABLE 1: Qualitative analysis of injection pressure for the half-the-air setting.

Compressed air volume	50% (half-the-air) (5 ml compression)			60% (6 ml compression)		
Repeated experiments	Set 1	Set 2	Set 3	Set 1	Set 2	Set 3
Step 1: not open to the needle
●	●	●	●	●	●
Step 2: maintaining step 1 pressure and then open to the needle
●	●	●	Unable to assess the pressure by alarm in a consecutive (step 1 and then step 2) manner. (already alarmed in step 1)		

By using the pressure management system of Injectomat Agilia pump (Fresenius Vial, Brezins, France) as an in-line manometer, pressure within the half-the-air setting (ten milliliters of air was aspirated into the syringe above 10 ml local anesthetic [LA]) was assessed by adding an extra three-way stopcock between the needle (the tip inserted 3 cm into the pork) and the low-dead space extension tube with the side female luer lock connecting to the Agilia syringe pump via a pressure tube. The syringe pump was set to a minimal infusion rate of 0.1 ml/h and a pressure limit of 750 mmHg. Each experiment was run in triplicate. Green dot: “without” pressure (occlusion) alarm within 5 sec of injection. Red dot: “with” pressure (occlusion) alarm within 5 sec of injection. D5W: 5% dextrose water.

Acknowledgments

The work reported in this corrigendum was supported by a grant from Hualien Armed Forces General Hospital. Ko-Huan Lin (Division of Psychiatry, Hualien Armed Forces General Hospital, Hualien, Taiwan) and Jui-An Lin contributed equally to this corrigendum.

References

[1] J.-A. Lin, R. Blanco, Y. Shibata, and T. Nakamoto, “Advances of techniques in deep regional blocks,” *BioMed Research International*, vol. 2017, Article ID 7268308, 4 pages, 2017.