Research Article

The Effect of C-Arm Mobility and Field of Vision on Radiation Exposure in the Treatment of Proximal Femoral Fractures: A Randomized Clinical Trial

Mahmut Kalem,¹ Kerem Başarır,¹ Hakan Kocaoğlu ,¹ Ercan Şahin ,² and Hakan Kınık¹

¹Department of Orthopedics and Traumatology, Ankara University Faculty of Medicine, Ankara, Turkey

²Department of Orthopedics and Traumatology, Bulent Ecevit University Faculty of Medicine, Zonguldak, Turkey

Correspondence should be addressed to Hakan Kocaoğlu; kocaoglu@ankara.edu.tr

Received 19 July 2017; Accepted 25 February 2018; Published 27 March 2018

Academic Editor: Amal Khoury

Copyright © 2018 Mahmut Kalem et al. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Objectives. To examine the effect of fluoroscopy devices with different sizes of image intensifier and C-arm maneuverability on operating time, fluoroscopy time, radiation dose and reduction, and fixation quality at intertrochanteric femoral fractures. **Design.** Single-center, randomized, prospective study. **Setting.** Academic Level I trauma hospital. **Patients and Intervention.** 34 patients treated with cephalomedullary nailing for a stable, intertrochanteric proximal femur fracture (OTA A1). **Main Outcome Measurement.** The total working time of the fluoroscopy device, the dose-area product (DAP), operating time, reduction quality (cortical continuity, symmetrical collodiaphyseal angle, and shortness), and fixation quality (Bosworth quadrants, the tip-apex distance, TAD). **Results.** There were no cases of poor reduction; also the placement of the blade was optimal for 14 patients and suboptimal in 3 patients in each group. Superior-posterior placement of the blade or TAD > 25 mm was not seen in any patient. Total operating time was significantly shorter when using device A compared to the use of device B (20.1 ± 3.4 mins versus 25.3 ± 5.4 mins, $p < 0.001$). Total radiation time was significantly shorter with device A compared to the use of device B (58.1 ± 19.4 secs versus 98.9 ± 55.4 secs, $p = 0.008$). The measured radiation dose was lower with the use of device A compared to device B (3.5 ± 1.2 Gy·cm² versus 7.3 ± 4.5 Gy·cm², $p = 0.002$). **Conclusion.** Physical properties of fluoroscopy devices used during the fixation of intertrochanteric fractures could yield significant differences in operating times and the radiation dose while having comparable clinical results.

1. Introduction

The use of fluoroscopy for guidance in orthopedic trauma surgery has significantly increased, allowing for smaller surgical exposures to achieve reductions and internal fixation of fractures [1–4]. Over the years, there has been a tremendous increase in the use of fluoroscopy in orthopedic surgery, especially in treating proximal femur fractures [5, 6].

According to the aim for which current fluoroscopy devices are used, there are differences in software, power, working distance, maneuver capability of the C-arm, and image intensifier size. The basic principle in reducing radiation exposure caused by fluoroscopy devices is to keep the ionized radiation dose as low as possible [7, 8]. In literature it has been reported that many parameters are known to affect

operating time, fluoroscopy time, and the consequent ionized radiation dose. It is thought that one of these could be the technical property of the fluoroscopy device. Although the effects of fluoroscopy devices with different properties on operating time and radiation exposure have been researched in orthopedic surgery, the effects on the outcomes of their use in proximal femur fracture percutaneous treatment are not known [1, 9–12]. This study was undertaken with the thought that high C-arm maneuver capability and a wide field of vision could have an effect on operating time and exposure to ionized radiation. Therefore, the aim of the study was to examine the effect on operating time, fluoroscopy time, radiation dose, and reduction and fixation quality of 2 fluoroscopy devices of the same make and with the same software but with different sizes of image intensifier, which affects the field

FIGURE 1: The distinct features of two fluoroscopy devices used (device A, OEC 9900 Elite, General Electric Healthcare, USA; device B, OEC Brivo 785 Essential, General Electric Healthcare, USA). ((1A) and (1B)) Device A having a bigger image intensifier size and longer C-arm depth, which affects the field of vision, when compared to device B. ((2A) and (2B)) Device A has more degree of freedom of C-arm maneuverability when compared to device B. This enables the technician to find the area of interest only moving the C-arm while keeping the device still thus facilitating the use during surgery.

of vision, and different C-arm maneuver capability, which facilitates use during surgery; the devices were used by the same surgeon in the treatment of stable intertrochanteric femoral fractures using single fixation device.

2. Patients and Method

This single-center, prospective study included 34 patients treated with cephalomedullary nailing for a stable, intertrochanteric proximal femur fracture (Orthopaedic Trauma Association [OTA] A1 classification). Approval for the study was granted by the Local Ethics Committee. Between December 2016 and March 2017, a total of 134 consecutive patients presented at the Emergency Department of a major trauma center following a fall and were diagnosed with intertrochanteric femoral fracture and were treated with closed reduction and fixation with cephalomedullary nailing. Demographic characteristics (age, gender, side, and body mass index) were recorded as well as data related to the fracture type, operating time, the fluoroscopy device used, fluoroscopy time, radiation dose and reduction, and fixation quality. Patients with an AO/OTA Type A1 (stable) fracture and age > 60 years were included in the study. Patients with an additional fracture or pathological fracture were excluded from the study. Informed consent was obtained from all the patients or from a first-degree relative in the case of patients with dementia. All patients were operated on within 48 hours of arrival at the hospital.

From the total 134 patients, 40 were A1 stable intertrochanteric fractures. Randomization was achieved by the availability of the devices, when the patient was operated on. 17 patients were operated on with device B (OEC Brivo 785 Essential, General Electric Healthcare, USA) on days when device A (OEC® 9900 Elite, General Electric Healthcare, USA), which has the property of being able to apply digital subtraction angiography (DSA), was used by the vascular surgeons and 23 patients were operated on with device A in the orthopedic operating theatre on days when it was available. These two fluoroscopy devices were of the same make and with the same software but with different sizes of image intensifier, which affects the field of vision, and different C-arm maneuver capability, which facilitates use during surgery (Figure 1). Patients operated on using device A were BMI matched to the 17 patients of device B and thus the 2 study groups were formed (Figure 2).

All the patients were given regional anesthesia and were operated on in a supine position on a radiolucent table with the fractured hip and whole extremity draped. The fluoroscopy device was set up on the opposite side in a manner to show the fracture region. Fracture reduction and protection were provided with manual traction only by a single assistant. Reduction evaluation and the fixation application were made under fluoroscopy guidance. All the operations were performed by a single surgeon (MK). In all patients, internal fixation was applied using cephalomedullary nailing (standard proximal femoral nail antirotation, PFNA; Synthes

FIGURE 2: Scheme explaining patient selection process.

GmbH, Oberdorf, Switzerland). The fluoroscopy device in automatic dose control mode was used by the same radiology technician in all cases.

The total working time of the fluoroscopy device during the surgical procedure and the radiation dose used were obtained from the records of the fluoroscopy device at the end of surgery. The dose-area product (DAP) recorded during fluoroscopy imaging was the measurement method of the radiation dose selected in this study. DAP is a quantity used in assessing the radiation risk from radiographic examinations and interventional procedures and is defined as the absorbed dose multiplied by the area irradiated [6, 11, 13]. The operating time was calculated as the time from the initial incision to closure of the wound.

All the radiographic measurements were made by one of the authors (ES, who is blinded to the surgical intervention) on the standard AP as well as lateral radiographs in internal rotation obtained at the end of surgery, using Centricity PACS-IW software (General Electric Healthcare). Reduction quality was classified as described by Schipper et al. as anatomic (cortical continuity, symmetrical collodiaphyseal angle, and no shortness), good (5° – 10° varus/valgus), and poor ($>10^{\circ}$ varus/valgus) [14]. For fixation quality, the Cleveland-Bosworth quadrants were used which evaluate the position of the neck screw within the femoral head and the tip-apex distance (TAD) was measured [15]. Optimal fixation was accepted as central-central and inferior-central placement of the blade and TAD < 25 mm. Placements outside of this were accepted as suboptimal fixation, with the worst being a superior-posterior placement.

Statistical evaluations were made using IBM SPSS 15 (SPSS Inc., Chicago, IL, USA) software. Conformity of the data to normal distribution was evaluated with the Shapiro-Wilk test. Numerical variables were stated as mean $\pm$ standard deviation (SD) and categorical variables as percentages (%). Differences between the groups for numerical variables not with normal distribution were examined with the Mann-Whitney U test. A value of $p < 0.05$ was accepted as statistically significant.

3. Results

The study included a total of 34 patients comprising 17 males and 17 females with a mean age of 77.8 ± 9.0 years in the device A group and 77.8 ± 12.4 years in the device B group ($p > 0.05$) (Table 1).

When the radiological results were evaluated, in the device B group, anatomic reduction was obtained in 15 patients (88.2%) and good reduction in 2 (11.8%). In the device A group, anatomic reduction was obtained in all 17 patients (100%). There were no cases of poor reduction in either group ($p > 0.05$) (Table 2).

When the fixation quality was evaluated, optimal placement of the blade was the same in both groups ($n = 14$, 82.3%). Suboptimal placement was observed in 3 patients in each group. Superior-posterior placement of the blade was not seen in any patient. In both groups, the TAD measurement was < 25 mm in all patients (Table 2).

Total operating time was significantly shorter when using device A compared to the use of device B (20.1 ± 3.4 mins

TABLE 1: The demographic distribution of patients.

Variable	Device A (N = 17)	Device B (N = 17)
Age (yr)	77,8 ± 9,0	77,8 ± 12,4
Gender		
Female	10 (58,8%)	7 (41,2%)
Male	7 (41,2%)	10 (58,8%)
Laterality		
Left	9 (52,9%)	7 (41,2%)
Right	8 (47,1%)	10 (58,8%)

TABLE 2: The demographic radiological results of patient.

	Device A (N = 17)	Device B (N = 17)
Reduction		
Anatomic	17 (100%)	15 (88,2%)
Good	-	2 (11,8)
Bad	-	-
Blade position		
Optimal	14 (82,3%)	14 (82,3%)
Suboptimal	3 (17,7%)	3 (17,7%)
TAD (mm)	19,4 ± 2,9	22,0 ± 3,8

versus 25.3 ± 5.4 mins, $p < 0.001$). Total radiation time was significantly shorter with device A compared to the use of device B (58.1 ± 19.4 secs versus 98.9 ± 55.4 secs, $p = 0.008$). The measured radiation dose was lower with the use of device A compared to device B (3.5 ± 1.2 Gy·cm² versus 7.3 ± 4.5 Gy·cm², $p = 0.002$) (Figure 3).

4. Discussion

The functional results following treatment of hip fractures with percutaneous methods are closely related to reduction and fixation quality and the most important factor in achieving this is obtaining a good quality field of vision during the surgical procedure [10]. However, Mastrangelo et al. stated the importance of exposure to radiation caused during imaging and reported that orthopedic surgeons were exposed to 4 times more radiation than other surgeons and 8 times more than other healthcare personnel and there were increased rates of cancer in those working in orthopedics [16]. The amount of ionized radiation that occurs during imaging is known to be associated with the exposure time, the distance of the fluoroscopy device from the patient, the BMI of the patient, the projection angle and the extremity level, the type of operating table, and the experience of the surgeon [5, 12, 13, 17]. The results of the current study revealed that the technical properties of the fluoroscopy device had a significant effect on operating time, fluoroscopy time, and the amount of ionized radiation. To discount variables other than the fluoroscopy device in this study, all the surgical procedures were performed by the same surgeon; the same fixation system was used with the consideration that there may have been an effect of differences created

in the application of different systems, and only AO/OTA A1 fractures were examined which could be reduced with manual traction only taking into consideration the effects that could have arisen from direct reduction methods.

The operating time for the treatment of intertrochanteric femoral fractures with closed reduction and nailing has been reported in literature at varying periods in the range of 20 mins to 67 mins [18–21]. The reasons for such differences in the measurement of operating time can be considered to be that studies have been made on different types of fractures; the period has included different surgical steps, the type of anesthesia, the use of manual traction or a traction table, the implant selection, and the experience of the surgeon. When the operating times of the current study are compared with those in literature, they are consistent and a significant difference was determined between the groups (device A: 20.1 ± 3.4 min; device B: 25.3 ± 5.4 min, $p < 0.001$).

The fluoroscopy time in the closed reduction and nailing of intertrochanteric fractures has been associated with factors such as the severity of the fracture, implant selection, patient position, and the experience of the surgeon and radiology technician [5, 22–25]. There has been mention in literature of fluoroscopy times in trochanteric fractures with PFNA fixation but there are insufficient data on the use in stable A1 fractures. In a study by Zehir et al., the application of PFNA was evaluated in 92 patients, most of whom had an A2 fracture, and fluoroscopy time was reported as a mean of 1.50 min [26]. The fluoroscopy times in the current study were consistent with those in literature and a statistically significant difference was determined between the groups (device A: 58.1 ± 19.4 secs; device B: 98.9 ± 55.4 secs, $p = 0.008$).

When the radiation doses in the treatment of intertrochanteric fractures with cephalomedullary nailing are evaluated in literature, the DAP measurements associated with the fluoroscopy times are between 0.8 and 11 Gy·cm² [27–29]. The differences in radiation dose are likely associated with surgical experience and competence and condition of surgical sites [27]. The DAP measurements in the current study are consistent with literature and a statistically significant difference was determined between the groups (device A: 3.5 ± 1.2 Gy·cm²; device B: 7.3 ± 4.5 Gy·cm², $p = 0.002$).

To acquire images in a different plane, changing the place of the fluoroscopy device takes a certain time and extra exposures during the procedure [2, 10]. Even there are two distinct papers in the literature that argue the use of two C-arms simultaneously to face these problems [10, 30]. In the current study, the significant difference between the groups in respect of the operating times is thought to be due to the difference in the C-arm maneuver capability. Due to the larger image intensifier it was possible to take anterior-posterior and lateral images of the fracture region of interest in a single shoot and this was considered to be a reason for the significant differences in fluoroscopy times and therefore the doses of radiation exposure.

The most important limitation of the study was that the power analysis results were not sufficient as the number of

FIGURE 3: Box-plot graphs showing the results. (a) Box-plot showing total operation time for device A and device B. (b) Box-plot showing total radiation time for device A and device B. (c) Box-plot showing total radiation dose for device A and device B.

patients was low. Future studies examining more extensive series would be beneficial in establishing which parameters directly affect operating time, fluoroscopy time, and the amount of ionized radiation.

5. Conclusion

The results of this study showed that different physical properties of fluoroscopy devices used during the fixation of intertrochanteric fractures with cephalomedullary nailing caused differences in operating times and the radiation dose without having any effect on the radiological results. It can be considered appropriate to make the selection of a fluoroscopy device taking into account the frequency of application and cost.

Disclosure

Level of evidence is therapeutic, Level I.

Conflicts of Interest

Mahmut Kalem, Kerem Başarır, Hakan Kocaoğlu, and Ercan Şahin declare that they have no conflicts of interest. Hakan Kınık has received a speaker honorarium from Smith & Nephew Plc.

References

- [1] B. D. Giordano, J. F. Baumhauer, T. L. Morgan, and G. R. R. Ii, "Patient and surgeon radiation exposure: Comparison of standard and Mini-C-Arm fluoroscopy," *The Journal of Bone & Joint Surgery*, vol. 91, no. 2, pp. 297–304, 2009.
- [2] F. Matthews, D. J. Hoigne, M. Weiser et al., "Navigating the fluoroscope's C-arm back into position: An accurate and practicable solution to cut radiation and optimize intraoperative workflow," *Journal of Orthopaedic Trauma*, vol. 21, no. 10, pp. 687–692, 2007.
- [3] B. D. Giordano, J. N. Grauer, C. P. Miller, T. L. Morgan, and G. R. Rehtine, "Radiation exposure issues in orthopaedics," *Journal of Bone and Joint Surgery (American Volume)*, vol. 69, pp. 1–10, 2011.

- [4] R. Baumgartner, K. Libuit, D. Ren et al., "Reduction of radiation exposure from C-arm fluoroscopy during orthopaedic trauma operations with introduction of real-time dosimetry," *Journal of Orthopaedic Trauma*, vol. 30, no. 2, pp. e53–e58, 2016.
- [5] P. V. Giannoudis, J. McGuigan, and D. L. Shaw, "Ionising radiation during internal fixation of extracapsular neck of femur fractures," *Injury*, vol. 29, no. 6, pp. 469–472, 1998.
- [6] M. T. Crawley and A. T. Rogers, "Dose-area product measurements in a range of common orthopaedic procedures and their possible use in establishing local diagnostic reference levels," *British Journal of Radiology*, vol. 73, no. 871, pp. 740–744, 2000.
- [7] I. A. Khan, S. Kamalasekaran, and M. Ali Fazal, "Risk of ionising radiation to trainee orthopaedic surgeons," *Acta Orthopaedica Belgica*, vol. 78, no. 1, pp. 106–110, 2012.
- [8] F. D. Shuler, J. L. Daigre, D. Pham, and V. L. Kish, "Laser targeting with C-arm fluoroscopy: effect on image acquisition and radiation exposure," *Journal of Orthopaedic Trauma*, vol. 27, no. 5, pp. e97–e102, 2013.
- [9] A. Shoaib, U. Rethnam, R. Bansal, A. De, and N. Makwana, "A comparison of radiation exposure with the conventional versus mini C arm in orthopedic extremity surgery," *Foot & Ankle International*, vol. 29, no. 1, pp. 58–61, 2008.
- [10] Y. S. Brin, E. Palmanovich, E. Aliev et al., "Closed reduction and internal fixation for intertrochanteric femoral fractures is safer and more efficient using two fluoroscopes simultaneously," *Injury*, vol. 45, no. 7, pp. 1071–1075, 2014.
- [11] E. J. C. Dawe, E. Fawzy, J. Kaczynski, P. Hassman, and S. H. Palmer, "A comparative study of radiation dose and screening time between mini C-arm and standard fluoroscopy in elective foot and ankle surgery," *Journal of Foot and Ankle Surgery*, vol. 17, no. 1, pp. 33–36, 2011.
- [12] A. W. Su, A. L. McIntosh, B. A. Schueler et al., "How does patient radiation exposure compare with low-dose o-arm versus fluoroscopy for pedicle screw placement in idiopathic scoliosis?" *Journal of Pediatric Orthopaedics*, vol. 37, no. 3, pp. 171–177, 2017.
- [13] S. Dalglish, A. Hince, and D. F. Finlayson, "Peri-operative radiation exposure: Are overweight patients at increased risks?" *Injury*, vol. 46, no. 12, pp. 2448–2451, 2015.
- [14] I. B. Schipper, E. W. Steyerberg, R. M. Castelein et al., "Treatment of unstable trochanteric fractures: randomised comparison of the gamma nail and the proximal femoral nail," *The Journal of Bone & Joint Surgery (British Volume)*, vol. 86-B, no. 1, pp. 86–94, 2004.
- [15] Eland CLE, "A ten-year of intertrochanteric femur," *Journal of Bone and Joint Surgery (American Volume)*, 1959.
- [16] G. Mastrangelo, U. Fedeli, E. Fadda, A. Giovanazzi, L. Scozzato, and B. Saia, "Increased cancer risk among surgeons in an orthopaedic hospital," *Occupational Medicine*, vol. 55, no. 6, pp. 498–500, 2005.
- [17] S. Madan and C. Blakeway, "Radiation exposure to surgeon and patient in intramedullary nailing of the lower limb," *Injury*, vol. 33, no. 8, pp. 723–727, 2002.
- [18] I. Takigami, K. Matsumoto, A. Ohara et al., "Treatment of trochanteric fractures with the PFNA (Proximal Femoral Nail Antirotation) nail system report of early results," *Bulletin of the NYU Hospital for Joint Diseases*, vol. 66, no. 4, pp. 276–279, 2008.
- [19] H. M. Cho and K. Lee, "Clinical and functional outcomes of treatment for type a1 intertrochanteric femoral fracture in elderly patients: comparison of dynamic hip screw and proximal femoral nail antirotation," *Hip & Pelvis*, vol. 28, no. 4, pp. 232–242, 2016.
- [20] E. Şahin, M. Songür, M. Kalem et al., "Traction table versus manual traction in the intramedullary nailing of unstable intertrochanteric fractures: A prospective randomized trial," *Injury*, vol. 47, no. 7, pp. 1547–1554, 2016.
- [21] C. D'Arrigo, A. Carcangiu, D. Perugia et al., "Intertrochanteric fractures: comparison between two different locking nails," *International Orthopaedics*, vol. 36, no. 12, pp. 2545–2551, 2012.
- [22] J. Zou, Y. Xu, and H. Yang, "A comparison of proximal femoral nail antirotation and dynamic hip screw devices in trochanteric fractures," *Journal of International Medical Research*, vol. 37, no. 4, pp. 1057–1064, 2009.
- [23] X. Yaozeng, G. Dechun, Y. Huilin, Z. Guangming, and W. Xianbin, "Comparative study of trochanteric fracture treated with the proximal femoral nail anti-rotation and the third generation of gamma nail," *Injury*, vol. 41, no. 12, pp. 1234–1238, 2010.
- [24] S. Zhang, K. Zhang, Y. Jia, B. Yu, and W. Feng, "InterTan nail versus proximal femoral nail antirotation-asia in the treatment of unstable trochanteric fractures," *Orthopedics*, vol. 36, no. 3, pp. e288–e294, 2013.
- [25] M. D. Baratz, Y.-Y. Hu, D. Zurakowski, P. Appleton, and E. K. Rodriguez, "The primary determinants of radiation use during fixation of proximal femur fractures," *Injury*, vol. 45, no. 10, pp. 1614–1619, 2014.
- [26] S. Zehir, E. Şahin, and R. Zehir, "Comparison of clinical outcomes with three different intramedullary nailing devices in the treatment of unstable trochanteric fractures," *Ulusal Travma ve Acil Cerrahi Dergisi*, vol. 21, no. 6, pp. 469–476, 2015.
- [27] Y.-S. Lee, H.-K. Lee, J.-H. Cho, and H.-G. Kim, "Analysis of radiation risk to patients from intra-operative use of the mobile X-ray system (C-arm)," *Journal of Research in Medical Sciences*, vol. 20, no. 1, pp. 7–12, 2015.
- [28] A. Roux, N. Bronsard, N. Blanchet, and F. de Peretti, "Can fluoroscopy radiation exposure be measured in minimally invasive trauma surgery?" *Orthopaedics & Traumatology: Surgery & Research*, vol. 97, no. 6, pp. 662–667, 2011.
- [29] J. Hardman, M. Elvey, N. Shah, N. Simson, S. Patel, and R. Anakwe, "Defining reference levels for intra-operative radiation exposure in orthopaedic trauma: A retrospective multicentre study," *Injury*, vol. 46, no. 12, pp. 2457–2460, 2015.
- [30] A. Kara, H. Celik, A. Seker, M. Uzun, M. M. Sonmez, and M. Erdil, "Procedural outcomes of double vs. single fluoroscopy for fixing intertrochanteric femur fractures," *Archives of Orthopaedic and Trauma Surgery*, vol. 136, no. 7, pp. 929–934, 2016.