Original Article

Quantitative Sasang Constitution Diagnosis Method for Distinguishing between Tae-eumin and Soeumin Types Based on Elasticity Measurements of the Skin of the Human Hand

Han Wook Song, SungJun Lee, Yon Kyu Park and Sam Yong Woo

Center for Mass & Force, Division of Physical Metrology, Korea Research Institute of Standards and Science, 1 Doryeong Dong, Yuseong Gu, Daejeon, South Korea

The usefulness of constitutional diagnoses based on skin measurements has been established in oriental medicine. However, it is very difficult to standardize traditional diagnosis methods. According to Sasang constitutional medicine, humans can be distinguished based on properties of the skin, including its texture, roughness, hardness and elasticity. The elasticity of the skin was previously used to distinguish between people with Tae-eumin (TE) and Soeumin (SE) constitutions. The present study designed a system that uses a compression method to measure the elasticity of hand skin and evaluated its measurement repeatability. The proposed system was used to compare the skin elasticity between SE and TE subjects, which produced a measurement repeatability error of <3%. The proposed system is suitable for use as a quantitative constitution diagnosis method for distinguishing between TE and SE subjects with an acceptable level of uncertainty.

Keywords: compression – laser sensor – palpation – skin elasticity

Introduction

The usefulness of constitutional diagnoses of humans using the skin of the hand is well established in medicine. There have been many studies of constitutions with human skin. Examples include the dichotomy of Hippocrates, the quartering method of Hall, the three constitution theory of Sigauard, the cosmic dual-force fifty people principle (陰陽五十人論) of Huang Di Nei Jing (黃帝內經), and three docha theory popularized in India (1). However, it is very difficult to standardize traditional diagnosis methods because they are dependent on qualitative judgments of inspectors themselves. A constitutional diagnosis using human skin is typically performed via palpation. Through Sasang constitutional medicine, humans can be distinguished based on

properties of the skin, including its texture, roughness, hardness and elasticity (2). The constitutional skin status has previously been reported as follows.

In *Dongeuoosebowoon* (3), the properties of different types of human skin were categorized as:

if a Tae-cumin (TE) has a hard and dense skin that does not sweat, this person is likely very sick. If a Soeumin (SE) has a hard and dense skin without a sweat, he is very healthy. A TE has a solid skin and a SE has a smooth skin (太陰人陽剛堅密 則大病也, 少陰人 陽剛堅密 則完實也, 太陰人肌肉 堅實, 少陰人肌肉 浮軟, in the original Chinese characters).

In *Sasangyolam* (4), ancient Korean doctors also described the constitutional skin status:

We inspect skins. The skin status of Soyangins is slippery and thin, but occasionally some Soyangins have a thick skin. The skin of TEs is

For reprints and all correspondence: Han Wook Song, Center for Mass & Force, Division of Physical Metrology, Korea Research Institute of Standards and Science, 1 Doryeong Dong, Yuseong Gu, Daejeon, South Korea. Tel: +82-42-868-5821; Fax: +82-42-868-5012; E-mail: hanugi16@kriss.re.kr

© 2009 The Author(s).

This is an Open Access article distributed under the terms of the Creative Commons Attribution Non-Commercial License (<http://creativecommons.org/licenses/by-nc/2.0/uk/>) which permits unrestricted non-commercial use, distribution, and reproduction in any medium, provided the original work is properly cited.

Table 1. Questionnaire Survey for Doctors of Oriental Medicine

What are the major factors of skin for a constitution diagnosis?	
Factor	Percentage
Thickness (thick, thin)	89.2
Tactile sensation (slippery, soft, rough)	89.2
Elasticity (high, low)	70.3

thick and stiff; but TEs occasionally have a soft skin. The skin of SEs is buoyant and soft.

A recent questionnaire given to doctors of oriental medicine is presented in Table 1. Among the factors in this table, elasticity was used as the determining factor to distinguish between the TE and SE types of constitutions in several previous research sources (3–5) and several qualitative clinical studies (6–8). However, research on quantitative measurements of skin elasticity for constitutional diagnoses is lacking.

In this study, a system that uses a compression method to measure the elasticity of hand skin was designed and its measurement repeatability was evaluated. The difference in the elasticity of the hand skin between TE and SE subjects was also investigated.

Experimental Setup

Before fabricating the measurement system, the force required by a doctor of oriental medicine to grip the skin of a human hand was measured for inclusion in the design of the system based on previous work (9) (Figure 1).

The concept of the system for measuring the elasticity of human skin utilized a mechanical technique that involved direct contact with the human hand based on the method used by doctors of oriental medicine during a palpation assessment (Figure 2). Figure 3 shows the proposed system and Figure 4 shows a schematic of its contact component. The equipment (Figure 3) consisted of a compression tube that made contact with the skin (solid line), a laser sensor that measured the displacement before and after the compression step (dotted line), a motor that controlled the motion of the compression tube (dot-dashed line) and a controller with a pump for compression and a valve for handling any leaks (double dot-dashed line). The laser sensor was selected to minimize dispersion of the hand skin and to prevent damage to the skin. Table 2 lists the specifications of the measurement system. A carrying jig (Figure 5) was fabricated to ensure perpendicular contact between the sensor tip and the hand. The jig was tilted as required to maximize the flatness of the contact point.

The measured data were transferred to a personal computer (PC) using a USB-type DAQ (data acquisition) device. The PC controlled the measurement

Figure 1. Measuring device and its design to test the muscular strength of finger grasping; (a) a device with amplifier; (b) a device; (c) a design drawing.

hardware using software written in LabView Version 8.2. Basic subject details (including their name, age and gender) were recorded, and the measurement time and the compression time of the pump were controlled, as shown in Figure 6.

Figure 2. A qualitative judgment of doctors of oriental medicine during a palpation assessment.

Figure 3. The measuring system for the elasticity of hand skin.

Figure 4. Contact component of the designed system.

Table 2. Specifications of the designed system

Component	Specification
Laser sensor (ZX-LD40, OMRON)	Diffusional reflection type, DC Power Detection length: 40 ± 10 mm Resolution: $2 \mu\text{m}$ Wavelength: 650 nm Power: 1 mW
Valve	Three-way solenoid valve
Pump	12 V, DC power
DAQ (USB, National Instrument)	14-bit, 48 kS/s Multifunction I/O and NI-DAQmx

Measurability and Repeatability of the Proposed System

The grasping force for a person using the thumb and index finger—as measured 10 times using the device shown in Figure 1—was ~ 10 N, with an accuracy of 1%. The compression force of the proposed system was designed to be the same as this grasping force. The diameter of the contact probe was 8 mm, and hence the compression pressure was maintained at 2000 kPa.

Figure 7 shows the variables analyzed during the measurements. U_e denotes the immediate deformation induced by the compression and U_v indicates the extra deformation caused by the viscoelasticity of the skin. U_f is the total deformation, which equals $U_e + U_v$. U_r denotes the immediate recovery of the skin upon removal of the compression. In this study, the elasticity was the major characteristic facilitating a constitution diagnosis between the TE and SE subjects. Therefore, the quotient of U_r/U_f was adopted, which corresponded to the elasticity of the skin and is known to be independent of the skin thickness (10–13).

First, the measurability of the proposed system was tested. A commercial rubber plate of known elastic constant was used as the reference material. Having shown a similar value the company suggested our system was acceptable to measure the elastic properties of subjects (14).

Using the proposed system, the measurement repeatability for human skins was tested using a male subject. The elasticity was measured five times per day and the long-term stability was assessed over a period of 1 week. Table 3 lists the measurement results—averaged value of five times measurements per day. The relative repeatability error (RRE)—defined as the ratio between the SD and the average value of the elasticity—was 2% for one measurement and 2.9% for the long-term test.

Optimum Position for the Measurements

Two male subjects—an SE type and a TE type, as determined through pretesting and diagnoses of their constitutions by doctors of oriental medicine in KIOM whose

Figure 5. Jig to hold the hand under analysis. (a) hand placeholder; (b) tilting screw for the vertical direction; (c) tilting screws for the X and Y directions on the horizontal plane.

specialty were Sasang constitution—were selected for a comparison of the elasticity of the human hand for a quantitative diagnosis using the proposed system. The temperature and humidity of the subjects' hands were

Figure 6. (a) The control component; (b) control software.

Figure 7. Variables of the designed system.

made consistent before each measurement, since the mechanical properties of human skin are typically affected by both these factors.

The measurement region was defined as an area comprising the width between the index finger and the ring

finger of the left hand and the height between the three middle fingers of the right hand, which were overlain onto the middle point of the middle finger bone of the left hand (Figure 8). This area was chosen because a doctor of oriental medicine typically palpates this region when performing a constitution diagnosis. Nine contact points in the region were defined to find the optimum contact point for a quantitative palpation measurement: their coordinates were defined from the top of index finger, top of the middle finger, top of the ring finger to bottom of the ring finger (Table 4). ('Top' is the point which is 1 cm below the end of each finger and 'bottom' is the point at which the thumb finger and index finger meet each other as shown in Figure 8)

Figure 9 shows the measurement results for the SE and TE subjects. Each data point was the averaged value of 10 times measurements at each contact point. The elasticity of the skin on the back of the hand was higher for the TE subject than for the SE subject, which is in good agreement with the qualitative measurements of a doctor of oriental medicine and several previous studies (3–5). The measurement repeatability was highest along the top line irrespective of the constitution, which is due to differences in the microstructure of the two skin types, which itself is due to variations in the viscoelasticity along the lower lines and the resulting lower repeatability (15). The measurement repeatability was higher for the TE subject than for the SE subject since the latter had a more uneven skin that resulted in wider dispersion of the laser beam on the skin surface (6).

Table 3. Measurability of the designed system

Time	1	2	3	4	5	6	RRE for one measurement	RRE for long-term measurement
Elasticity	0.669	0.645	0.663	0.669	0.698	0.646	2%	2.9%

Clinical Data between TE and SE

Table 5 lists the clinical data of elasticity from the two constitutions using the proposed system. Each data point was the averaged value of 10 times measurements collected from five TE and three SE male subjects using the top of the ring finger as contact point. Despite the presence of variations in the characteristics of the subjects with each type of constitution (e.g. age, height and weight), the elasticity was uniformly higher in TEs than in SEs with an acceptable level of uncertainty; this tendency is in good agreement with findings reported in the literature (3–5).

Conclusion

In conclusion, a system for measuring the elasticity of human skin was constructed using a compression tube and a laser sensor for the purpose of constitutional diagnoses. The uncertainty in elasticity measurements when using this system was <3%. The elasticity was higher for the TE type than for the SE type with an acceptable level of uncertainty, which is consistent with the findings of previous studies. It can be concluded that the proposed system is suitable as a quantitative constitutional diagnosis method for distinguishing between TE and SE subjects with an acceptable level of uncertainty.

Table 4. RRE of elasticity values for SE and TE types

Position	(T,R)	(M,R)	(B,R)	(T,L)	(M,L)	(B,L)	(T,I)	(M,I)	(B,I)
SE type	3.3	7.2	10.6	3.9	7.8	12.1	3.1	10.7	12.5
TE type	2.0	3.3	12.5	3.5	4.0	5.0	2.6	4.5	5.7

R, ring finger; L, long finger; I, index finger; T, top; M, middle; B, bottom.

Figure 8. Inspection regions of the back of left hand.

Figure 9. Comparison of the elasticity values between an SE type and a TE type along (a) the index finger line, (b) the middle finger line and (c) the ring finger line.

Table 5. Clinical data of elasticity for TE and SE types

	TE					SE				
	1	2	3	4	5	1	2	3	4	5
Elasticity	0.697	0.712	0.688	0.691	0.703	0.536	0.554	0.544	0.556	0.537

Acknowledgement

This work formed part of the project entitled 'Development of Diagnosis Systems using the Senses' (No. 10028423) funded by the Ministry of Knowledge Economy of Korea.

References

1. Shim JP. Sasang constitution in Sasang medicine and constitution principle from eastern and western cultures. *Master's Thesis* 1998, Kyung Hee University (in Korean).
2. Shim EB, Lee SW, Kim JY, Earm YE. Physiome and Sasang constitutional medicine. *J Physiol Sci* 2008;58:433–40.
3. Lee JJM. *Dongeuoseobowoon (Longevity & Life Preservation in Oriental Medicine)*. Seoul: Kyuhg Hee University Press, 1996.
4. Lee DG. *Sasangyolam*. Korea: WoonBooGyo, 1995 (in Korean).
5. Park IS. *Dongusa Sangyogyul*. Korea: SoNaMoo, 1992 (in Korean).
6. Lee SH, Joo JC, Yoon YS, Kim JY. Clinical study of the relations of the refines and the tactile of back skin of the hand to Sasang constitutions depending on sex and age. *Korean J Ori Physio Path* 2005;19:536–43 (in Korean).
7. Lee SH, Choi SM, Kim HG, Kim JY. Clinical study of the relations of the thickness and the stiffness of back skin of the hand to Sasang constitutions depending on sex and age. *Korean J Ori Physio Path* 2005;19:561–7 (in Korean).
8. Lee SH, Han SS, Choi SM, Joo JC, Kim JY. Clinical study of properties of back skin of the hand to Sasang constitutions. *J Ori Physio Path* 2005;19:15–18 (in Korean).
9. Lee JT, Song HW, Park YK, Oh CH. Measurement of finger muscular force to set up a standard for intelligent artificial arms. *Proceeding of APMF 2007* 2008;213–19.
10. Barel AO, Courage W, Clays P. Suction method for measurement of skin mechanical properties: the Cutometer®. In: Serup J, Jemec GBE (eds). *Handbook of Noninvasive Methods and the Skin*. Boca Raton, FL: CRC Press, 1995, 335–40.
11. Dobrev HP. *In vivo* study of skin mechanical properties in patients with systemic sclerosis. *J Am Acad Dermatol* 1999;40:436–42.
12. Dobrev HP. *In vivo* study of skin mechanical properties in scleroderma of buschke. *Acta Derm Venereol* 1998;78:103–6.
13. Serup J, Notrtheved A. Skin elasticity in psoriasis. *In vivo* measurement of tensile distensibility, hysteresis, and resilience distension with a new method. Comparison with skin thickness as measured with high-frequency ultrasound. *J Dermatol* 1985;12:318–24.
14. Song HW. A Project Report for 'Establishment of Standard for New Demand' in KRISS 2008.
15. Langer K. On the anatomy and physiology of the skin (I–IV). *Br J Plast Surg* 1978;31:3–8, 93–106, 185–99, 273–8.

Received March 11, 2009; accepted June 23, 2009