Research Article

Sensitive Determination of Terazosin in Pharmaceutical Formulations and Biological Samples by Ionic-Liquid Microextraction Prior to Spectrofluorimetry

Mohsen Zeeb^{1,2} and Mahdi Sadeghi¹

¹ Agricultural, Medical and Industrial Research School, Nuclear Science and Technology Research Institute,
P.O. Box 31485/498, Karaj, Iran

² Department of Analytical Chemistry, Faculty of Chemistry, University College of Science, University of Tehran,
P.O. Box 14155-6455, Tehran, Iran

Correspondence should be addressed to Mahdi Sadeghi, mahdisadeghi2003@yahoo.com

Received 10 September 2011; Accepted 2 November 2011

Academic Editor: P. Haglund

Copyright © 2012 M. Zeeb and M. Sadeghi. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

An efficient and environmentally friendly sample preparation method based on the application of hydrophobic 1-Hexylpyridinium hexafluorophosphate [Hpy][PF₆] ionic liquid (IL) as a microextraction solvent was proposed to preconcentrate terazosin. The performance of the microextraction method was improved by introducing a common ion of pyridinium IL into the sample solution. Due to the presence of the common ion, the solubility of IL significantly decreased. As a result, the phase separation successfully occurred even at high ionic strength, and the volume of the settled IL-phase was not influenced by variations in the ionic strength (up to 30% w/v). After preconcentration step, the enriched phase was introduced to the spectrofluorimeter for the determination of terazosin. The obtained results revealed that this system did not suffer from the limitations of that in conventional ionic-liquid microextraction. Under optimum experimental conditions, the proposed method provided a limit of detection (LOD) of 0.027 $\mu\text{g L}^{-1}$ and a relative standard deviation (R.S.D.) of 2.4%. The present method was successfully applied to terazosin determination in actual pharmaceutical formulations and biological samples. Considering the large variety of ionic liquids, the proposed microextraction method earns many merits, and will present a wide application in the future.

1. Introduction

Terazosin is a selective potent α_1 adreno-receptor antagonist. It is an effective drug for hypertension and benign prostatic hyperplasia [1, 2]. The needs of monitoring trace levels of terazosin in treated patients with initiate or chronic treatment routine in chronic regimen, especially in renal impairment cases, are necessary. To prevent excess employed dosage, it is critical to develop the sensitive and accurate techniques for its quantitative analysis. Several analytical methods have been developed for the measurement of terazosin in bulk, pharmaceuticals, or biological fluids [3–8].

Sample preparation is a critical step to isolate the analytes of interest from the sample matrix as well as to increase the concentration of analytes prior to their determination, especially when their levels are lower than

the detection limit of analytical techniques. Liquid-liquid extraction (LLE) and solid-phase extraction (SPE) have been widely used as basic, simple, and adaptable methods for sample preconcentration. However, some drawbacks such as the use of large amount of hazardous organic solvents, and the employment of multistep procedures associated with the high risk of analyte losses make LLE and SPE tedious, expensive, and environmentally unfriendly. In recent years, reduction or complete removal of hazardous organic solvents in sample pretreatment techniques has received great attention. For this goal, many microextraction methods such as dispersive liquid-liquid microextraction (DLLME) [9–12], single-drop microextraction (SDME) [13–15], stir bar sorptive extraction [16], and so forth, have been developed to reduce the amount of toxic solvents. In some of these procedures, however, the use of toxic materials like benzene,

toluene, or chloroform as the microextraction solvent is a common practice.

Ionic liquids (ILs) are ionic media resulting from combinations of organic cations and different anions. ILs have a variety of chemical and physical properties such as low vapor pressure, excellent thermal stabilities, adjustable miscibility, moderate solubility of organic compounds and metal ions, and so forth, which make them highly attractive in microextraction methods [17]. ILs have been widely applied in many sample pretreatment procedures as green extraction solvent including ionic liquid-based liquid-liquid microextraction (IL-LLME) [18, 19], cold induced aggregation microextraction (CIAME) [20], ionic liquid-based single-drop microextraction (IL-SDME) [21, 22], ionic liquid-based dispersive liquid-liquid microextraction (IL-DLLME) [23, 24], in situ solvent formation microextraction (ISFME) [25], temperature-controlled ionic liquid dispersive liquid phase microextraction (TILDLM) [26], and so forth.

The use of separation and preconcentration steps based on IL-based microextraction methods offers several advantages such as high recovery, no need of hazardous extraction solvent, simplicity of operation, and compatibility with many analytical techniques. Despite the many advantages of using ionic liquid-based microextraction techniques, under high salt condition, the fine droplets of the extraction phase can not be settled due to increase in sample solution density. The solubility of ionic liquids increases as the salt amount of sample increases. As a result, the volume of the settled phase depends significantly on the ionic strength of the sample solution. On the other hand, accuracy and performance of the extraction method are significantly influenced by variations in the ionic strength. In this work, in order to solve these problems and obtain higher accuracy, a novel, simple, and powerful ionic-liquid microextraction method which put very few dangers to environment was developed. In the proposed approach, a hydrophobic pyridinium IL was added to the aqueous sample solution containing one of the IL's ions as a common ion to act as the extraction phase. Hydrophobic 1-hexylpyridinium hexafluorophosphate [Hpy][PF₆] ionic liquid and NaPF₆ were used as the extraction phase and common ion source, respectively. Due to the presence of the common ion in the sample, the solubility of IL decreases, even at high ionic strength. Therefore, the volume of the enriched phase is independent from the ionic strength. In this work, [Hpy][PF₆] was employed as the extractant solvent due to some physicochemical properties such as high hydrophobicity, water immiscibility, and suitable density. The present methodology does not need heating, antisticking agent, long equilibration time, and cooling before or after centrifugation prior to analysis. To the best of our knowledge, up until now, no method based on the microextraction with ILs and spectrofluorimetric detection has been proposed for preconcentration, extraction, and determination of drugs in their trace levels. Spectrofluorimetric method was applied due to ease, proper selectivity, sensitivity, wide dynamic range, and low cost of operation. The goal of this paper is to study the applicability of the proposed microextraction method followed by spectrofluorimetry to determine terazosin as a model compound in pharmaceutical and biological

samples. Furthermore, the obtained results reveal that the proposed methodology is simple and practical for the routine determination of terazosin in aqueous media.

2. Experimental

2.1. Instrumentation. Fluorescence measurements were performed using a Perkin-Elmer LS 50 spectrofluorimeter equipped with xenon discharge lamp, and quartz microcell with a volume of 100 μ L. A centrifuge from Hettich (Tuttlingen, Germany) was utilized to accelerate the phase separation process. An adjustable sampler (10–100 μ L) was prepared from Eppendorf (Hamburg, Germany). The pH-meter model 692 (Herisau, Switzerland) supplied with a glass-combined electrode and universal pH indicator (pH 0–14) from Merck (Darmstadt, Germany) were used for the pH measurements.

2.2. Reagents and Materials. All chemicals used were of analytical reagent grade and all solutions were prepared with ultrapure water. 1-Hexylpyridinium hexafluorophosphate [Hpy][PF₆] (97%) (Acros organics, Belgium) was applied as an extractant solvent. Acetone, acetonitrile, methanol, ethanol and sodium hexafluorophosphate (NaPF₆) were purchased from Merck (Darmstadt, Germany). Stock solution of terazosin at concentration of 1000 mg L⁻¹ was prepared by dissolving the required amount of pure drug in ultrapure water, and working standard solutions were obtained by serial dilutions of this stock solution. [Hpy][PF₆] IL is not liquid at room temperature (melting point: 45°C). Therefore, this IL was dissolved in acetonitrile to obtain a working solution of 75 mg mL⁻¹. A solution of 250 mg mL⁻¹ NaPF₆ was obtained by dissolving appropriate amount of this salt in ultrapure water. 0.5 mol L⁻¹ of sodium hydroxide and concentrated hydrochloric acid were used for adjusting the pH value of sample solutions. Terazosin tablets (labeled as containing 2 and 5 mg terazosin per tablet) were obtained from Daroupakhsh (Iran, Tehtan).

2.3. General Analytical Procedure. Aliquots of 10.0 mL sample or standard solution (pH 10) containing terazosin in the range of 0.1 to 115 μ g L⁻¹ was placed into a glass test tube with conic bottom, 1.4 mL of NaPF₆ (250 mg mL⁻¹) was added. Then, 150 mg of [Hpy][PF₆] (dissolved in acetonitrile) as the extraction solvent was added to the sample solution. After shaking, a cloudy condition was immediately formed at room temperature. Due to the formation a large contact area between extraction solvent and analyte, the uncharged form of terazosin was rapidly trapped into IL-phase. In order to accelerate phase separation, the resultant solution was centrifuged for 5 min at 4000 rpm. The upper aqueous solution was then removed by a pipette. Finally, in order to reduce the viscosity of IL-phase and facilitate sample handling prior to analysis, the settled phase (about 16 μ L) in the test tube was diluted to 250 μ L by adding appropriate amount of ethanol. 100 μ L of the diluted settled phase was removed by a sampler and transferred to the microcell of the fluorimeter. The fluorescence intensity was measured

FIGURE 1: (a) Emission spectra obtained for 8 standard solutions of terazosin with different concentrations. (b) Emission spectrum of terazosin (75 $\mu\text{g L}^{-1}$) treated the same as previously described in the general analytical procedure (a), and emission spectrum of reagent blank in [Hpy][PF₆] ionic liquid (b). Inset: chemical structure of [Hpy][PF₆] ionic liquid. Experimental conditions were as indicated in the Table 1.

at 376 nm with the excitation wavelength set at 330 nm. A reagent blank was prepared using a similar procedure but without adding terazosin.

2.4. Analysis of Tablets. Five terazosin tablets, labeled as containing 2 mg terazosin each, were weighed and the average mass per tablet was determined. An amount of the powder equivalent to 2 mg of terazosin was accurately weighed and dissolved in 50 mL ultrapure water. The solution was sonicated for 5 min and filtered into a 100 mL volumetric flask. Further dilutions were made up with ultrapure water to achieve a final concentration of 20 $\mu\text{g mL}^{-1}$. In addition, appropriate dilution was performed for terazosin analysis.

2.5. Analysis of Spiked Human Plasma. Human plasma samples (1.0 mL) were spiked with terazosin solutions, deproteinized by adding 4 mL acetonitrile, and centrifuged at 4000 rpm for 15 min. Then, 2.0 mL of the clear supernatant was diluted to 50 mL and subjected to the proposed method.

2.6. Analysis of Spiked Human Urine. Fresh matutinal human urine (10 mL) was transferred into graduated centrifuge tubes. These solutions were centrifuged for 5 min at 4000 rpm. Then, aliquots of 2 mL from clear supernatant were put in new centrifuge tubes and spiked with different amounts of terazosin (2 to 100 $\mu\text{g L}^{-1}$); the general analytical procedure was followed.

3. Results and Discussion

In this work, a practical and simple ionic-liquid microextraction method was combined with spectrofluorimetry for pre-concentration and determination of terazosin. Hydrophobic [Hpy][PF₆] and NaPF₆ were applied as the extraction phase and common ion source, respectively. In the proposed microextraction method, the volume of the enriched phase was independent from the ionic strength of the sample solution (up to 30% w/v) which as an advantage significantly improved the performance of the extraction method. The influence of different factors affecting extraction conditions were studied and optimized to obtain a compromise between sensitivity, simplicity, and reproducibility.

3.1. Spectral Characteristics of Terazosin and [Hpy][PF₆]. The intense fluorescence of terazosin is due to its cyclic conjugated structure which benefits of having π -electron system. The emission spectra of terazosin were recorded as described in the general analytical procedure (see Figure 1(a)). As it can be ascertained, the emission peak of terazosin is at 376 nm, while its excitation peak is at 330 nm. In order to study the effect of the extraction phase on fluorescence spectrum of the analyte of interest. For this goal, fluorescence spectrum of the reagent blank was evaluated. As it can be seen in Figure 1(b), the emission of [Hpy][PF₆] has no notable effect on the determination of terazosin. Thus, the aforementioned wavelengths were selected for quantitative analysis of terazosin.

FIGURE 2: Influence of amount of [Hpy][PF₆] on the fluorescence intensities. Experimental conditions: terazosin concentration 75 $\mu\text{g L}^{-1}$; pH 10; NaPF₆ 350 mg.

FIGURE 3: Influence of NaPF₆ on the analytical signals obtained for terazosin. Experimental conditions: terazosin concentration 75 $\mu\text{g L}^{-1}$; pH 10; [Hpy][PF₆] 150 mg.

3.2. Influence of IL Amount. The amount of [Hpy][PF₆] is a critical parameter which can affect the analyte recoveries. Thus, the microextraction procedure was carefully evaluated in order to achieve a compromise between the amount of IL and the analytical responses. The Effect of [Hpy][PF₆] amount on the analytical responses was studied within the range of 25–400 mg. Figure 2 shows the variation of the fluorescence signal versus the amount of [Hpy][PF₆]. As it can be ascertained, the signal intensity increases as the amount of [Hpy][PF₆] increases and remains nearly constant from 125 mg. No significant variations were obtained on the microextraction efficiency for higher amounts of IL. Thus, in order to achieve sensitive analytical response, 150 mg of IL was selected as an optimum value.

3.3. Influence of Common Ion. The effect of NaPF₆ amount as a common ion source on the fluorescence signal was studied within the range of 0–400 mg (see Figure 3). Due to the common ion influence, an increase in the amount of NaPF₆ causes a notable decrease in the solubility of [Hpy][PF₆]. As a result, the extraction performance and the analytical intensity increase. As it can be seen in Figure 3, the fluorescence intensity increases with increasing the amount of NaPF₆ up to 325 mg and then remains nearly stable. However, 350 mg of NaPF₆ was chosen through the rest of the work.

3.4. Influence of Diluting Agent. Due to the high viscosity of the IL, it must be conditioned with a diluting agent prior to its introduction to the spectrofluorimeter. Several diluting agents including methanol, ethanol, acetone, and acetonitrile were evaluated in order to choose a diluting agent which can dissolve the IL-phase completely and provide the best sensitivity. Type of the organic solvent had no significant effect on the analytical response. Finally, due to less toxicity of ethanol, it was applied as diluting agent in all experiments.

3.5. Influence of pH. The pH of the sample is an important factor which affects the chemical structure of terazosin. The effect of pH on the extraction of terazosin from aqueous

FIGURE 4: Influence of pH on the fluorescence intensities. Experimental conditions: terazosin concentration 75 $\mu\text{g L}^{-1}$; [Hpy][PF₆] 150 mg; NaPF₆ 350 mg. Inset: chemical structure of terazosin. Acid base equilibrium involved at working pH. (a) Protonated form of this terazosin; (b) dissociated form of terazosin.

samples was studied over the range of 5–12 using HCl and NaOH. For ionizable organic molecules, maximum extraction efficiency is obtained at pH values where the uncharged form of the analyte overcomes, and therefore, the analyte of interest is favored to be partitioned into the hydrophobic IL-phase. The results illustrated in Figure 4 reveal that analytical signals, obtained for terazosin, depend on pH. The best conditions of terazosin microextraction were achieved at alkaline pH values due to dissociation of the aromatic amine group, which acts as a weak base (see Figure 4) ($\text{pK}_b = 6.9$). Based on the results obtained, pH 10 was selected as the optimum value.

3.6. Influence of Equilibration Temperature and Extraction Time. In order to obtain complete extraction of terazosin and easy phase separation, the effect of equilibration temperature was studied from 5 to 40°C. Based on the results obtained in this study, the temperature had no significant and benefic effect on the extraction efficiency. Therefore, room temperature as an equilibration temperature was

used through the rest of the work. Microextraction time is one of the important factors affecting the extraction efficiency, especially in microextraction methods such as SPME and LPME. The dependence of extraction efficiency upon extraction time was studied from 5 sec to 30 min. The results obtained in this experiment revealed that signal variations versus extraction time were not significant. It was demonstrated that after formation of the cloudy condition, the surface area between extraction phase and aqueous media was infinitely large. Thus, the transfer of the target analyte from aqueous media to IL-phase was very fast. In order to keep extraction time as short as possible, the cloudy solution was centrifuged immediately after the preparation.

3.7. Influence of Centrifugation Condition. The effect of centrifugation rate on the fluorescence intensity was evaluated from 1000 to 6000 rpm. The obtained results showed that over 3500 rpm the IL-phase was completely transferred to the bottom of the test tube, and the analytical response remained constant. Thus, 4000 rpm was selected for all experiments. At the optimum rate, the effect of centrifugation time upon analytical responses was investigated within the range of 1–15 min. Over 4 min, the fluorescence intensity remained stable indicating entire transfer of IL-phase to the bottom of the centrifuge tube. Therefore, a centrifugation time of 5 min was chosen as the optimum value.

3.8. Influence of Salt Concentration. Influence of salt concentration on the extraction performance was studied by adding various amounts of NaNO_3 from 0 to 40% (w/v) while other experimental conditions were kept constant. In the proposed microextraction method, due to the presence of PF_6^- as a common ion in the sample solution, the volume of the enriched phase was independent from the ionic strength, and phase separation occurred up to 30% w/v of NaNO_3 . Table 1 summarizes both instrumental and experimental conditions selected.

3.9. Influence of Interfering Substances. The effect of different chemical species on the determination of target analyte was studied using solutions containing $75 \mu\text{g L}^{-1}$ of terazosin, and adding 100 mg L^{-1} of the possible interfering species. The tolerance limit of each interfering substance was taken into account as the largest amount yields an error in the determination of the target analyte which does not exceed 5%. No interference was observed from commonly interfering species such as Na^+ , NH_4^+ , Ca^{2+} , Zn^{2+} , Mg^{2+} , Cl^- , PO_4^{3-} , SO_4^{2-} , starch, glucose, lactose, fructose, sucrose, ascorbic acid, citric acid, dye species (as yellow quinoline), urea, and saccharin. The obtained results revealed the excellent selectivity of the proposed methodology in detecting the studied drug in dosage form and biological samples.

4. Application

4.1. Analytical Performance. Under the optimum conditions, calibration graph was achieved by analyzing 10.0 mL of standard solutions containing known amounts of terazosin.

TABLE 1: The experimental conditions for terazosin determination.

Microextraction parameters	Studied range	Selected condition
Amount of NaPF_6 (mg)	0–400	350
Amount of $[\text{Hpy}][\text{BF}_4]$ (mg)	25–400	150
pH	5–12	10
Equilibration temperature ($^\circ\text{C}$)	5–40	Room temperature
Centrifugation time (min)	1–15	5
Ionic strength (% w/v)	0–40	NE^a
Instrumental parameters		
Excitation wavelength (nm)	200–350	330
Emission wavelength (nm)	350–700	376
Excitation and emission slit widths (nm)	5–20	10

^aNo effect.

TABLE 2: Determination of terazosin in tablets by the proposed methodology and by a reported voltammetric method [8].

Claimed (mg/tablet)	Proposed methodology (mg) ^a	Reported method (mg) ^a	Error (%) ^b	Error (%) ^c
2	1.95 (± 0.08)	2.07 (± 0.10)	−2.5	−5.8
5	5.1 (± 0.21)	5.27 (± 0.37)	+2.0	−3.2

^aValues in parenthesis show the standard deviation based on three replicates.

^bError against the declared value.

^cError against the reported method.

The settled ionic phase was diluted to $250 \mu\text{L}$ with ethanol and the fluorescence intensity was recorded. Therefore, a pre-concentration factor of 40 was achieved. The enhancement factor calculated as the slope ratio of calibration graph after and before extraction was about 33. The linear concentration range was from 0.1 to $115 \mu\text{g L}^{-1}$ of terazosin with linear regression equation as

$$I = 51.33 + 7.5608 C, \quad (1)$$

where I is the fluorescence intensity at $376 \pm 3 \text{ nm}$ and C is the terazosin concentration in $\mu\text{g L}^{-1}$. The relative standard deviation (R.S.D.) obtained for the determination of $75 \mu\text{g L}^{-1}$ of terazosin was 2.4% ($n = 5$). The limit of detection (LOD), calculated as three times the standard deviation of the measurement of blanks divided by the slope of the calibration curve, was found to be $0.027 \mu\text{g L}^{-1}$.

4.2. Analysis of Terazosin in Pharmaceutical Formulations. In order to show the validity of the proposed method, it was applied for terazosin determination in commercial tablets. Three replicate determinations were performed, and satisfactory results were achieved. Table 2 shows the results obtained by applying the present method and those obtained by a reported voltammetric method [8]. Until date no official method has been reported for terazosin determination. The results show the applicability of the proposed methodology and its accuracy for quantitative analysis of terazosin in this type of samples.

TABLE 3: Determination of terazosin in spiked urine and spiked plasma by present work.

Drug	Amount added ($\mu\text{g L}^{-1}$)	Spiked urine		Amount added ($\mu\text{g L}^{-1}$)	Spiked plasma	
		Amount found ($\mu\text{g L}^{-1}$) $\pm$ S.D. ^a	Recovery (%)		Amount found ($\mu\text{g L}^{-1}$) $\pm$ S.D. ^a	Recovery (%)
Terazosin	2	1.83 $\pm$ 0.09	91.5	2	1.73 $\pm$ 0.11	86.5
	5	4.81 $\pm$ 0.28	96.2	5	4.82 $\pm$ 0.33	96.4
	10	9.40 $\pm$ 0.30	94.0	10	9.60 $\pm$ 0.51	96.0
	100	105.81 $\pm$ 4.11	105.8	100	93.10 $\pm$ 4.88	93.1

^a Average of three independent measurements.

TABLE 4: Comparison of the proposed approach with other reported methods for determination of terazosin.

Method	Sample	LOD ($\mu\text{g L}^{-1}$)	R.S.D. (%)	LR ($\mu\text{g L}^{-1}$)	References
Spectrofluorimetry	Serum, urine	210	2.5	Up to 3486.9	[4]
X-ray fluorescence spectrometry	Drug	732	—	732–843000	[5]
HPLC with fluorescence detection	Plasma	0.25	<7	Up to 100	[6]
HPLC with electrospray ionization mass spectrometry detection	Plasma	0.0625	—	—	[7]
Square-wave adsorptive cathodic stripping voltammetry	Plasma, drug	0.0058	<1.1	0.39–11.62	[8]
ionic-liquid microextraction spectrofluorimetry	Urine, plasma, drug	0.027	2.4	0.1–115	This work

LOD: limit of detection, LR: linear range, R.S.D.: relative standard deviation.

4.3. Analysis of Terazosin in Spiked Human Plasma and Urine. The accuracy of the present method was evaluated by determination of terazosin in spiked human urine and spiked human plasma. The recovery of the studied drug was investigated at four concentration levels. The obtained results are shown in Table 3. As can be seen, calculated amounts of recoveries varied between 91.5–105.8% and 86.5–96.0% for human urine and human plasma, respectively, indicating both accuracy and precision.

4.4. Comparison with Other Reported Methods. Determination of terazosin by the proposed methodology was compared with other reported techniques. The results are shown in Table 4. As it can be seen, compared to the previously reported methods, the present method has a relatively low LOD, wide dynamic range, and excellent reproducibility. The method developed in this work is proposed as a suitable alternative to expensive instrumental methods for trace determination of terazosin in pharmaceutical and biological samples. These results reveal that the proposed approach is a very sensitive, low-cost, rapid, environmentally friendly and accurate technique that can be used for terazosin determination in routine analytical laboratories.

5. Conclusion

A novel and efficient mode of ionic-liquid microextraction was combined with spectrofluorimetric detection to preconcentrate and determine trace levels of terazosin. Hydrophobic [Hpy][PF₆] ionic liquid was applied as a green extraction solvent and an alternative to hazardous organic solvents. Due to the presence of one of the IL's ions as a

common ion in the sample solution, the performance of the microextraction method was not influenced by variations in the ionic strength of the aqueous solution (up to 30%). The method developed in this work was demonstrated to be simple, rapid, practical, inexpensive, and environmentally friendly for preconcentration and separation of the trace analytes. In addition, the present method revealed to be a practical tool for routine quality control of drugs in pharmaceutical and biological samples with low operation cost and simplicity of instrumentation.

Acknowledgments

Support of this investigation by the Research Council of Agricultural, Medical and Industrial Research School through grant is gratefully acknowledged. The authors also thank Miss Shakila Savadkouhi for help to check voltammetric procedure as a reference method.

References

- [1] R. Mudiya and A. Ahmed, "Effect of terazosin on clinical benign prostatic hyperplasia in older adults," *Journal of the American Geriatrics Society*, vol. 51, no. 3, pp. 424–426, 2003.
- [2] H. Lopor, K. Jones, and W. Williford, "The mechanism of adverse events associated with terazosin: An analysis of the Veterans Affairs cooperative study," *Journal of Urology*, vol. 163, no. 4, pp. 1134–1137, 2000.
- [3] H. H. Abdine, F. A. El-Yazbi, S. M. Blaih, and R. A. Shaalan, "Spectrophotometric and spectrofluorimetric methods for the determination of terazosin in dosage forms," *Spectroscopy Letters*, vol. 31, no. 5, pp. 969–980, 1998.

- [4] C. Q. Jiang, M. X. Gao, and J. X. He, "Study of the interaction between terazosin and serum albumin -Synchronous fluorescence determination of terazosin," *Analytica Chimica Acta*, vol. 452, no. 2, pp. 185–189, 2002.
- [5] R. A. Silva and A. N. Masi, "Sensitive determination of terazosin by x-ray fluorescence spectrometry based on the formation of ion-pair associates with zinc thiocyanate," *X-Ray Spectrometry*, vol. 36, no. 4, pp. 279–283, 2007.
- [6] P. Y. Cheah, K. H. Yuen, and M. L. Liong, "Improved high-performance liquid chromatographic analysis of terazosin in human plasma," *Journal of Chromatography B*, vol. 745, no. 2, pp. 439–443, 2000.
- [7] A. P. Zavitsanos and T. Alebic-Kolbah, "Enantioselective determination of terazosin in human plasma by normal phase high-performance liquid chromatography-electrospray mass spectrometry," *Journal of Chromatography A*, vol. 794, no. 1–2, pp. 45–56, 1998.
- [8] M. M. Ghoneim, M. A. El Ries, E. Hammam, and A. M. Beltagi, "A validated stripping voltammetric procedure for quantification of the anti-hypertensive and benign prostatic hyperplasia drug terazosin in tablets and human serum," *Talanta*, vol. 64, no. 3, pp. 703–710, 2004.
- [9] M. Zeeb, M. R. Ganjali, and P. Norouzi, "Dispersive liquid-liquid microextraction followed by spectrofluorimetry as a simple and accurate technique for determination of thiamine (vitamin B1)," *Microchimica Acta*, vol. 168, no. 3, pp. 317–324, 2010.
- [10] P. Hemmatkhah, A. Bidari, S. Jafarvand, M. R. M. Hosseini, and Y. Assadi, "Speciation of chromium in water samples using dispersive liquid-liquid microextraction and flame atomic absorption spectrometry," *Microchimica Acta*, vol. 166, no. 1–2, pp. 69–75, 2009.
- [11] S. Jafarvand, A. Bidari, P. Hemmatkhah, M. R. M. Hosseini, and Y. Assadi, "Dispersive liquid-liquid microextraction of silver prior to determination by microsample introduction-flame atomic absorption spectrometry," *Analytical Letters*, vol. 42, no. 14, pp. 2214–2231, 2009.
- [12] F. Rezaei, A. Bidari, A. P. Birjandi, M. R. Milani Hosseini, and Y. Assadi, "Development of a dispersive liquid-liquid microextraction method for the determination of polychlorinated biphenyls in water," *Journal of Hazardous Materials*, vol. 158, no. 2–3, pp. 621–627, 2008.
- [13] A. Jain, A. K. K. V. Pillai, N. Sharma, and K. K. Verma, "Headspace single-drop microextraction and cuvetteless microspectrophotometry for the selective determination of free and total cyanide involving reaction with ninhydrin," *Talanta*, vol. 82, no. 2, pp. 758–765, 2010.
- [14] A. Garbi, V. Sakkas, Y. C. Fiamegos, C. D. Stalikas, and T. Albanis, "Sensitive determination of pesticides residues in wine samples with the aid of single-drop microextraction and response surface methodology," *Talanta*, vol. 82, no. 4, pp. 1286–1291, 2010.
- [15] Y. He and H. K. Lee, "Liquid-phase microextraction in a single drop of organic solvent by using a conventional microsyringe," *Analytical Chemistry*, vol. 69, no. 22, pp. 4634–4640, 1997.
- [16] A. Sanchez-Ortega, N. Unceta, A. Gómez-Caballero et al., "Sensitive determination of triazines in underground waters using stir bar sorptive extraction directly coupled to automated thermal desorption and gas chromatography-mass spectrometry," *Analytica Chimica Acta*, vol. 641, no. 1–2, pp. 110–116, 2009.
- [17] S. Pandey, "Analytical applications of room-temperature ionic liquids: a review of recent efforts," *Analytica Chimica Acta*, vol. 556, no. 1, pp. 38–45, 2006.
- [18] N. Hirayama, M. Deguchi, H. Kawasumi, and T. Honjo, "Use of 1-alkyl-3-methylimidazolium hexafluorophosphate room temperature ionic liquids as chelate extraction solvent with 4,4,4-trifluoro-1-(2-thienyl)- 1,3-butanedione," *Talanta*, vol. 65, no. 1, pp. 255–260, 2005.
- [19] P. Berton, E. M. Martinis, L. D. Martinez, and R. G. Wuilloud, "Room temperature ionic liquid-based microextraction for vanadium species separation and determination in water samples by electrothermal atomic absorption spectrometry," *Analytica Chimica Acta*, vol. 640, no. 1–2, pp. 40–46, 2009.
- [20] M. Gharehbaghi, F. Shemirani, and M. D. Farahani, "Cold-induced aggregation microextraction based on ionic liquids and fiber optic-linear array detection spectrophotometry of cobalt in water samples," *Journal of Hazardous Materials*, vol. 165, no. 1–3, pp. 1049–1055, 2009.
- [21] L. Vidal, A. Chisvert, A. Canals, and A. Salvador, "Ionic liquid-based single-drop microextraction followed by liquid chromatography-ultraviolet spectrophotometry detection to determine typical UV filters in surface water samples," *Talanta*, vol. 81, no. 1–2, pp. 549–555, 2010.
- [22] E. Aguilera-Herrador, R. Lucena, S. Cárdenas, and M. Valcárcel, "Determination of trihalomethanes in waters by ionic liquid-based single drop microextraction/gas chromatographic/mass spectrometry," *Journal of Chromatography A*, vol. 1209, no. 1–2, pp. 76–82, 2008.
- [23] M. Gharehbaghi, F. Shemirani, and M. Baghdadi, "Dispersive liquid-liquid microextraction based on ionic liquid and spectrophotometric determination of mercury in water samples," *International Journal of Environmental Analytical Chemistry*, vol. 89, no. 1, pp. 21–33, 2009.
- [24] H. Abdolmohammad-Zadeh and G. H. Sadeghi, "Combination of ionic liquid-based dispersive liquid-liquid microextraction with stopped-flow spectrofluorimetry for the pre-concentration and determination of aluminum in natural waters, fruit juice and food samples," *Talanta*, vol. 81, no. 3, pp. 778–785, 2010.
- [25] M. Zeeb, M. R. Ganjali, P. Norouzi, and M. R. Kalaei, "Separation and preconcentration system based on microextraction with ionic liquid for determination of copper in water and food samples by stopped-flow injection spectrofluorimetry," *Food and Chemical Toxicology*, vol. 49, no. 5, pp. 1086–1091, 2011.
- [26] Q. Zhou, H. Bai, G. Xie, and J. Xiao, "Temperature-controlled ionic liquid dispersive liquid phase micro-extraction," *Journal of Chromatography A*, vol. 1177, no. 1, pp. 43–49, 2008.