with extreme values (at low $p\text{CO}_2$ and high $p\text{O}_2$). The findings also suggest that fully mechanised systems are neither more accurate nor more precise under all circumstances. Mechanised equipment lacked replicability in $p\text{O}_2$ measurements and its use seemed to be associated with a certain degree of contamination.

An important conclusion of this study is that two-point calibration is to be preferred for all measurements, and that equipment involving single-point calibration (non-mechanised) for pH measurement, mechanised for measurements of $p\text{CO}_2$ and $p\text{O}_2$) is inadequate.

Undoubtedly, tonometry is still the best method of controlling blood gas analysis, but it is beyond the reach of many laboratories. For this reason, the results of the present survey have important fractional implications.

REFERENCES

- [1] Minty, B.D. and Nunn, J.F., *Annals of Clinical Biochemistry*, 1977, 14, 245.
- [2] Evans, J.R., *Annals of Clinical Biochemistry*, 1978, 15, 168.
- [3] Komjathy, Z.L., Mathies, J.C., Parker, J.A. and Schreiber, H.A., *Clinical Chemistry*, 1976, 22, 1399.
- [4] Maas, A.H., Veefkind, A.H., Van-den-Camp, R.A., Teunissen, A.J. and Winkers, E.K., *Clinical Chemistry*, 1977, 23, 1718.
- [5] Noonan D.C. and Komjathy Z.L., *Clinical Chemistry*, 1976, 22, 1818.
- [6] Schwartz, O., *Methodes statistiques a l'usage des Medecins et des Biologistes*, Editions Flammarion II, 1969, 173.
- [7] Ksorensen, Radiometer A.J. Copenhagen, Denmark Quality control and standards National Bureau of Standards, Special Publication 450, Blood pH gases and electrolytes U.S. Dept of Commerce, Washington D.C. June 1977 285.

Erratum

The Journal of Automatic Chemistry, October 1979, 1, 5, 273-281.

An evaluation of the Kodak Ektachem system for the determination of glucose and urea

R. Haeckel, O. Sonntag and K. Pethy.

The authors have asked that the following errors in the above paper be brought to the attention of readers.

On page 276 in the second paragraph under the heading "Accuracy" reference is twice made to "Figures 4 and 7". In both cases this should read "Figures 4 to 7".

In Figures 11 and 12 on page 280, the line and reference [14] should have been deleted.

In Table 10 the methotrexate concentration should have been given in μmol .

Some additional information has been provided for Table 9 and for completeness, we have reproduced the entire table.

Table 9 Refound values of glucose and urea in pool sera to which various components were added. In the absence of exogenic compounds mean values ($\bar{x}$) and standard deviations (s) were calculated from 15 determinations. Results of analysis which are outside the range of confidence ($\bar{x} \pm 3s$), are marked with an asterisk. The range of confidence is for glucose 11.14 – 14.08 mmol/l and for urea 6.23 – 6.68 mmol/l.

Trade name	I.N.N. (1)	concentration mg/l	glucose (mmol/l)	urea (mmol/l)
Amuno	indometacinum	4	13.28	6.47
Butazolidin	phenylbutazonum	12	13.06	6.51
Metalcaptase	D-penicillaminum	36	12.35	6.53
Prolixan	azopropazon-d ihydrat	36	13.45	6.57
Resochin	chloroquinum	5	12.70	6.57
Tanderil	oxyphenbutazonum	12	12.67	6.46
Aponal	doxepinum	6	12.76	6.43
Megaphen	phenothiazinum	20	13.11	6.55
Multum	chlordiazeposidum	1.8	12.26	6.56
Aspirin	acidum acetylosali- cyclicum	100	13.59	6.45
Dolviran	acidum acetylocali- cyclicum, etc.	—	—	—
Novalein	novaminsulfonum	80	12.79	6.45
Benemid	probenecidum	40	13.36	6.42
Uriovac	benzbromaronum	8	13.17	6.44
Zyloric	allopurinolum	18	12.71	6.39
Anglografin	acidum trijodbenzoicum	1300	13.07	6.55
Biligrafin	adipinyltrijodanilidum	200	12.20	6.39
Urografin	acidum trijodbenzoicum	—	—	—
Binotal 500	aminobenzylpenicillinum	180	13.23	6.46
Hostacyclin	tetracyclinum	40	16.99*	6.77*
Paraxin	chloramphenicolum	600	16.01*	6.64
Buscopan	hyoscin-N-butylbrominum	2	13.16	6.59
Cebion	acidum ascorbicum	80	13.24	6.42
Polybion	Vitamin B complex	2.3	12.57	6.39

Table continues overleaf

Table 9 continued

Trade name	I.N.N. ⁽¹⁾	concentration mg/l	glucose (mmol/l)	urea (mmol/l)
Euglycon 5	glibenclamidum	0.3	13.51	6.46
Rastinon	tolbutamidum	40	13.24	6.46
Dulcolax	bisacodylum	0.2	12.58	6.40
Durenat	sulfanilamidopyrimidinum	20	13.75	6.50
Endoxan	cyclophosphamidum	8	13.41	6.42
Methotrexat	acidum methylpteroylglut- aminicum	8	13.30	6.62
Intensain	carbocromenum	18	12.64	6.52
Furadantin	nitrofurantoinum	5	13.93	6.48
Lanicor	digoxinum	0.015	13.03	6.49
Lasix	furosemidum	40	13.13	6.56
Luminal	acidum phenyl- aethylbarbituricum	6	12.74	6.56
Macrodex 6%	dextranum 6%	1800	13.09	6.45
Modenol	thiabutazide, etc.	0.2	12.64	6.56
Presinol	methyl dopa	80	12.35	6.64
Nicobion	nicotinamidum	12	12.15	6.55
Novadral	norfenefrinum	0.48	12.23	6.64
Marcumar	phenprocoumonum	0.36	12.31	6.64
Na-Citrat	Na-citrate	500	12.21	6.64
Liquemin	Na-heparinat	75	12.21	6.68
Na-Fluorid	Na-fluoride	200	17.55*	7.40
Na-Oxalat	Na-oxalate	300	12.34	6.72*
EDTA	Titriplex III	100	12.73	6.61
Aldactone	spiro lactonum	8	11.80	6.51

(1) International designation as recommended by WHO

Notes for Contributors

Presentation of manuscripts

Manuscripts should be typed (double-spaced) on one side of the paper only and with generous margins. The title should be brief and informative avoiding the word "new" and its synonyms. The full list of authors with their affiliations and full address(es) should appear on the title page. On a separate sheet an abstract of no more than 150 words is required. This should succinctly describe the scope of the contribution and highlight significant findings or innovations. It should be written in a style which can easily be translated into French and German.

The Concise Oxford Dictionary and Fowler's *Modern English Usage* (both published by Oxford University Press) should be used as the standard for spelling and grammar. Abbreviations should be limited to those generally recognised, or where a frequently occurring term is abbreviated it should, in the first instance, be explained thus "flow injection analysis (FIA) ..." and the abbreviation used thereafter. Abbreviations, for standard measures and units should follow SI recommendations. There are various publications giving guidance on the use of SI units.

References should be indicated in the text by numerals following the author's name, i.e. Skeggs [6]. On a separate sheet of paper, list all references in numerical order thus: [6] Skeggs, L.T., *American Journal of Clinical Pathology*, 1959, 28, 311.

Note that journal titles are given in full. Where there is more

than one author, the form Foreman et al. should be used in the text but all authors should be named in the list of references. When reference is made to a chapter in a book the reference should take the following form:

[7] Malmstadt, H.V. in "Topics in Automatic Chemistry" Ed. Stockwell P.B. and Foreman J.K. 1978 Horwood, Chichester, pp. 68-70.

Only work which has been published or has been accepted for publication should be cited. Avoid the citation of documents which are subject to restricted circulation, patent literature, unpublished work and personal communications. The latter can be mentioned in the text in parenthesis.

To illustrate a paper line diagrams are preferred to photographs. Photographs should only be used when they significantly add to the discussion. Diagrams, charts and graphs should be carefully drawn in black ink on stout card or heavy quality tracing paper. Most illustrations are reduced for publication; to allow for this originals should be between 16 and 36 cm wide (the depth must not exceed 50 cm). The lettering of diagrams should be sufficiently clear to withstand reduction. Except in the case of proper names, all lettering should be in lower case print. If photographs are used they must be supplied in the form of clear, unmounted, glossy, black and white prints. "Instant" photographs are not normally acceptable. All illustrations must be identified on the reverse showing the figure number and the author's name.

Each illustration should have a fully explanatory caption. Captions should be typed together on a separate sheet of paper; they must *not* be an inseparable part of the illustration.