Correspondence: A method for determining total nitrogen . . .

From Professor A. Fleck

I read with interest the paper by Geiger, Dennis *et al.* in the journal's Volume 9, Number 2, pages 72–76: 'A method for determining total nitrogen in Kjeldahl digestion solution using a centrifugal analyser'.

Without wishing to detract from the value and considerable interest of this application of a centrifugal analyser, there are a few points which it is unfortunate that the authors did not comment on or note in their paper.

The digestion method using a small proportion of a strong oxidizing agent perchloric acid seems to have been taken from a 'continuous flow method', which was not designed for discrete digestion and analysis. There is considerable evidence that satisfactory digestion of even fairly refractory substances can be achieved without the use of perchloric acid (1) and it is now usually recommended that perchloric acid digestion should be avoided if at all possible because of the hazards of explosion. There is also considerable evidence that the use of selenium dioxide as a catalyst may lead to incomplete recovery of nitrogen (see [1]). This is also a hazard with temperatures above 400 °C if that was the temperature of the digest [1].

It seems to be a pity that the authors did not explore the possibility of using the rate of colour development to quantitate the ammonia, preliminary observations of which have been reported by Weichselbaum *et al.* [2].

Finally, the authors give no evidence of the recovery of the analyte from real samples.

Thus, although the authors show that, as might be anticipated, the centrifugal analyser can be satisfactorily used for the determination of ammonia in solution; I would strongly recommend that others wishing to use this procedure for determining ammonia in Kjeldahl digests should carefully check on the recommended conditions for the Kjeldahl digestion before following the details of the authors' procedure.

Department of Chemical Pathology, Charing Cross and Westminster Hospital Medical School, Fulham Palace Road, London W6 8RF.

References

- FLECK, A. Micro determination of nitrogen. *Critical Reviews in Analytical Chemistry* (1974), 141–154.
WEICHSELBAUM, T. E., HAGGARTY, J. C., MARK, H. B. JNR., *Analytical Chemistry*, **41** (1969), 848.

FORTHCOMING PAPERS

The following are amongst the articles to be published in the *Journal of Automatic Chemistry*, Volume 10:

Instrument selection and evaluation
F. L. Mitchell

Automated processing of whole blood samples into microlitre aliquots of plasma
C. A. Burtis et al.

Robots in the clinical laboratory
L. B. Roberts

A novel approach to non-segmented flow analysis. Part 2. A prototype high-performance analyser
D. T. Malcolm-Lawes and C. Pasquini

Automated serum chloride analysis using the Apple computer
P. J. Taylor and R. A. Bouska