Erratum

Erratum to “Development of a Low-Cost SIA-Based Analyser for Water Samples”

Moisés Knochen,¹ Alejandro Caamaño,¹ and Heinkel Bentos²

¹DEC, Analytical Chemistry, School of Chemistry, University of the Republic (UdelaR), Avenue Gral. Flores 2124, 11800 Montevideo, Uruguay

²DETEMA, Instruments Workshop, School of Chemistry, University of the Republic (UdelaR), Avenue Gral. Flores 2124, 11800 Montevideo, Uruguay

Correspondence should be addressed to Moisés Knochen, mknochen@fq.edu.uy

Received 30 March 2011; Accepted 24 May 2011

Copyright © 2011 Moisés Knochen et al. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

We are hereby presenting an erratum to our paper “Development of a Low-Cost SIA-Based Analyser for Water Samples.” We have noticed an error in Table 1: units for concentrations

were erroneously labelled as $\mu\text{g L}^{-1}$, when in fact they should be labelled as mg L^{-1} . The corrected version of Table 1 is presented.

TABLE 1: Analytical figures of merit for the automated methods developed for chloride, nitrite, and nitrate in water samples. LD and LQ: detection (3σ) and quantification (10σ) limits. Precision was measured as repeatability ($n = 10$).

Analyte/figure of merit	LD (mg L^{-1})	LQ (mg L^{-1})	Linear range (mg L^{-1})	Precision sr (%)	Difference with reference method (%)	Sampling rate (hour^{-1})
Chloride	0.6	1.90	1.9–54	1.4	2.0	60
Nitrite	0.02	0.076	0.11–3.8	2.9	–2.8	60
Nitrate	0.5	1.6	1.6–54	2.4	–0.3	20