Corrigendum

Corrigendum to “Hepatic Iron Quantification on 3 Tesla (3 T) Magnetic Resonance (MR): Technical Challenges and Solutions”

Muhammad Anwar ¹, John Wood,^{2,3} Deepa Manwani,⁴ Benjamin Taragin ¹,
Suzette O. Oyeku,⁵ and Qi Peng¹

¹Department of Pediatric Radiology, Children’s Hospital at Montefiore, Bronx, NY, USA

²Department of Pediatric Radiology, Children’s Hospital Los Angeles, Los Angeles, CA, USA

³Division of Pediatric Cardiology, Children’s Hospital Los Angeles, Los Angeles, CA, USA

⁴Department of Pediatrics, Division of Pediatric Hematology/Oncology, Children’s Hospital at Montefiore, Bronx, NY, USA

⁵Department of Pediatrics, Division of General Pediatrics, Children’s Hospital at Montefiore, Bronx, NY, USA

Correspondence should be addressed to Muhammad Anwar; masaeed@excite.com

Received 1 March 2018; Accepted 4 March 2018; Published 3 June 2018

Copyright © 2018 Muhammad Anwar et al. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

In the article titled “Hepatic Iron Quantification on 3 Tesla (3 T) Magnetic Resonance (MR): Technical Challenges and Solutions” [1], there was an error in equation (2), where the R_2^* values for 1.5 T and 3 T are reversed. The corrected equation is as follows:

$$R_2^* (1.5 \text{ T}, C_{\text{Fe}}) = \frac{(R_2^* (3 \text{ T}, C_{\text{Fe}}) + 11)}{2.0}. \quad (2)$$

References

- [1] M. Anwar, J. Wood, D. Manwani, B. Taragin, S. O. Oyeku, and Q. Peng, “Hepatic iron quantification on 3 tesla (3 T) magnetic resonance (MR): technical challenges and solutions,” *Radiology Research and Practice*, vol. 2013, Article ID 628150, 7 pages, 2013.