Research Article

Direct-Contact Low-Frequency Ultrasound and Pulse Lavage Eradicates Biofilms on Implant Materials In Vitro

Xiaoqing Wu,¹ Xuanren Shi,² Mingcong Chen,¹ Xiaoyong Chen,¹ Chi Zhang,³ Xiaonan Zhang,⁴ and Jinyu Zhu^{1,5}

¹Department of Orthopaedics and Traumatology, Shenzhen University General Hospital, Shenzhen 518055, China

²Department of Hematology and Oncology, Shenzhen University General Hospital, Shenzhen 518055, China

³Department of Laboratory Medicine, Shenzhen University General Hospital, Shenzhen 518055, China

⁴Guangdong Key Laboratory for Biomedical Measurements and Ultrasound Imaging, School of Biomedical Engineering, Health Science Center, Shenzhen University, Shenzhen 518060, China

⁵Department of Orthopaedics, Southern University of Science and Technology Hospital, Shenzhen 518055, China

Correspondence should be addressed to Jinyu Zhu; zhujinyu@hotmail.com

Received 31 May 2021; Revised 25 July 2021; Accepted 12 August 2021; Published 31 August 2021

Academic Editor: Isabella Cavalcanti

Copyright © 2021 Xiaoqing Wu et al. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Pulse lavage (PL) debridement and ultrasound are both known to be the treatment of biofilm-related periprosthetic joint infection (PJI). However, the efficacy of these in combination is unknown in eradicating biofilm from the orthopaedic metal implant surface. This study was conducted to understand the efficacy of PL and ultrasound in combination in eradicating bacterial biofilms on titanium alloy in vitro. Biofilms of *Staphylococcus aureus* strains were grown on titanium alloy coupons for 24 h. Then, the coupons were taken to each treatment group: (i) debrided with PL, (ii) exposed to ultrasound, or (iii) exposed to both. An untreated biofilm was set as a control group. Viable plate count and confocal microscopy using live/dead staining was used to measure the amount of biofilm. Viable plate count showed an approximate two-log reduction in CFU/cm² in PL alone, from an initial cell count on the metal surface of approximately 10⁹ CFU/cm². The ultrasound caused an approximate seven-log reduction, and the combination group eradicated viable biofilm bacteria completely. Confocal imaging corroborated the CFU data. Our results indicate that PL and ultrasound both are remarkably in eradicating biofilm, and the combination of PL and ultrasound is more effective than alone in reducing biofilm.

1. Introduction

Periprosthetic joint infection (PJI) is one of the most dreaded complications in joint replacement surgery, which is associated with pain, prolonged hospital stays, multiple surgeries, functional incapacitation, and even mortality [1]. Although the incidence of PJI is below 1-2%, with increasing number of patients undergoing joint replacement surgery, more implant-associated infections could happen [2].

These infections include acute infections (within the first 4 weeks after implantation) and late infections, which can be derived from either a perioperative contamination of the joint or an hematogenous spreading of bacteria to the joint

[3]. The incidence of pathogenic microorganisms depends on the origin and the time interval after the index surgery. *Staphylococcus aureus* is the most commonly isolated bacteria in acute PJI cases, while coagulase-negative *Staphylococcus* and *Streptococcus* are dominant in late infections and hematogenous infections, respectively [4–6].

Infections by *S. aureus* are characterized by rapid form biofilms [7]. Biofilms, aggregates of bacteria cells embedded in an extracellular polymeric substances (EPS) matrix, are fundamental with respect to the pathogenesis and persistence of PJIs which protect the dividing bacteria from the immune system, antibiotics, and even mechanical debridement [8–10]. Despite recent improvements in understanding in biofilm, clinical success in eradicating PJI

through revision surgery remains poor, even in the early postoperative period [11–13].

For acute PJI, debridement, antibiotics, and implant retention (DAIR) have been increasingly used due to less invasion and lower cost in contrast to two-stage exchange of the device. However, DAIR have higher failure rate (16–57.4%) [14–16], and *S. aureus* PJI appears to have lower success rate than other organisms [17, 18]. Inadequate removal of biofilms during the debridement is regarded as the major reason of DAIR failures. Many different debridement techniques in vitro have been used to mechanically disrupt and remove bacterial biofilm established on implant materials, such as iodine immersion, pulse lavage (PL), or even mechanical brushing. PL is a common method which mechanically disrupts and removes bacterial biofilm established on bone, soft tissue, and implant materials. However, some studies have found that PL is inadequate at removal of biofilm from the surface of implant materials in vivo [19, 20].

For the last decade, sonication has emerged and gradually becomes a practical and effective method to dislodge biofilm and the associated bacteria from the surface of the implant. It was found that the biofilm on the prosthesis surface could be removed by ultrasonic vibration after the PJI prosthesis was taken out; the bacteria in the biofilm could be released, so as to improve the positive rate of bacterial culture. Unfortunately, although ultrasound has destroyed the biofilm almost completely, residual bacterial viability could still be detected [21–23]. Recently, a direct-contact low-frequency ultrasound (DCLFU) device was introduced for the purpose of wound debridement [24]. Further investigation found that DCLFU is a promising method to treat biofilm infections [25]. Therefore, we assume whether DCLFU and PL in combination can act on the surface of prosthesis to remove the biofilm. Here, we extend the study to biofilms grown on titanium alloy coupons and compare the efficacy of eradicating *S. aureus* biofilms using PL alone, DCLFU alone, and the two treatments in combination. If it can completely remove the biofilm on the surface of prosthesis, it will greatly improve the success rate of DAIR, which is undoubtedly a great news for PJI patients.

2. Materials and Methods

2.1. Bacterial Strain. As previously described, a clinically isolated *S. aureus* strain was used in this study [19]. *S. aureus* strain was grown on tryptic soy agar (TSA) (Oxoid, Cambridge, UK). Then, representative colonies were picked and suspended in trypticase soy broth (TSB; Rishui Biotechnology, Qingdao, China), growing at 37°C overnight with agitation (200 rpm). Bacteria were harvested and resuspended in TSB, adjusted to a turbidity equivalent to that of a 1 McFarland standard and diluted 1:3000, achieving the final cell concentration of approximately 1×10^5 CFU/ml.

2.2. Biofilm Production. The biofilm was cultivated according to previously described steps with some modifications [20]. Briefly, biofilms were grown on titanium alloy

coupons ($10 \times 10 \times 1$ mm, roughness 0.47, Beijing AK Medical Co., Ltd.). The coupons were put in the 24-well clear bottom microtiter plates (Corning, Inc, Corning, NY). Subsequently, 2 ml bacteria suspension was added to each well and incubated for 24 hours at 37°C, 5% CO₂.

2.3. Debridement. Coupons were followed by either (i) debridement with PL, (ii) debridement with DCLFU, or (iii) both. 3 L of normal saline was taken to irrigate each coupon with PL irrigation set at the high setting (Five Continents, Ningbo, China). During the operation, the nozzle was kept upright with about 3 cm distance from the surface of the coupon. A single operator moved the nozzle over the entire surface of coupon randomly but in an equal fashion, which aimed to create the same conditions in the operating room. According to the manufacturer's instructions, we performed sonication treatment (Scientz, Ningbo, China) with the following parameters: frequency of 25 kHz, flow rate of 15 ml/min saline solution, and a processing time of 1 min. The DCLFU device wide hatch probe was placed 2 mm above the coupons. Experiments were performed in triplicate.

2.4. Viable Cell Count. After exposure to the treatments, phosphate buffered saline (PBS; Dulbecco's, Gibco, Grand Island, NY) were used to rinse each coupon. Sonication for 15 min at a frequency of 35 kHz in the 10 ml of PBS was used to remove the biofilm. Totally, sonication was duplicated three times. Between each time, there is a 10 s vortex period. A 10-fold serial dilution was prepared and plated onto a solid agar, which were incubated for 24 h (37°C, 5% CO₂), and then, the number of colony forming units (CFU) was counted, expressed as CFU/cm².

2.5. Confocal Laser Scanning Microscopy (CLSM). Confocal laser scanning microscopy (CLSM, Olympus FV10i, Waltham, MA) was used in the control group and treatment group to image the bacterial biofilms and conform the CFU data. The bacterial biofilms were observed using live-dead staining (Invitrogen Molecular Probes, USA) following the manufacturer's instructions. The Live-Dead kit contains SYTO-9 that stains viable bacterial DNA green, and dead cells appear red when propidium iodide (PI) enters compromised bacterial cell membranes. After exposure to the treatments, the coupons were lightly dipped in sterile water three times in order to remove nonfirmly attached bacteria and debris. Then staining for 15 min at room temperature in the dark, biofilms were rinsed with PBS to remove the extracellular dyes and observed with CLSM.

2.6. Statistical Analysis. CFU data were first log₁₀ transformed. Statistical comparisons between the geometric means of CFU/cm² from control and treatment groups were performed using SPSS Version 19 (IBM SPSS Statistics for Windows, Version 19.0 Armonk, NY: IBM Corp.) using an unpaired, two tailed Student's *t*-test assuming equal variance. Statistical significance was determined if $p < 0.05$.

3. Results

3.1. Viable Cell Count. In this study, the number of CFU on coupons aims to quantify bacteria. In the control group, the biofilms had grown to approximately 10^9 CFU/cm² after 24 h (Figure 1). In the treatment group, PL debridement showed an approximate two-log reduction in CFU/cm² compared to the control group ($p < 0.05$), whereas exposure to DCLFU indicated an approximate seven-log reduction ($p < 0.05$). The bacteria cannot be detected in debridement PL and DCLFU exposure, which is accounting for a nine-log reduction ($p < 0.05$).

3.2. Confocal Microscopy. Confocal microscopy was a useful method which can measure the level of biofilm debrided following pulse lavage irrigation and DCLFU (Figure 2). After PL the biofilm cell density and viability had been reduced (Figure 2). The coupon exposed to DCLFU showed even more reduction of surface attached bacteria (Figure 2). The combination of PL and DCLFU showed the least cells remaining on the surface, and they were almost all dead (red) (Figure 2).

4. Discussion

With the formation of biofilms and bacterial growth, it results in a refractory infection of the implants. Since high risks of infection could occur perioperatively, there is an urgent need for a novel and effective way to remove bacterial biofilms from the implants. In this study, we investigated the effect of PL and DCLFU on eradication of biofilm formed by *S. aureus* on the titanium alloy surface.

PL irrigation is a commonly used technique for debridement. Our results showed that PL reduced bacterial colonization by approximately two log number of cells. CLSM also showed that the biofilm had been reduced to a monolayer of cells on the surface, and the remaining cells appeared to be mostly viable after PL. These findings are consistent with previous studies that reported approximately two-log reduction in cell numbers following PL [20, 26]. Knecht et al. [20] have demonstrated that PL reduced the CFU count of strain of biofilms by approximately two orders of magnitude, from an initial cell count on the metal surface of approximately 10^9 CFU/cm². These studies illustrate that PL irrigation of implant materials removed a substantial mass of biofilm. Besides, it is also found that a substantial mass of biofilm still remains. Furthermore, some authors have found that residual bacteria on the coupons after PL were sufficient to restore a biofilm after incubation [27]. Therefore, a new debridement method which can completely remove the biofilm is needed.

Low-frequency ultrasound is a promising method to treat biofilm infections due to its advantages, such as beam directivity and capability of treating deep tissue targets without tissue damage [28]. Previous studies have demonstrated the effectiveness of ultrasound in treating biofilms [29, 30]. In our study, DCLFU alone resulted in approximately seven log CFU/cm² reduction of biofilm cells. CLSM

FIGURE 1: Biofilm cell density quantified after treatment demonstrating a reduction in biofilm mass. C, control group; PL, pulse lavage group; DCLFU, direct-contact low-frequency ultrasound group; PL + DCLFU, combination of the PL and DCLFU group. *, **, and *** indicate statistically significant log reductions compared to successive groups.

show that there is a significant reduction in biofilm burden but not completely with regard to DCLFU alone. These findings were supported by some other studies. Singh et al. found that sonication treatment did not manage to destroy the biofilm completely [21]. Ensing et al. concluded that ultrasound alone did not negatively affect bacterial viability, neither in planktonic state nor in biofilm [31]. In contrast, Nishikawa et al. showed that *Staphylococcus* biofilm can be effectively removed by ultrasonic exposure in the animal model [32]. Granick et al. have demonstrated that the biofilm was cleared off completely from titanium and stainless steels implant materials when treated with DCLFU [25]. Numbers of factors could cause the observation bias. Such as the frequency, the configuration, the intensity, the exposure time, and the material, these are all important factors in consideration whether biofilms can be eradicated. However, there is growing evidence that ultrasonic treatment alone does not seem to be capable of completely eradicating biofilms on its own.

The combination of sonication with other technologies such as high pressure requires further investigations as pointed out in a review by Piyasena et al. [33]. To the best of our knowledge, it is the first time to investigate the effects of DCLFU in combination with PL against *S. aureus* biofilm in vitro. In the present study, reductions of approximately two log CFU/cm² and seven log CFU/cm², respectively, were observed with PL and DCLFU alone. When PL were combined with DCLFU, the reduction of cell increased, reaching proximately nine log CFU/cm², implying a synergistic effect. The synergistic effect of PL and DCLFU remains unclear. It seems likely that PL reduces the load of bioburden, DCLFU mechanical destruction due to cavitation and acoustic microstreaming. It still needs to be clarified in future study.

There were some limitations in our study. First, we did not assess the implant surface modifications such as surface roughness, which significantly impacts on implant longevity [34]. Second, it may not be applicable to other types of bacteria, such as MRSA, because we used a single strain experimental model. Third, our study only focused on

FIGURE 2: Confocal laser scanning microscopic images showing *S. aureus* of control, pulse lavage (PL), and direct-contact low-frequency ultrasound (DCLFU) treatment groups. Live cells are stained green (appears bright in grayscale) and dead cells are stained red (appears dim in grayscale). The scale bar represents 100 μm magnification.

titanium alloy, and other materials of commonly used orthopaedic implants still need to be further investigated.

The present study provided insight into the elimination of biofilm formed by *S. aureus* on titanium alloy using PL or DCLFU alone or in combination. Our experimental evidence strongly indicated that PL or DCLFU technique alone did not have the ability to completely eradicate the biofilms in our model system. Conversely, the combination could clear the biofilms off from the implant materials totally. Our results suggest that the combination of PL and DCLFU might be an optimal technique in debridement for infected total joint implants. It will improve the success rate of DAIR in the early PJI while to avoid two-stage reimplantation that would bring patients more costs and risks.

Data Availability

The data generated or used to support the findings of this study are included within the article.

Conflicts of Interest

The authors declare that they have no conflicts of interest.

Authors' Contributions

Xiaoqing Wu, Xuanren Shi, and Mingcong Chen are the equal contributors.

Acknowledgments

This study was supported by the Natural Science Foundation of Shenzhen University General Hospital, Grant/Award Number: SUGH2018QD064.

References

- [1] C. Li, N. Renz, and A. Trampuz, "Management of peri-prosthetic joint infection," *Hip & Pelvis*, vol. 30, no. 3, pp. 138–146, 2018.
- [2] G. Tsaras, D. R. Osmon, T. Mabry et al., "Incidence, secular trends, and outcomes of prosthetic joint infection: a population-based study, Olmsted county, Minnesota, 1969–2007," *Infection Control & Hospital Epidemiology*, vol. 33, no. 12, pp. 1207–1212, 2012.
- [3] W. Zimmerli, "Clinical presentation and treatment of orthopaedic implant-associated infection," *Journal of Internal Medicine*, vol. 276, no. 2, pp. 111–119, 2014.
- [4] V. Zeller, Y. Kerroumi, V. Meyssonier et al., "Analysis of postoperative and hematogenous prosthetic joint-infection microbiological patterns in a large cohort," *Journal of Infection*, vol. 76, no. 4, pp. 328–334, 2018.
- [5] C. R. Arciola, D. Campoccia, and L. Montanaro, "Implant infections: adhesion, biofilm formation and immune evasion," *Nature Reviews Microbiology*, vol. 16, no. 7, pp. 397–409, 2018.
- [6] T. N. Peel, A. C. Cheng, K. L. Buising, and P. F. M. Choong, "Microbiological aetiology, epidemiology, and clinical profile of prosthetic joint infections: are current antibiotic prophylaxis guidelines effective?" *Antimicrobial Agents and Chemotherapy*, vol. 56, no. 5, pp. 2386–2391, 2012.
- [7] M. Betz, S. Abrassart, P. Vaudaux et al., "Increased risk of joint failure in hip prostheses infected with *Staphylococcus aureus* treated with debridement, antibiotics and implant retention compared to streptococcus," *International Orthopaedics*, vol. 39, no. 3, pp. 397–401, 2015.
- [8] P. S. Stewart and J. William Costerton, "Antibiotic resistance of bacteria in biofilms," *The Lancet*, vol. 358, no. 9276, pp. 135–138, 2001.
- [9] C. R. Arciola, D. Campoccia, P. Speziale, L. Montanaro, and J. W. Costerton, "Biofilm formation in staphylococcus implant infections. a review of molecular mechanisms and implications for biofilm-resistant materials," *Biomaterials*, vol. 33, no. 26, pp. 5967–5982, 2012.
- [10] J. Y. Sharahi, T. Azimi, A. Shariati, H. Safari, M. Khanzadeh, and A. Hashemi, "Advanced strategies for combating bacterial biofilms," *Journal of Cellular Physiology*, vol. 234, pp. 14689–14708, 2019.
- [11] G. Grammatopoulos, M.-E. Bolduc, B. L. Atkins et al., "Functional outcome of debridement, antibiotics and implant retention in periprosthetic joint infection involving the hip: a case-control study," *The Bone & Joint Journal*, vol. 99-B, no. 5, pp. 614–622, 2017.
- [12] W. S. Son, O.-J. Shon, D.-C. Lee, S.-J. Park, and H. S. Yang, "Efficacy of open debridement and polyethylene exchange in strictly selected patients with infection after total knee arthroplasty," *Knee Surgery & Related Research*, vol. 29, no. 3, pp. 172–179, 2017.
- [13] M. Maillet, P. Pavese, D. Bruley, A. Seigneurin, and P. François, "Is prosthesis retention effective for chronic infections in hip arthroplasties? a systematic literature review," *European Journal of Clinical Microbiology & Infectious Diseases*, vol. 34, no. 8, pp. 1495–1502, 2015.

- [14] K. L. Urish, A. G. Bullock, A. M. Kreger et al., "A multicenter study of irrigation and debridement in total knee arthroplasty periprosthetic joint infection: treatment failure is high," *The Journal of Arthroplasty*, vol. 33, no. 4, pp. 1154–1159, 2018.
- [15] L. d. Vries, W. v. der, W. Neve, H. Das, B. Ridwan, and J. Steens, "The effectiveness of debridement, antibiotics and irrigation for periprosthetic joint infections after primary hip and knee arthroplasty. a 15 years retrospective study in two community hospitals in The Netherlands," *Journal of Bone and Joint Infection*, vol. 1, no. 1, pp. 20–24, 2016.
- [16] A. F. Duque, Z. D. Post, R. W. Lutz, F. R. Orozco, S. H. Pulido, and A. C. Ong, "Is there still a role for irrigation and debridement with liner exchange in acute periprosthetic total knee infection?" *The Journal of Arthroplasty*, vol. 32, no. 4, pp. 1280–1284, 2017.
- [17] B. S. Konigsberg, C. J. D. Valle, N. T. Ting, F. Qiu, and S. M. Sporer, "Acute hematogenous infection following total hip and knee arthroplasty," *The Journal of Arthroplasty*, vol. 29, no. 3, pp. 469–472, 2014.
- [18] I. Byren, P. Bejon, B. L. Atkins et al., "One hundred and twelve infected arthroplasties treated with "DAIR" (debridement, antibiotics and implant retention): antibiotic duration and outcome," *Journal of Antimicrobial Chemotherapy*, vol. 63, no. 6, pp. 1264–1271, 2009.
- [19] K. L. Urish, P. W. DeMuth, D. W. Craft, H. Haider, and C. M. Davis, "Pulse lavage is inadequate at removal of biofilm from the surface of total knee arthroplasty materials," *The Journal of Arthroplasty*, vol. 29, no. 6, pp. 1128–1132, 2014.
- [20] C. S. Knecht, J. P. Moley, M. S. McGrath, J. F. Granger, P. Stoodley, and D. H. Dusane, "Antibiotic loaded calcium sulfate bead and pulse lavage eradicates biofilms on metal implant materials in vitro," *Journal of Orthopaedic Research*, vol. 36, no. 9, pp. 2349–2354, 2018.
- [21] G. Singh, R. Hameister, B. Feuerstein, F. Awiszus, H. Meyer, and C. H. Lohmann, "Low-frequency sonication may alter surface topography of endoprosthetic components and damage articular cartilage without eradicating biofilms completely," *Journal of Biomedical Materials Research Part B: Applied Biomaterials*, vol. 102, no. 8, pp. 1835–1846, 2014.
- [22] J. Xu, T. A. Bigelow, L. J. Halverson, J. M. Middendorf, and B. Rusk, "Minimization of treatment time for in vitro 1.1 MHz destruction of *Pseudomonas aeruginosa* biofilms by high-intensity focused ultrasound," *Ultrasonics*, vol. 52, no. 5, pp. 668–675, 2012.
- [23] T. A. Bigelow, T. Northagen, T. M. Hill, and F. C. Sailer, "The destruction of *Escherichia coli* biofilms using high-intensity focused ultrasound," *Ultrasonics in Medicine & Biology*, vol. 35, no. 6, pp. 1026–1031, 2009.
- [24] C. A. Murphy, P. Houghton, T. Brandys, G. Rose, and D. Bryant, "The effect of 22.5 kHz low-frequency contact ultrasound debridement (LFCUD) on lower extremity wound healing for a vascular surgery population: a randomised controlled trial," *International Wound Journal*, vol. 15, no. 3, pp. 460–472, 2018.
- [25] M. S. Granick, C. Paribathan, M. Shanmugam, and N. Ramasubbu, "Direct-contact low-frequency ultrasound clearance of biofilm from metallic implant materials," *Eplasty*, vol. 17, p. e13, 2017.
- [26] H. Poilvache, A. Ruiz-Sorribas, G. Sakoulas, H. Rodriguez-Villalobos, O. Cornu, and F. Van Bambeke, "Synergistic effects of pulsed lavage and antimicrobial therapy against *Staphylococcus aureus* biofilms in an in-vitro model," *Frontiers in Medicine*, vol. 7, p. 527, 2020.
- [27] E. M. Schwechter, D. Folk, A. K. Varshney, B. C. Fries, S. J. Kim, and D. M. Hirsh, "Optimal irrigation and debridement of infected joint implants: an in vitro methicillin-resistant *Staphylococcus aureus* biofilm model," *The Journal of Arthroplasty*, vol. 26, no. 6, pp. 109–113, 2011.
- [28] Y. Cai, J. Wang, X. Liu, R. Wang, and L. Xia, "A review of the combination therapy of low frequency ultrasound with antibiotics," *BioMed Research International*, vol. 2017, Article ID 2317846, 14 pages, 2017.
- [29] T. A. Bigelow, C. L. Thomas, H. Wu, and K. M. F. Itani, "Histotripsy treatment of *S. Aureus* biofilms on surgical mesh samples under varying pulse durations," *IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control*, vol. 64, no. 10, pp. 1420–1428, 2017.
- [30] A. K. Seth, K. T. Nguyen, M. R. Geringer et al., "Noncontact, low-frequency ultrasound as an effective therapy against *Pseudomonas aeruginosa*-infected biofilm wounds," *Wound Repair and Regeneration*, vol. 21, no. 2, pp. 266–274, 2013.
- [31] G. T. Ensing, D. Neut, J. R. v. Horn, H. C. van der Mei, and H. J. Busscher, "The combination of ultrasound with antibiotics released from bone cement decreases the viability of planktonic and biofilm bacteria: an in vitro study with clinical strains," *Journal of Antimicrobial Chemotherapy*, vol. 58, no. 6, pp. 1287–1290, 2006.
- [32] T. Nishikawa, A. Yoshida, A. Khanal et al., "A study of the efficacy of ultrasonic waves in removing biofilms," *Gerodontology*, vol. 27, no. 3, pp. 199–206, 2010.
- [33] P. Piyasena, E. Mohareb, and R. C. McKellar, "Inactivation of microbes using ultrasound: a review," *International Journal of Food Microbiology*, vol. 87, no. 3, pp. 207–216, 2003.
- [34] I. Catelas and M. A. Wimmer, "New insights into wear and biological effects of metal-on-metal bearings," *Journal of Bone and Joint Surgery*, vol. 93, no. Suppl 2, pp. 76–83, 2011.