Research Article

Cortical Tasks-Based Optimal Filter Selection: An fNIRS Study

Rayyan Azam Khan ¹, **Noman Naseer** ², **Sajid Saleem** ³, **Nauman Khalid Qureshi** ⁴,
Farzan Majeed Noori ⁵, and **Muhammad Jawad Khan**⁶

¹Department of Mechanical Engineering, University of Saskatchewan, Saskatoon, Saskatchewan S7N 5A9, Canada

²Department of Mechatronics Engineering, Air University, Islamabad 46000, Pakistan

³Department of Electrical Engineering, National University of Modern Languages, Islamabad 46000, Pakistan

⁴Department of Biomedical Engineering, Faculty of Electronic Information and Electrical Engineering,
Dalian University of Technology, Dalian 116000, China

⁵Department of Informatics, Faculty of Mathematics and Natural Sciences, University of Oslo, Oslo 0316, Norway

⁶School of Mechanical and Manufacturing Engineering, National University of Science and Technology,
Islamabad 46000, Pakistan

Correspondence should be addressed to Noman Naseer; noman.naseer@mail.au.edu.pk

Received 16 December 2019; Accepted 25 April 2020; Published 22 May 2020

Academic Editor: Ahmed F. Hussein

Copyright © 2020 Rayyan Azam Khan et al. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Functional near-infrared spectroscopy (fNIRS) is one of the latest noninvasive brain function measuring technique that has been used for the purpose of brain-computer interfacing (BCI). In this paper, we compare and analyze the effect of six most commonly used filtering techniques (i.e., Gaussian, Butterworth, Kalman, hemodynamic response filter (hrf), Wiener, and finite impulse response) on classification accuracies of fNIRS-BCI. To conclude with the best optimal filter for a specific cortical task owing to a specific cortical region, we divided our experimental tasks according to the three main cortical regions: prefrontal, motor, and visual cortex. Three different experiments were performed for prefrontal and motor execution tasks while one for visual stimuli. The tasks performed for prefrontal include rest (R) vs mental arithmetic (MA), R vs object rotation (OB), and OB vs MA. Similarly, for motor execution, R vs left finger tapping (LFT), R vs right finger tapping (RFT), and LFT vs RFT. Likewise, for the visual cortex, R vs visual stimuli (VS) task. These experiments were performed for ten trials with five subjects. For consistency among extracted data, six statistical features were evaluated using oxygenated hemoglobin, namely, slope, mean, peak, kurtosis, skewness, and variance. Combination of these six features was used to classify data by the nonlinear support vector machine (SVM). The classification accuracies obtained from SVM by using hrf and Gaussian were significantly higher for R vs MA, R vs OB, R vs RFT, and R vs VS and Wiener filter for OB vs MA. Similarly, for R vs LFT and LFT vs RFT, hrf was found to be significant ($p < 0.05$). These results show the feasibility of using hrf for effective removal of noises from fNIRS data.

1. Introduction

Brain-computer interface (BCI) also known as human-machine interface (HMI) or brain-machine interface (BMI) provides a communication mean between the user and external devices through a combination of hardware and software systems [1–3]. These systems are trained to generate control commands based on a specific set of patterns of brain signals [4].

Brain signal acquisition is categorized between invasive and noninvasive techniques. However, due to surgical risks and limited access to the cortical region, noninvasive

techniques are common in practice [5]. Noninvasive modalities include functional magnetic resonance interference (fMRI), functional near-infrared spectroscopy (fNIRS) [6], and electroencephalography (EEG) [4]. fNIRS is a comparatively new modality that has better spatial resolution and low artifacts, cost, and portability [4, 7]. So far, promising results have been shown by fNIRS-BCI [8–10]. Acquired brain signals for a specific task may contain noises that can contaminate signals and can effect informative data. These noises are categorized between physiological noise, experimental noise, and instrumental noise [8, 11]. In fNIRS,

the experimental noises are removed prior to the change of the raw signal to its magnitude through the modified Beer-Lambert law [8, 12, 13]. Noises produced due to hardware or by surrounding are known to be instrumental noises. These noises usually have high frequency that can be removed using low-pass filter; furthermore, keeping isolation from external sources such as light can reduce such type of noises. Experimental noises include motion artifacts such as head motion during signal acquisition that can cause the dislocation of optodes from the assigned position, thus generates a spike-like noise due to change in light intensity. Various studies [13–15] have utilized commonly developed filtering techniques randomly for noise removal. However, obtained signals can be corrupted by different kinds of noises that could affect further analysis. Noises can be physiologically produced due to Mayer waves (~ 0.1 Hz), respiration ($0.2\sim 0.5$ Hz), and heartbeat ($1\sim 1.5$ Hz), mainly due to fluctuations of blood pressure [8, 16–19]. These noises can be removed using adaptive or bandpass filtering [20, 21]. After preprocessing, useful information is extracted from the filtered data afterward classified using different classifiers mainly named as linear discriminant analysis (LDA), support vector machine (SVM), quadratic discriminant analysis (QDA), and naïve bayes (NB) [22] to generate control commands, hence completing the loop for BCI. Previous studies [13, 14, 23] show that an appropriate filter for correctness of data is the key to achieve more accurate results.

In this study, we hypothesized to find an optimal filter for commonly used cortical tasks, owing to a particular cortical region. Hence, we compared six commonly used filters to remove previously discussed noises. These filters include discrete Kalman [24], time-varying Wiener [25], 4th order Butterworth, hemodynamic response filter (hrf), Gaussian [26], and window-based finite impulse response (FIR) [27]. For the said purpose, cortical data were acquired from the three main regions of the brain, namely, prefrontal (PFC), motor (MC), and visual cortex (VC). Since the data acquired from PFC relate to thinking tasks [28–31], hence arithmetic and object rotation tasks were performed for this cortical region [32–34]. Similarly, tasks related to movement of limbs or fingers is related to the motor cortex [35]; therefore, the finger tapping tasks were performed for desired data acquisition [36–38]. Likewise, flickering of checker box was performed for visual cortex data [39]. Keeping in view the target of an optimal filter for a specific cortical region, a previous study [22] reported the different combinations of statistical features. Therefore, for consistency of extracted data, statistical features were kept the same for all experimental tasks, hence making combinations of six features, namely, signal mean (SM), signal slope (SS), signal peak (SP), signal skewness (SK), signal kurtosis (KR), and signal variance (SV). For classification, a number of studies [22, 40, 41] reported nonlinear SVM classifiers for comparatively better accuracies, hence all experimental tasks were classified using a nonlinear SVM classifier. Therefore, the main contribution of this work is (1) to analyze the effect of the six most commonly used filtering techniques and (2) to propose the optimal one among the most frequently

discussed noise removal techniques. For the aforementioned experiment, we select three main cortical regions: prefrontal, motor, and visual cortex with seven various paradigms. The canonical hemodynamic response filter (hrf) [26] performed overall best among the opted techniques. On the basis of these systematic and explicit analyses, it can be seen that selection of an optimal filter has a significant role in enhancing accuracies. Hence, these observations can serve as a standard guide for others to test the effect of noise correction algorithms for fNIRS experiments, and therefore can select a significant methodology.

2. Materials and Methods

2.1. Experimental Setup. To acquire experimental data, seven paradigms owing to three main cortical regions were designed and explicitly performed using a dynamic near-infrared optical tomography (DYNOT-232; NIRx Medical Technologies, NY, USA) device at Pusan National University. It operates on two wavelengths that are 760 and 830 nm where the signal acquisition sampling frequency was 1.81 Hz. Five healthy subjects with normal or corrected-to-normal vision took part in the experiment with a verbal consent before experimentation. All subjects were right-handed with an age range of 26 ± 3 . Right-handed subjects were selected to minimize hemodynamic response variation due to hemispheric-dominance difference. The experimental participants had no history of alcoholism, psychiatric, neurological, and visual disorder, cardiovascular and respiratory disease, mental illness, or any motor disability. Moreover, three hours before the commencement of the study, participants were asked to refrain from caffeinated drinks. As discussed in the literature [28, 30, 31, 35, 37, 38], for thinking-related task, signals were acquired from PFC, similarly for motor execution tasks from the primary motor cortex (PMC), and visual stimuli task from VC. Performed experiments were according to the latest Declaration of Helsinki.

2.2. Experimental Paradigms. In accordance with the literature [35, 39], subjects were seated on a comfortable chair and were asked to take rest with restricted movements as they can, so that the hemodynamic response activation owing to previous activities can be avoided. Hence, each paradigm related to PFC, MC, and VC starts with a rest of 20 s period to set up the baseline conditions. As the literature [42–45] show, (10–12) s task is adequate to acquire hemodynamic response of brain activity, hence 20 s initial rest was followed by 10 s task, and this was followed in turn by another 20 s rest period permitting signals to return to their baseline values before the start of the next trial in paradigm (a). For paradigm (b), 20 s rest after 10 s task 1 was again followed by 10 s task 2. The 20 s rest between two 10 s tasks was added to differentiate two classes through the baseline value. Figure 1 depicts paradigm (a) and paradigm (b). For optimal filter selection, tasks were selected concerning specific cortical regions such that for PFC, three different experiments were performed, rest (R) vs mental arithmetic (MA), R vs object rotation (OB), and OB vs MA task. Similarly, for MC, R vs left finger tapping (LFT), R vs right

FIGURE 1: Experimental paradigms for cortical tasks. (a) Task vs rest. (b) Task 1 vs task 2.

finger tapping (RFT), and LFT vs RFT. Likewise, for VC, R vs visual stimuli (VS) task.

2.2.1. Prefrontal Tasks

(1) *Mental Arithmetic*. According to previous studies [44, 46, 47], for R vs MA task, subjects were asked to perform a series of arithmetic calculations for 10 s based on the pseudorandom order, such that to subtract the random two-digit number (between 10 and 20) successively from the previous result of the three-digit number subtraction appearing on the screen (e.g., 400-11, 389-17, and 372-14). Afterwards, the screen was turned black so that subjects does not go beyond the 10 s task.

(2) *Object Rotation*. For the object rotation task, subjects were asked to imagine a cube rotating for the 10 s task while the “object rotation” word appeared for 10 s on the screen at about 2 m distance [48–50].

(3) *Mental Arithmetic vs Object Rotation*. In this protocol, subjects were asked to perform the aforementioned MA task versus OB task in between 20 s of rest to distinguish two tasks. The experimental paradigm is depicted in Figure 1(b).

2.2.2. Motor Tasks

(1) *Finger Tapping*. According to the literature [8, 17, 19, 23, 31, 37, 38, 51], subjects were asked to tap the self-paced index finger of one hand for 10 s afterwards the 20 s rest task was performed allowing signals to return to their reference values. Also, repetition for 10 times was performed as depicted in Figure 1(a). Similar trials were performed on the other hand, while for the LFT vs RFT task, 20 s rest was performed in between two tasks for the restoration of the signal to the baseline level as shown in Figure 1(b).

2.2.3. Visual Task

(1) *Checker Box Flickering*. In this experiment [4, 18, 39], a screen was placed in front of the subjects at a distance of approximately 2 m, and also subjects were requested to avoid

eye blinking during the experiment. The 10 s task of checker box flickering at 4 Hz was performed followed by 20 s rest of the black screen. The sound was also generated during the transition between rest and task. The paradigm followed for visual stimuli is shown in Figure 1(a).

2.3. *Experimental Setup*. Since the mental imagery task activates the PFC [34, 43], a total 11 of near-infrared (NI) light optodes were placed on PFC, 3 of which were detectors and 8 were the source in accordance with the literature [46, 47]. Similarly, for the motor execution task, the primary motor cortex (PMC) is activated [37, 43], hence 15 optodes were placed on PMC out of which 8 were the source and 7 were detectors. To extract data for LFT, optodes were placed on the right hemisphere, while for RFT on the left hemisphere [43]. Similarly, for the visual stimuli task data acquisition from the visual cortex [4, 18, 39], eleven optodes were placed having eight sources and three detectors. The distance between the source and the detector was 3 cm. Optode placement with channel configuration for MC, VC, and PFC is shown in Figure 2.

2.3.1. *Signal Acquisition*. In accordance with the literature [6, 8, 29], the raw optical density signal is converted to oxyhemoglobin ($\Delta c_{\text{HbO}}(t)$) and deoxyhemoglobin ($\Delta c_{\text{HbR}}(t)$) concentration using the modified Beer–Lambert law (MBLL) as described in the following equation:

$$\begin{bmatrix} \Delta c_{\text{HbO}}(t) \\ \Delta c_{\text{HbR}}(t) \end{bmatrix} = \frac{\begin{bmatrix} \beta_{\text{HbO}}(\lambda_1) & \beta_{\text{HbR}}(\lambda_1) \\ \beta_{\text{HbO}}(\lambda_2) & \beta_{\text{HbR}}(\lambda_2) \end{bmatrix}^{-1} \begin{bmatrix} \Delta \psi(t, \lambda_1) \\ \Delta \psi(t, \lambda_2) \end{bmatrix}}{d * l}, \quad (1)$$

where $\beta_{\text{HbX}}(\lambda)$ is the HbX extinction coefficient in $\{\mu\text{M}^{-1}\text{cm}^{-1}\}$, d is the differential path length factor for the curved path in (mm), l is the detector and emitter distance in (mm), and $\Delta \psi_{\text{HbX}}(t)$ is the absorbance difference of the light emitter wavelength of λ_i .

2.4. *Signal Processing*. The acquired raw signals of the brain contain various noises that can be categorized into physiological, experimental, and instrumental noise [8, 52]. In

FIGURE 2: Optode placement configuration in accordance with the 10–20 international system. (a) Motor cortex. (b) Visual cortex. (c) Prefrontal cortex.

fNIRS, the instrumental and experimental noises are removed prior to change of the raw signal to its magnitude through MBLL [8, 12, 14].

2.4.1. Instrumental Noises. Noises produced due to hardware or by surrounding are categorized under instrumental noises. These noises usually have high frequency that can be removed using a low-pass filter; furthermore, keeping isolation from external sources such as light can reduce such type of noises [8].

2.4.2. Experimental Noises. These errors contain motion artifacts due to unintentional body movements like head motion during signal acquisition. It can cause the dislocation of optodes from the assigned positions that result in a spike-like noise due to change in light intensity. Filters like Kalman and Wiener can be used to remove such type of noises [12–15].

2.4.3. Physiological Noises. These noises are produced due to Mayer waves (~ 0.1 Hz), respiration (~ 0.5 Hz), and heartbeat ($1\sim 1.5$ Hz) that occur due to fluctuations of blood pressure [16–19]. These noises can be removed using adaptive or bandpass filtering [8, 20, 38].

2.5. Data Analysis. NIRS-SPM is a toolbox designed for fNIRS data analysis. For signal processing, it provides common filtering techniques, namely, Butterworth, Gaussian, and hrf [26]. The comparative data analysis was performed by implementing Gaussian and hrf filtering using the NIRS-SPM toolbox, while other techniques on MATLAB® 2017b. The generalized mathematical models with necessary details are as follows.

2.5.1. Gaussian Filtering. A Gaussian filter is used in various forms depending upon the nature of the signal. Generally, a Gaussian filter is based on a Gaussian function which defines

the probability distribution of noise or data. It can also be used as a smoothing operator. A Gaussian kernel is used for smoothing the signal in which each value is replaced with the weighted average of itself and its neighboring values [14, 21, 26]. A simple representation of the 2D Gaussian filter can be defined in the following equation as

$$G(x, y) = \frac{1}{2\pi\sigma^2} e^{-(x^2 + y^2)/2\sigma^2}, \quad (2)$$

where x and y are the distance from the origin in horizontal and vertical axis and σ is the known standard deviation of the distribution.

2.5.2. Hemodynamic Response Filter (hrf). The hrf is based on the canonical representation of the hemodynamic response functions (HRF) and is used for the temporal smoothing of the fNIRS time series signal. In NIRS-SPM, the given functional data were smoothen using the least square estimate with ideal HRF. The hrf and Gaussian filter model details are in accordance with the literature [26].

2.5.3. Butterworth Filter. Butterworth filter is a model-based bandpass filter which performs on frequency attenuation using high and low-pass filter. The filtered value not only depends on the weighted average of the unfiltered time series, but also recursively on the previous values of the filtered time series. This filter aims to have a flat frequency response in the desired pass band [26]. The 4th order Butterworth filter with a band pass of (0.01–0.1) Hz was applied by MATLAB® build in a library for the desired experiments.

2.5.4. Finite Impulse Response Filter. FIR filter is designed by finding the coefficients and filter order so that it performs a cross-correlation between the input signal and the time reversed impulse response; therefore, by sampling the pulse shape, coefficients of the filter are designed [14]. FIR filter of order N can be defined as the following equation:

$$x[n] = \sum_{i=0}^N a_i y[n-i], \quad (3)$$

where $y[n]$ is the input signal, $x[n]$ is the output signal, N is the filter order, and b_i is the value of impulse response at i^{th} instance. Here, the 4th order FIR filter with a low-pass band of 0.1 Hz was utilized. The coefficients were estimated using the least square estimate. However, a time-varying Wiener filter, based on the short-time Fourier series, was implemented as in [25, 53].

2.5.5. Kalman Filter. Kalman filters the input signal containing statistical and other noises by linear quadratic estimation. The estimated unknown variables based on the Bayesian inference and joint probability distribution tend to be more accurate [13, 14]. Its simple model can be seen in the following equation:

$$x_k = F_k x_{k-1} + B_k u_k + w_k, \quad (4)$$

where F_k is the state transition model applied to previous state x_{k-1} , B_k is the control input model applied to the control vector u_k , and w_k is the process noise. A discrete model of Kalman was implemented in accordance with [24].

2.6. Feature Selection. For the consistency between extracted data across all paradigms, six statistical features (SV, KR, SS, SM, SK, and SP) were used to extract information across data [28, 43, 46, 54, 55]. SM is calculated in the following equation as

$$SM = \frac{1}{n} \sum_{X=1}^n A_x, \quad (5)$$

where A_x is the input signal such that $\Delta c_{\text{HbO}}(t)$ and n is the total number of observations. Signal variance is calculated in equation (6) as

$$\sigma^2 = \frac{1}{n-1} \sum_{x=1}^{n-1} (Z_x - \mu)^2, \quad (6)$$

where Z_x is the input signal, μ is the mean found from equation (6), n is the number of samples, and σ is the standard deviation. For KR, calculation was made by the following equation:

$$\text{kurt}(X) = E \left[\left(\frac{X - \mu}{\sigma} \right)^4 \right], \quad (7)$$

where X is the input signal and E is the expected value of X .

SK is the asymmetry of values relative to normal distribution around the mean, hence calculated in the following equation:

$$\text{skew}(X) = E \left[\left(\frac{X - \mu}{\sigma} \right)^3 \right]. \quad (8)$$

MATLAB® polyfit function fits the line to all input data points, therefore used to calculate SS. Similarly, max

function was used to calculate SP. Statistical features were rescaled between 0 and 1 using the following equation:

$$Z' = \frac{Z - \min(Z)}{\max(Z) - \min(Z)}, \quad (9)$$

where Z' is the rescaled feature and Z refers to the original feature values. The scatter plot across all six statistical features for the OB vs MA task of subject 1 is shown in Figure 3.

2.7. Support Vector Machine. Statistical significance of accuracy is analyzed for selection of an optimal filter for a specific cortical region; therefore, for higher classification performance, nonlinear SVM is used [35, 38]. It can rescale high-dimensional data and can control errors explicitly by maximizing the margins between two or more classes thus creating hyperplanes named as support vectors [40, 41, 56].

The cost function that is to be maximized for the SVM classifier gives a correlation between training data and hyperplane as defined in equations (10) and (11), respectively,

$$\text{Minimize } \frac{1}{2} \|z\|^2 + k \sum_{x=1}^N \varepsilon_x, \quad (10)$$

$$\text{Provided } y_x (Z^T \xi_x + a)^3 \geq 1 - \varepsilon_x, \quad \varepsilon_x \geq 0, \quad (11)$$

where $z, \xi_x \in R^2, b \in R^1, z^2 = z^T z$, k is the positive regularization parameter, ε_x is the measure of the training error, and y_x is the class label for the n th sample. Here, the third-degree polynomial kernel function with $k = 0.5$ and 10-fold cross-validation was applied for the estimation of classification accuracies [53].

3. Results and Discussion

3.1. Results. In this study, an optimal filter was chosen based on cortical tasks. Activities were categorized based on three main regions such as PFC, MC, and VC. $\Delta c_{\text{HbO}}(t)$ signals were filtered using six filters. Figures 4–6 show the averaged $\Delta c_{\text{HbO}}(t)$ filtered signals across trials of prefrontal, motor execution and, visual stimuli tasks. The three different signals show various paradigms for each brain region, while the horizontal axis is aligned with one complete event. The pictorial analysis of filtered responses shows that using the 4th order Butterworth changes data form altogether while discrete Kalman and Wiener filtered the signals, but there remain some noises in the output response, whereas, canonical hrf and Gaussian give much smoother response as compared to any other technique. Moreover, previous studies have commonly utilized Butterworth for signal processing, while the obtained visual does not show any significant improvement in comparison. Likewise, the Gaussian filter is also utilized in various studies that assume the acquired data with some normal distribution; therefore, it considers Gaussian function which defines the probability distribution of noise or data. It can also be used as a smoothing operator. A Gaussian kernel is used for smoothing the signal in which each value is replaced with the weighted average of itself and its neighbouring values.

FIGURE 3: Extracted feature scatter plot for OB vs MA.

However, the hrf filter considers ideal hemodynamic response as a reference for smoothing the data that are closely related to chi-squared distribution. Due to this fact, hrf outperforms any other considered technique.

For consistency among extracted information, classification accuracies were obtained across six statistical features. The accuracies obtained across these features are shown in Table 1. The two-tailed independent t -test was performed to check statistical significance with Holm–Bonferroni for multiple comparison of filters. Results obtained in Table 2 show significance using hrf or the Gaussian filter for R vs MA and R vs OB while for the OB vs MA task, no significant filter was found; however, the time-varying Wiener filter outperformed others. Similarly, in motor execution tasks, hrf was found to be significant for R vs LFT, likewise for R vs RFT hrf and Gaussian. Moreover, for LFT vs RFT, the hrf filter shows better performance. In the VS task, hrf and Gaussian were significant. Moreover, we plot the mean accuracies of filters with respect to cortical regions as depicted in Figure 7. The Gaussian and canonical hrf performed consistently better across all three cortical regions. However, comparatively better accuracies were obtained using hrf only. The statistical results validate our hypothesis of optimal filter selection by assuring that hrf generally has better performance for fNIRS-based studies, which is in accordance with a previous study [53].

3.2. Discussion. Despite of the fact that fNIRS offers portability, low cost, and ease of equipment setup, there still remains a challenge of removing noises like systematic physiological (Mayer waves, muscle activity, blood pressure,

and respiration and heart rate) and artifacts [17, 57], as fNIRS signals are highly contaminated by measurement noises and physiology-based systemic interference [58]. Several studies have proposed methodologies that can remove noises robustly. In [59], temporal filtering using a low-pass filter with 0.6 Hz cutoff frequency and canonical hemodynamic response function with 4 s full width at half maximum was applied. In [60], exponential moving average and Chebyshev filter were used to remove artifacts from the fNIRS data. Similarly, in [61], only low-pass filter with a 0.14 Hz cutoff was applied to remove physiological noises from fNIRS signals. Aqil et al. [62] used recursive least square estimation for online imaging. This adaptive approach provided a spatial filtering with low and high pass, detrending, and baseline correction. Similarly, Seo et al. [63] assessed the utility of NIRS in removal of physiological noises in fMRI data by reducing variance of the residual error in the baseline fMRI signal through the NIRS signal in the model. Similarly, in [64], adaptive filtration with the affine projection algorithm was used to accelerate convergence with colored noise, but it increases computational cost. In [65], a bandpass filter based on the 5th order Butterworth filter was used to filter motor execution signals based on EEG. In literature [52], fNIRS-based walking signals and walking signals while talking were acquired. These signals were preprocessed using a low-pass with finite impulse response filter, while the talking task results in low-amplitude artifacts with a similar frequency of hemodynamic response, hence it may affect cortical activity [57]. In [66], multiple filters were applied for EEG- and ECG-based signal preprocessing such as for removal of motion artifacts, a median filter with 5-point, for systemic component

cancellation of motion artifacts using Wiener and discrete Kalman filter was discussed; meanwhile, t -tests performed in comparison showed better performance with Kalman. Similarly, in [12, 13], four techniques, namely, Kalman, principal component analysis, wavelet analysis, and spline interpolation, were compared to remove NIRS data artifacts. Results showed that spline interpolation and wavelet analysis were significant for such noise removals. In [67], systemic noise was removed using wavelet minimum description length detrending approach and artifact using moving standard deviation and spline interpolation. Eggenberger

TABLE 1: Accuracies across six features for multiple cortical paradigms.

<i>Prefrontal tasks</i>						
R vs MA						
Subjects	Butter	FIR	Gaussian	hrf	Kalman	Wiener
S1	57.03	62.22	81.48	82.96	82.96	80
S2	57.77	59.25	71.85	83.7	68.148	69.62
S3	58.51	54.07	88.14	90.37	73.33	82.22
S4	64.44	71.85	82.22	82.96	72.59	74.81
S5	55.55	69.62	85.18	82.22	61.48	74.81
R vs OB						
S1	63.7	68.14	80	86.66	71.48	75.18
S2	54.81	68.88	78.51	80	68.88	73.33
S3	60	65.92	82.22	89.62	89.22	74.81
S4	63.7	71.11	97.03	97.03	90.37	94.81
S5	60	60	82.96	77.03	68.14	65.18
OB vs MA						
S1	50	64.44	68.88	75.55	85.55	90
S2	53.33	66.66	72.22	72.22	84.44	86.66

TABLE 1: Continued.

S3	54.44	68.88	85.55	80	76.66	74.44
S4	56.66	71.11	81.11	83.33	85.55	86.66
S5	57.77	75.55	82.22	82.22	85.55	88.88
<i>Motor execution tasks</i>						
R vs LFT						
S1	65.18	57.03	77.03	81.48	57.03	66.66
S2	60	57.77	78.51	85.18	45.18	50.37
S3	55.55	49.62	67.4	72.59	59.25	59.25
S4	53.33	56.29	68.14	73.33	60	61.48
S5	45.18	57.77	76.29	76.29	45.18	45.18
R vs RFT						
S1	61.48	43.7	78.51	89.62	66.66	65.92
S2	59.25	58.51	68.88	79.25	59.25	59.25
S3	57.77	65.92	68.88	62.96	46.66	60.74
S4	54.81	51.85	66.66	74.07	60	47.4
S5	62.96	58.51	71.85	77.03	55.77	56.29
LFT vs RFT						
S1	54.44	46.66	87.77	92.22	64.44	60
S2	55.55	61.11	64.44	76.66	53.33	51.11
S3	58.88	58.88	64.44	68.88	63.33	45.55
S4	54.44	57.77	78.88	80	48.88	52.22
S5	60	62.22	70	80	55.55	56.66
<i>Visual stimuli task</i>						
R vs VS						
S1	53.33	60.74	71.11	65.18	48.88	51.85
S2	54.81	66.66	74.07	75.55	59.25	62.96
S3	59.25	64.44	76.29	82.22	71.11	69.62
S4	57.77	47.4	68.14	73.33	49.62	57.77
S5	68.88	62.96	73.33	78.51	56.29	61.48

TABLE 2: Statistical significance of filters across multiple cortical paradigms.

Filters	<i>p</i> values
<i>Prefrontal tasks</i>	
R vs MA	
Butter vs hrf, FIR, Gaussian, Kalman, Wiener	0.089, 0.012, 0.012, 0.012, 0.012
hrf vs Gaussian, Wiener, FIR, Kalman	0.089, 0.012, 0.012, 0.022
Gaussian vs FIR, Kalman, Wiener	0.012, 0.089, 0.012
Kalman vs Wiener, FIR	0.089, 0.089
Wiener vs FIR	0.012
R vs OB	
Butter vs FIR, Gaussian, Kalman, Wiener, hrf	0.022, 0.012, 0.012, 0.012, 0.012
hrf vs Kalman, Gaussian, Wiener, FIR	0.012, 0.158, 0.012, 0.012
Gaussian vs FIR, Kalman, Wiener	0.012, 0.089, 0.012
Kalman vs Wiener, FIR	0.327, 0.022
Wiener vs FIR	0.012
OB vs MA	
Butter vs FIR, Gaussian, Kalman, Wiener, hrf	0.012, 0.012, 0.012, 0.012, 0.012
hrf vs Kalman, Gaussian, Wiener, FIR	0.089, 0.281, 0.089, 0.012
Gaussian vs FIR, Kalman, Wiener	0.012, 0.089, 0.089
Kalman vs Wiener, FIR	0.089, 0.012
Wiener vs FIR	0.012
<i>Motor execution tasks</i>	
R vs LFT	
Butter vs FIR, Gaussian, Kalman, Wiener, hrf	0.65, 0.025, 1, 0.158, 0.025
hrf vs Kalman, Gaussian, Wiener, FIR	0.012, 0.022, 0.012, 0.025
Gaussian vs FIR, Kalman, Wiener	0.012, 0.012, 0.012
Kalman vs Wiener, FIR	0.041, 1
Wiener vs FIR	0.32

TABLE 2: Continued.

Filters	<i>p</i> values
R vs RFT	
Butter vs FIR, Gaussian, Kalman, Wiener, hrf	0.089, 0.012, 0.5, 0.5, 0.012
hrf vs Kalman, Gaussian, Wiener, FIR	0.012, 0.089, 0.012, 0.089
Gaussian vs FIR, Kalman, Wiener	0.012, 0.012, 0.012
Kalman vs Wiener, FIR	0.5, 0.32
Wiener vs FIR	0.32
LFT vs RFT	
Butter vs FIR, Gaussian, Kalman, Wiener, hrf	0.158, 0.012, 0.32, 0.089, 0.012
hrf vs Kalman, Gaussian, Wiener, FIR	0.012, 0.012, 0.012, 0.012
Gaussian vs FIR, Kalman, Wiener	0.012, 0.012, 0.012
Kalman vs Wiener, FIR	0.32, 0.32
Wiener vs FIR	0.089
<i>Visual stimuli</i>	
R vs VS	
Butter vs FIR, Gaussian, Kalman, Wiener, hrf	0.32, 0.012, 0.32, 0.5, 0.012
hrf vs Kalman, Gaussian, Wiener, FIR	0.012, 0.089, 0.012, 0.012
Gaussian vs FIR, Kalman, Wiener	0.012, 0.012, 0.012
Kalman vs Wiener, FIR	0.089, 0.32
Wiener vs FIR	0.32

Filters mean accuracies with standard deviation

FIGURE 7: Filters' mean accuracies.

[84] removed channels with large movement or technical artifacts. Smaller artifacts were corrected with the correlation-based signal improvement method while a low-pass filter with a 5.0 s moving average filter was applied. In [85–87], signals were analyzed with SPM99 (Statistical Parametric Mapping software; Wellcome Department of Cognitive Neurology, London, UK). In [88], moving standard deviation-based artifact removal (moving artifact reduction algorithm: MARA) with a threshold of 0.45 for HbO and 0.18 for HbR was applied and signals were linearly detrended and low-pass filtered at 0.1 Hz.

Keeping in view the discussions in the literature, various methodologies were adopted for artifact removal while for systematic physiological noises, mostly high and low-pass filter were applied. Therefore, to make use of the best optimal filter for a specific cortical task, we performed an explicit and systematic analysis to find the filter based on statistical significance for mostly used cortical tasks in fNIRS study applications. The result shows that hrf outperforms the discussed techniques due to the fact that it considers ideal hemodynamic response distribution for smoothing the data. Keeping this in view, our future work involves designing of

an adaptive filter technique, considering different distributions of data and noise. Therefore, opting more accurate separation of data.

4. Conclusion

To the best of our knowledge, this is the first study to propose filter selection for commonly used cortical tasks of functional near-infrared spectroscopy (fNIRS) contaminated with artifacts and systematic physiological noises. Six filters, namely, Gaussian, hemodynamic response filter (hrf), Butterworth, time-varying Wiener, discrete Kalman, and window-based finite impulse response, were tested. The results obtained have validated the overall statistical significance of the hrf for prefrontal, motor, and visual cortex tasks. Furthermore, signals acquired from different cortical regions may contain different types of noises. Hence, the prime goal of this study was to suggest an optimal filter for a specific task, owing to the specific cortical region so that further studies can achieve maximum accuracies by reliably improving the recovered hemodynamic response function. Outcome of this study shows that there is a significant impact of filter selection on the accuracy of the classified data, therefore facilitating users to avoid analysis of complex signal techniques by themselves.

Data Availability

The data used to support the findings of this study are available from the corresponding author upon request.

Conflicts of Interest

The authors declare that there are no conflicts of interest regarding the publication of this paper.

Acknowledgments

The authors would like to thank Prof. Keum-Shik Hong from Pusan National University for providing them an opportunity to visit his lab and use the equipment in his lab for data acquisition.

References

- [1] N. Naseer and K.-S. Hong, "fNIRS-based brain-computer interfaces: a review," *Frontiers in Human Neuroscience*, vol. 9, pp. 1–15, 2015.
- [2] L. F. Nicolas-Alonso and J. Gomez-Gil, "Brain computer interfaces, a review," *Sensors*, vol. 12, no. 2, pp. 1211–1279, 2012.
- [3] J. R. Wolpaw, N. Birbaumer, D. J. McFarland, G. Pfurtscheller, and T. M. Vaughan, "Brain-computer interfaces for communication and control," *Clinical Neurophysiology*, vol. 113, no. 6, pp. 767–791, 2002.
- [4] A. Ferreira, W. C. Celeste, F. A. Cheein, T. F. Bastos-Filho, M. Sarcinelli-Filho, and R. Carelli, "Human-machine interfaces based on EMG and EEG applied to robotic systems," *Journal of NeuroEngineering and Rehabilitation*, vol. 5, no. 1, p. 10, 2008.
- [5] S. Waldert, "Invasive vs. Non-invasive neuronal signals for brain-machine interfaces: will one prevail?" *Frontiers in Neuroscience*, vol. 10, pp. 1–4, 2016.
- [6] M. Aqil, K.-S. Hong, M.-Y. Jeong, and S. S. Ge, "Detection of event-related hemodynamic response to neuroactivation by dynamic modeling of brain activity," *Neuroimage*, vol. 63, no. 1, pp. 553–568, 2012.
- [7] S. Fok, R. Schwartz, M. Wronkiewicz et al., "An EEG-based brain computer interface for rehabilitation and restoration of hand control following stroke using ipsilateral cortical physiology," in *Proceedings of the 2011 Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, pp. 6277–6280, IEEE, Boston, MA, USA, August 2011.
- [8] T. J. Huppert, S. G. Diamond, M. A. Franceschini, and D. A. Boas, "HomER: a review of time-series analysis methods for near-infrared spectroscopy of the brain," *Applied Optics*, vol. 48, no. 10, pp. 280–298, 2009.
- [9] M. Schneider and I. Miyai, "Review of functional near-infrared spectroscopy in neurorehabilitation," *Neurophotonics*, vol. 3, Article ID 031414, 2016.
- [10] A. Venkatakrishnan, G. E. Francisco, and J. L. Contreras-Vidal, "Applications of brain-machine interface systems in stroke recovery and rehabilitation," *Current Physical Medicine and Rehabilitation Reports*, vol. 2, no. 2, pp. 93–105, 2014.
- [11] F. Jobsis, "Noninvasive, infrared monitoring of cerebral and myocardial oxygen sufficiency and circulatory parameters," *Science*, vol. 198, no. 4323, pp. 1264–1267, 1977.
- [12] S. Brigadoi, L. Ceccherini, S. Cutini et al., "Motion artifacts in functional near-infrared spectroscopy: a comparison of motion correction techniques applied to real cognitive data," *Neuroimage*, vol. 85, pp. 181–191, 2014.
- [13] R. J. Cooper, J. Selb, L. Gagnon et al., "A systematic comparison of motion artifact correction techniques for functional near-infrared spectroscopy," *Frontiers in Neuroscience*, vol. 6, pp. 1–10, 2012.
- [14] M. Izzetoglu, P. Chitrapu, S. Bunce, and B. Onaral, "Motion artifact cancellation in NIR spectroscopy using discrete Kalman filtering," *BioMedical Engineering OnLine*, vol. 9, no. 1, p. 16, 2010.
- [15] M. Izzetoglu, A. Devaraj, S. Bunce, and B. Onaral, "Motion artifact cancellation in NIR spectroscopy using wiener filtering," *IEEE Transactions on Biomedical Engineering*, vol. 52, no. 5, pp. 934–938, 2005.
- [16] D. a. Boas, K. Chen, D. Grebert, and M. A. Franceschini, "Improving the diffuse optical imaging spatial resolution of the cerebral hemodynamic response to brain activation in humans," *Optics Letters*, vol. 29, no. 13, p. 1506, 2004.
- [17] D. A. Boas, A. M. Dale, and M. A. Franceschini, "Diffuse optical imaging of brain activation: approaches to optimizing image sensitivity, resolution, and accuracy," *Neuroimage*, vol. 23, pp. S275–S288, 2004.
- [18] M. A. Franceschini, D. K. Joseph, T. J. Huppert, S. G. Diamond, and D. A. Boas, "Diffuse optical imaging of the whole head," *Journal of Biomedical Optics*, vol. 11, no. 5, Article ID 054007, 2006.
- [19] Y. Zhang, D. H. Brooks, M. A. Franceschini, and D. A. Boas, "Eigenvector-based spatial filtering for reduction of physiological interference in diffuse optical imaging," *Journal of Biomedical Optics*, vol. 10, no. 1, Article ID 011014, 2005.
- [20] G. Bauernfeind, S. C. Wriessnegger, I. Daly, and G. R. Müller-Putz, "Separating heart and brain: on the reduction of physiological noise from multichannel functional near-infrared spectroscopy (fNIRS) signals," *Journal of Neural Engineering*, vol. 11, no. 5, Article ID 056010, 2014.

- [21] E. Kirilina, N. Yu, A. Jelzow, H. Wabnitz, A. M. Jacobs, and I. Tachtsidis, "Identifying and quantifying main components of physiological noise in functional near infrared spectroscopy on the prefrontal cortex," *Frontiers in Human Neuroscience*, vol. 7, pp. 1–17, 2013.
- [22] N. Naseer, N. K. Qureshi, F. M. Noori, and K.-S. Hong, "Analysis of different classification techniques for two-class functional near-infrared spectroscopy-based brain-computer interface," *Computational Intelligence and Neuroscience*, vol. 2016, Article ID 5480760, 11 pages, 2016.
- [23] M. D. Pfeifer, F. Scholkmann, and R. Labruyère, "Signal processing in functional near-infrared spectroscopy (fNIRS): methodological differences lead to different statistical results," *Frontiers in Human Neuroscience*, vol. 11, pp. 1–12, 2018.
- [24] M. J. Grimble, *Robust Industrial Control Systems*, John Wiley & Sons, Chichester, UK, 2006.
- [25] A. Y. Kostiev, A. Y. Butrym, and S. N. Shulga, "Time-varying wiener filtering based on short-time fourier transform," in *Proceedings of the 2012 6th International Conference on Ultrawideband and Ultrashort Impulse Signals*, pp. 305–308, IEEE, Kiev, Ukraine, 2012.
- [26] J. Ye, S. Tak, K. Jang, J. Jung, and J. Jang, "NIRS-SPM: statistical parametric mapping for near-infrared spectroscopy," *Neuroimage*, vol. 44, no. 2, pp. 428–447, 2009.
- [27] R. B. Lake, "Programs for digital signal processing," *Proceedings of the IEEE*, vol. 69, no. 7, pp. 856–857, 1981.
- [28] K.-S. Hong, N. Naseer, and Y.-H. Kim, "Classification of prefrontal and motor cortex signals for three-class fNIRS-BCI," *Neuroscience Letters*, vol. 587, pp. 87–92, 2015.
- [29] M. Mihara, I. Miyai, M. Hatakenaka, K. Kubota, and S. Sakoda, "Role of the prefrontal cortex in human balance control," *Neuroimage*, vol. 43, no. 2, pp. 329–336, 2008.
- [30] S. D. Power and T. Chau, "Automatic single-trial classification of prefrontal hemodynamic activity in an individual with Duchenne muscular dystrophy," *Developmental Neurorehabilitation*, vol. 16, no. 1, pp. 67–72, 2012.
- [31] S. D. Power, T. H. Falk, and T. Chau, "Classification of prefrontal activity due to mental arithmetic and music imagery using hidden Markov models and frequency domain near-infrared spectroscopy," *Journal of Neural Engineering*, vol. 7, no. 2, Article ID 026002, 2010.
- [32] N. Naseer, F. M. Noori, N. K. Qureshi, and K.-S. Hong, "Determining optimal feature-combination for LDA classification of functional near-infrared spectroscopy signals in brain-computer interface application," *Frontiers in Human Neuroscience*, vol. 10, pp. 1–10, 2016.
- [33] F. M. Noori, N. Naseer, N. K. Qureshi, H. Nazeer, and R. A. Khan, "Optimal feature selection from fNIRS signals using genetic algorithms for BCI," *Neuroscience Letters*, vol. 647, pp. 61–66, 2017.
- [34] N. K. Qureshi, N. Naseer, F. M. Noori, H. Nazeer, R. A. Khan, and S. Saleem, "Enhancing classification performance of functional near-infrared spectroscopy- brain-computer interface using adaptive estimation of general linear model coefficients," *Frontiers in Neurobotics*, vol. 11, 2017.
- [35] K.-S. Hong and N. Naseer, "Reduction of delay in detecting initial dips from functional near-infrared spectroscopy signals using vector-based phase Analysis," *International Journal of Neural Systems*, vol. 26, no. 3, Article ID 1650012, 2016.
- [36] P.-H. Chang, S.-H. Lee, K.-M. Gu et al., "The cortical activation pattern by a rehabilitation robotic hand: a functional NIRS study," *Frontiers in Human Neuroscience*, vol. 8, pp. 1–7, 2014.
- [37] S. P. Perrey, "Possibilities for examining the neural control of gait in humans with fNIRS," *Frontiers in Physiology*, vol. 5, pp. 10–13, 2014.
- [38] N. Thanh Hai, N. Q. Cuong, T. Q. Dang Khoa, and V. Van Toi, "Temporal hemodynamic classification of two hands tapping using functional near-infrared spectroscopy," *Frontiers in Human Neuroscience*, vol. 7, pp. 1–12, 2013.
- [39] X. Liu and K.-S. Hong, "fNIRS based color detection from human visual cortex," in *Proceedings of the 2015 54th Annual Conference of the Society of Instrument and Control Engineers of Japan (SICE)*, pp. 1156–1161, IEEE, Hangzhou, China, July 2015.
- [40] C. J. C. Burges, "A tutorial on support vector machines for pattern recognition," *Data Mining and Knowledge Discovery*, vol. 2, pp. 121–167, 1998.
- [41] C. Ashina and V. Vapnik, "Support-vector networks," *Machine Learning*, vol. 20, no. 3, pp. 273–297, 1995.
- [42] M. Naito, Y. Michioka, K. Ozawa, Y. Ito, M. Kiguchi, and T. Kanazawa, "A communication means for totally locked-in ALS patients based on changes in cerebral blood volume measured with near-infrared light," *IEICE Transactions on Information and Systems*, vol. E90-D, no. 7, pp. 1028–1037, 2007.
- [43] N. Naseer and K.-S. Hong, "Classification of functional near-infrared spectroscopy signals corresponding to the right- and left-wrist motor imagery for development of a brain-computer interface," *Neuroscience Letters*, vol. 553, pp. 84–89, 2013.
- [44] N. Naseer and K.-S. Hong, "Decoding answers to four-choice questions using functional near infrared spectroscopy," *Journal of Near Infrared Spectroscopy*, vol. 23, no. 1, pp. 23–31, 2015.
- [45] R. Lima-Saad, H. Zhang, C. Guan et al., "Temporal classification of multichannel near-infrared spectroscopy signals of motor imagery for developing a brain-computer interface," *Neuroimage*, vol. 34, no. 4, pp. 1416–1427, 2007.
- [46] N. Naseer, M. J. Hong, and K.-S. Hong, "Online binary decision decoding using functional near-infrared spectroscopy for the development of brain-computer interface," *Experimental Brain Research*, vol. 232, no. 2, pp. 555–564, 2014.
- [47] S. D. Power, A. Kushki, and T. Chau, "Towards a system-paced near-infrared spectroscopy brain-computer interface: differentiating prefrontal activity due to mental arithmetic and mental singing from the no-control state," *Journal of Neural Engineering*, vol. 8, no. 6, Article ID 066004, 2011.
- [48] B. Abibullaev and J. An, "Classification of frontal cortex haemodynamic responses during cognitive tasks using wavelet transforms and machine learning algorithms," *Medical Engineering & Physics*, vol. 34, no. 10, pp. 1394–1410, 2012.
- [49] A. Faress and T. Chau, "Towards a multimodal brain-computer interface: combining fNIRS and fTCD measurements to enable higher classification accuracy," *Neuroimage*, vol. 77, pp. 186–194, 2013.
- [50] H.-J. Hwang, J.-H. Lim, D.-W. Kim, and C.-H. Im, "Evaluation of various mental task combinations for near-infrared spectroscopy-based brain-computer interfaces," *Journal of Biomedical Optics*, vol. 19, Article ID 077005, 2014.
- [51] N. Naseer, F. Ali, S. Ahmed, S. Iftikhar, R. A. Khan, and H. Nazeer, "EMG based control of individual fingers of robotic hand," in *Proceedings of the 2018 International Conference on Sustainable Information Engineering and Technology (SIET)*, Malang, Indonesia, November 2018.

- [52] R. Holtzer, J. R. Mahoney, M. Izzetoglu, C. Wang, S. England, and J. Verghese, "Online fronto-cortical control of simple and attention-demanding locomotion in humans," *Neuroimage*, vol. 112, pp. 152–159, 2015.
- [53] R. A. Khan, N. Naseer, N. K. Qureshi, F. M. Noori, H. Nazeer, and M. U. Khan, "fNIRS-based Neurobotic Interface for gait rehabilitation," *Journal of NeuroEngineering and Rehabilitation*, vol. 15, no. 1, p. 7, 2018.
- [54] R. Khan, N. Naseer, H. Nazeer, and M. N. Khan, "Control of a prosthetic leg based on walking intentions for gait rehabilitation: an fNIRS study," *Frontiers in Human Neuroscience*, vol. 12, 2018.
- [55] F. M. Noori, N. K. Qureshi, R. A. Khan, and N. Naseer, "Feature selection based on modified genetic algorithm for optimization of functional near-infrared spectroscopy (fNIRS) signals for BCI," in *Proceedings of the 2016 2nd International Conference on Robotics and Artificial Intelligence (ICRAI)*, pp. 50–53, IEEE, Los Angeles, CA, USA, April 2016.
- [56] R. A. Khan, N. Naseer, and M. J. Khan, "Drowsiness detection during a driving task using fNIRS," in *Neuroergonomics*, pp. 79–85, Elsevier, Amsterdam, Netherlands, 2019.
- [57] R. Vitorio, S. Stuart, L. Rochester, L. Alcock, and A. Pantall, "fNIRS response during walking-artefact or cortical activity? A systematic review," *Neuroscience & Biobehavioral Reviews*, vol. 83, pp. 160–172, 2017.
- [58] S. Kubota and J. C. Ye, "Statistical analysis of fNIRS data: a comprehensive review," *Neuroimage*, vol. 85, pp. 72–91, 2014.
- [59] R. Rosas-Romero, E. Guevara, K. Peng et al., "Prediction of epileptic seizures with convolutional neural networks and functional near-infrared spectroscopy signals," *Computers in Biology and Medicine*, vol. 111, Article ID 103355, 2019.
- [60] A. Rahimpour, H. A. Noubari, and M. Kazemian, "A case-study of NIRS application for infant cerebral hemodynamic monitoring: a report of data analysis for feature extraction and infant classification into healthy and unhealthy," *Informatics in Medicine Unlocked*, vol. 11, pp. 44–50, 2018.
- [61] R. Lamb, P. Antonenko, E. Etopio, and A. Seccia, "Comparison of virtual reality and hands on activities in science education via functional near infrared spectroscopy," *Computers & Education*, vol. 124, pp. 14–26, 2018.
- [62] M. Aqil, K.-S. Hong, M.-Y. Jeong, and S. S. Ge, "Cortical brain imaging by adaptive filtering of NIRS signals," *Neuroscience Letters*, vol. 514, no. 1, pp. 35–41, 2012.
- [63] R. J. Seo, L. Gagnon, D. M. Goldenholz, D. A. Boas, and D. N. Greve, "The utility of near-infrared spectroscopy in the regression of low-frequency physiological noise from functional magnetic resonance imaging data," *Neuroimage*, vol. 59, no. 4, pp. 3128–3138, 2012a.
- [64] M. A. Kamran and K.-S. Hong, "Linear parameter-varying model and adaptive filtering technique for detecting neuronal activities: an fNIRS study," *Journal of Neural Engineering*, vol. 10, no. 5, Article ID 056002, 2013.
- [65] J. Höhne, E. Holz, P. Staiger-Sälzer, K.-R. Müller, A. Kübler, and M. Tangermann, "Motor imagery for severely motor-impaired patients: evidence for brain-computer interfacing as superior control solution," *PLoS One*, vol. 9, no. 8, Article ID e104854, 2014.
- [66] R. Zimmermann, L. Marchal-Crespo, J. Edelmann et al., "Detection of motor execution using a hybrid fNIRS-biosignal BCI: a feasibility study," *Journal of NeuroEngineering and Rehabilitation*, vol. 10, no. 1, p. 4, 2013.
- [67] R. Beurskens, I. Helmich, R. Rein, and O. Bock, "Age-related changes in prefrontal activity during walking in dual-task situations: a fNIRS study," *International Journal of Psychophysiology*, vol. 92, no. 3, pp. 122–128, 2014.
- [68] P. Eggenberger, M. Wolf, M. Schumann, and E. D. de Bruin, "Exergame and balance training modulate prefrontal brain activity during walking and enhance executive function in older adults," *Frontiers in Aging Neuroscience*, vol. 8, pp. 1–16, 2016.
- [69] A. Izzetoglu, C. Bohil, J. Lewis, and M. Neider, "Prefrontal cortex activity during walking while multitasking," *Proceedings of the Human Factors and Ergonomics Society Annual Meeting*, vol. 57, no. 1, pp. 1224–1228, 2013.
- [70] R. Holtzer, J. R. Mahoney, M. Izzetoglu, K. Izzetoglu, B. Onaral, and J. Verghese, "fNIRS study of walking and walking while talking in young and old individuals," *The Journals of Gerontology Series A: Biological Sciences and Medical Sciences*, vol. 66A, no. 8, pp. 879–887, 2011.
- [71] M. E. Hernandez, R. Holtzer, G. Chaparro et al., "Brain activation changes during locomotion in middle-aged to older adults with multiple sclerosis," *Journal of the Neurological Sciences*, vol. 370, pp. 277–283, 2016.
- [72] R. Holtzer, C. Schoen, E. Demetriadou et al., "Stress and gender effects on prefrontal cortex oxygenation levels assessed during single and dual-task walking conditions," *European Journal of Neuroscience*, vol. 45, no. 5, pp. 660–670, 2017.
- [73] R. Verghese, J. Verghese, G. Allali, M. Izzetoglu, C. Wang, and J. R. Mahoney, "Neurological gait abnormalities moderate the functional brain signature of the posture first hypothesis," *Brain Topography*, vol. 29, no. 2, pp. 334–343, 2016.
- [74] R. Holtzer, J. Yuan, J. Verghese, J. R. Mahoney, M. Izzetoglu, and C. Wang, "Interactions of subjective and objective measures of fatigue defined in the context of brain control of locomotion," *The Journals of Gerontology Series A: Biological Sciences and Medical Sciences*, vol. 72, pp. 167–423, 2016.
- [75] I. Maidan, H. Bernad-Elazari, E. Gazit, N. Giladi, J. M. Hausdorff, and A. Mirelman, "Changes in oxygenated hemoglobin link freezing of gait to frontal activation in patients with Parkinson disease: an fNIRS study of transient motor-cognitive failures," *Journal of Neurology*, vol. 262, no. 4, pp. 899–908, 2015.
- [76] A. Mirelman, I. Maidan, H. Bernad-Elazari et al., "Increased frontal brain activation during walking while dual tasking: an fNIRS study in healthy young adults," *Journal of NeuroEngineering and Rehabilitation*, vol. 11, no. 1, p. 85, 2014.
- [77] J. Verghese, C. Wang, E. Ayers, M. Izzetoglu, and R. Holtzer, "Brain activation in high-functioning older adults and falls," *Neurology*, vol. 88, no. 2, pp. 191–197, 2017.
- [78] M.-I. B. Lin and K.-H. Lin, "Walking while performing working memory tasks changes the prefrontal cortex hemodynamic activations and gait kinematics," *Frontiers in Behavioral Neuroscience*, vol. 10, 2016.
- [79] H. Y. Kim, S. P. Yang, G. L. Park, E. J. Kim, and J. H. You, "Best facilitated cortical activation during different stepping, treadmill, and robot-assisted walking training paradigms and speeds: a functional near-infrared spectroscopy neuroimaging study," *NeuroRehabilitation*, vol. 38, no. 2, pp. 171–178, 2016.
- [80] K. L. M. Koenraadt, E. G. J. Roelofsen, J. Duyens, and N. L. W. Keijsers, "Cortical control of normal gait and precision stepping: an fNIRS study," *Neuroimage*, vol. 85, pp. 415–422, 2014.
- [81] C.-F. Lu, Y.-C. Liu, Y.-R. Yang, Y.-T. Wu, and R.-Y. Wang, "Maintaining gait performance by cortical activation during dual-task interference: a functional near-infrared spectroscopy study," *PLoS One*, vol. 10, no. 6, Article ID e0129390, 2015.

- [82] I. Maidan, F. Nieuwhof, H. Bernad-Elazari et al., “The role of the frontal lobe in complex walking among patients with Parkinson’s disease and healthy older adults,” *Neuro-rehabilitation and Neural Repair*, vol. 30, no. 10, pp. 963–971, 2016.
- [83] M. J. Kurz, T. W. Wilson, and D. J. Arpin, “Stride-time variability and sensorimotor cortical activation during walking,” *Neuroimage*, vol. 59, no. 2, pp. 1602–1607, 2012.
- [84] F. G. Metzger, A.-C. Ehliis, F. B. Haeussinger et al., “Functional brain imaging of walking while talking—an fNIRS study,” *Neuroscience*, vol. 343, pp. 85–93, 2017.
- [85] I. Hausdorff, H. C. Tanabe, I. Sase et al., “Cortical mapping of gait in humans: a near-infrared spectroscopic topography study,” *Neuroimage*, vol. 14, no. 5, pp. 1186–1192, 2001.
- [86] M. Shimizu, I. Miyai, T. Ono, and K. Kubota, “Activities in the frontal cortex and gait performance are modulated by preparation. An fNIRS study,” *Neuroimage*, vol. 39, no. 2, pp. 600–607, 2008.
- [87] M. Suzuki, I. Miyai, T. Ono et al., “Prefrontal and premotor cortices are involved in adapting walking and running speed on the treadmill: an optical imaging study,” *Neuroimage*, vol. 23, no. 3, pp. 1020–1026, 2004.
- [88] F. Nieuwhof, M. F. Reelick, I. Maidan et al., “Measuring prefrontal cortical activity during dual task walking in patients with Parkinson’s disease: feasibility of using a new portable fNIRS device,” *Pilot and Feasibility Studies*, vol. 2, no. 1, p. 59, 2016.